

Continuing Calibration Report for SequenceID : FC061125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069174.D

Aliphatic C9-C12	5936222.000	60.000	3.303	6.605	98937.033	105717.491	6.414
Aliphatic C12-C16	3691689.000	40.000	6.606	10.010	92292.225	101985.435	9.505
Aliphatic C16-C21	5157507.000	60.000	10.011	13.380	85958.450	97865.081	12.166
Aliphatic C21-C28	7042755.000	80.000	13.381	17.048	88034.438	93803.915	6.151
Aliphatic C28-C40	10687110.000	120.000	17.049	22.033	89059.250	94439.035	5.697
Aliphatic EPH	32515283.000	360.000	3.303	22.033	90320.231	97587.136	7.447