

Continuing Calibration Report for SequenceID : FD060925AR

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FD049414.D

|                  |              |         |        |        |            |            |        |
|------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aromatic C10-C12 | 6066366.000  | 40.000  | 4.424  | 6.169  | 151659.150 | 167344.833 | 9.373  |
| Aromatic C12-C16 | 9855692.000  | 60.000  | 6.170  | 8.798  | 164261.533 | 187525.524 | 12.406 |
| Aromatic C16-C21 | 12307850.000 | 80.000  | 8.799  | 13.089 | 153848.125 | 184740.440 | 16.722 |
| Aromatic C21-C36 | 25693832.000 | 180.000 | 13.090 | 18.512 | 142743.511 | 145057.816 | 1.595  |
| Aromatic EPH     | 53923740.000 | 360.000 | 4.424  | 18.512 | 149788.167 | 163430.463 | 8.347  |