

Continuing Calibration Report for SequenceID : FD060925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049421.D

Aromatic C10-C12	6920053.000	40.000	4.424	6.169	173001.325	167344.833	-3.380
Aromatic C12-C16	11045426.000	60.000	6.170	8.798	184090.433	187525.524	1.832
Aromatic C16-C21	13902265.000	80.000	8.799	13.089	173778.313	184740.440	5.934
Aromatic C21-C36	28417137.000	180.000	13.090	18.512	157872.983	145057.816	-8.835
Aromatic EPH	60284881.000	360.000	4.424	18.512	167458.003	163430.463	-2.464