

Continuing Calibration Report for SequenceID : FD061025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049424.D

Aromatic C10-C12	6278489.000	40.000	4.425	6.169	156962.225	167344.833	6.204
Aromatic C12-C16	10256841.000	60.000	6.170	8.798	170947.350	187525.524	8.840
Aromatic C16-C21	12991778.000	80.000	8.799	13.089	162397.225	184740.440	12.094
Aromatic C21-C36	26792100.000	180.000	13.090	18.511	148845.000	145057.816	-2.611
Aromatic EPH	56319208.000	360.000	4.425	18.511	156442.244	163430.463	4.276