

Continuing Calibration Report for SequenceID : FD061025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049427.D

Aromatic C10-C12	6485261.000	40.000	4.425	6.169	162131.525	167344.833	3.115
Aromatic C12-C16	10630607.000	60.000	6.170	8.798	177176.783	187525.524	5.519
Aromatic C16-C21	13401211.000	80.000	8.799	13.089	167515.138	184740.440	9.324
Aromatic C21-C36	27453656.000	180.000	13.090	18.511	152520.311	145057.816	-5.144
Aromatic EPH	57970735.000	360.000	4.425	18.511	161029.819	163430.463	1.469