

Continuing Calibration Report for SequenceID : FD061125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049430.D

Aromatic C10-C12	6205695.000	40.000	4.424	6.168	155142.375	167344.833	7.292
Aromatic C12-C16	9970470.000	60.000	6.169	8.797	166174.500	187525.524	11.386
Aromatic C16-C21	12701998.000	80.000	8.798	13.088	158774.975	184740.440	14.055
Aromatic C21-C36	25903906.000	180.000	13.089	18.509	143910.589	145057.816	0.791
Aromatic EPH	54782069.000	360.000	4.424	18.509	152172.414	163430.463	6.889