

Continuing Calibration Report for SequenceID : FD061125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049443.D

Aromatic C10-C12	6392686.000	40.000	4.424	6.168	159817.150	167344.833	4.498
Aromatic C12-C16	10225061.000	60.000	6.169	8.797	170417.683	187525.524	9.123
Aromatic C16-C21	13364565.000	80.000	8.798	13.088	167057.063	184740.440	9.572
Aromatic C21-C36	27341316.000	180.000	13.089	18.509	151896.200	145057.816	-4.714
Aromatic EPH	57323628.000	360.000	4.424	18.509	159232.300	163430.463	2.569