

Continuing Calibration Report for SequenceID : FD061125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049449.D

Aromatic C10-C12	6540731.000	40.000	4.424	6.168	163518.275	167344.833	2.287
Aromatic C12-C16	10481876.000	60.000	6.169	8.797	174697.933	187525.524	6.840
Aromatic C16-C21	13688224.000	80.000	8.798	13.088	171102.800	184740.440	7.382
Aromatic C21-C36	27718945.000	180.000	13.089	18.509	153994.139	145057.816	-6.161
Aromatic EPH	58429776.000	360.000	4.424	18.509	162304.933	163430.463	0.689