

Continuing Calibration Report for SequenceID : FD061325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049481.D

Aromatic C10-C12	6695791.000	40.000	4.424	6.168	167394.775	167344.833	-0.030
Aromatic C12-C16	10586658.000	60.000	6.169	8.796	176444.300	187525.524	5.909
Aromatic C16-C21	13509794.000	80.000	8.797	13.086	168872.425	184740.440	8.589
Aromatic C21-C36	27183534.000	180.000	13.087	18.508	151019.633	145057.816	-4.110
Aromatic EPH	57975777.000	360.000	4.424	18.508	161043.825	163430.463	1.460