

Continuing Calibration Report for SequenceID : FD061325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049489.D

Aromatic C10-C12	6667078.000	40.000	4.424	6.168	166676.950	167344.833	0.399
Aromatic C12-C16	10603581.000	60.000	6.169	8.796	176726.350	187525.524	5.759
Aromatic C16-C21	13634458.000	80.000	8.797	13.086	170430.725	184740.440	7.746
Aromatic C21-C36	27639906.000	180.000	13.087	18.508	153555.033	145057.816	-5.858
Aromatic EPH	58545023.000	360.000	4.424	18.508	162625.064	163430.463	0.493