

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q2198 **SAS No :** Q2198 **SDG No:** Q2198
Sample No : PB168408BS **Datafile:** FC069158.D **Client ID :** PB168408BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	4.97	50		(40-140)
Aliphatic C12-C16	6.7	4.11	62		(40-140)
Aliphatic C16-C21	10	7.02	70		(40-140)
Aliphatic C21-C28	13.3	10.0	75		(40-140)
Aliphatic C28-C40	20.0	14.6	73		(40-140)