

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q2198 **SAS No :** Q2198 **SDG No:** Q2198
Sample No : PB168408BSD **Datafile:** FC069159.D **Client ID :** PB168408BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.03	50		1.3 (40-140)	25
Aliphatic C12-C16	6.7	4.15	62		0.98 (40-140)	25
Aliphatic C16-C21	10	7.07	71		0.8 (40-140)	25
Aliphatic C21-C28	13.3	10.1	76		1.1 (40-140)	25
Aliphatic C28-C40	20.0	15.0	75		2.8 (40-140)	25