



|                |               |                      |              |                   |                   |          |              |
|----------------|---------------|----------------------|--------------|-------------------|-------------------|----------|--------------|
| Contract:      | <u>PORT06</u> |                      |              |                   |                   |          |              |
| Lab Code:      | <u>CHEM</u>   | Case No.:            | <u>Q2198</u> | SAS No.:          | <u>Q2198</u>      | SDG NO.: | <u>Q2198</u> |
| Instrument ID: | <u>ECD_P</u>  | Calibration Date(s): |              | <u>05/19/2025</u> | <u>05/19/2025</u> |          |              |
|                |               | Calibration Times:   |              | <u>09:42</u>      | <u>18:25</u>      |          |              |

|                     |           |                   |          |                   |
|---------------------|-----------|-------------------|----------|-------------------|
| LAB FILE ID:        | RT 1000 = | <u>PP072164.D</u> | RT 750 = | <u>PP072165.D</u> |
| RT 500 = PP072166.D | RT 250 =  | PP072167.D        | RT 050 = | PP072168.D        |

[illegible]

**RETENTION TIMES OF INITIAL CALIBRATION**

|                      |       |       |       |       |       |       |       |       |
|----------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| Decachlorobiphenyl   | 10.22 | 10.22 | 10.22 | 10.22 | 10.22 | 10.22 | 10.12 | 10.32 |
| Tetrachloro-m-xylene | 4.50  | 4.50  | 4.50  | 4.50  | 4.50  | 4.50  | 4.40  | 4.60  |
| Aroclor-1268-1 (1)   | 8.72  | 8.72  | 8.72  | 8.72  | 8.72  | 8.72  | 8.62  | 8.82  |
| Aroclor-1268-2 (2)   | 8.81  | 8.82  | 8.81  | 8.81  | 8.81  | 8.81  | 8.71  | 8.91  |
| Aroclor-1268-3 (3)   | 9.04  | 9.05  | 9.04  | 9.04  | 9.05  | 9.04  | 8.94  | 9.14  |
| Aroclor-1268-4 (4)   | 9.46  | 9.46  | 9.46  | 9.46  | 9.46  | 9.46  | 9.36  | 9.56  |
| Aroclor-1268-5 (5)   | 9.88  | 9.88  | 9.88  | 9.88  | 9.88  | 9.88  | 9.78  | 9.98  |
| Decachlorobiphenyl   | 10.22 | 10.22 | 10.22 | 10.21 | 10.22 | 10.22 | 10.12 | 10.32 |
| Tetrachloro-m-xylene | 4.50  | 4.51  | 4.50  | 4.50  | 4.50  | 4.50  | 4.40  | 4.60  |

### RETENTION TIMES OF INITIAL CALIBRATION

|                |               |                      |              |                   |                   |          |              |
|----------------|---------------|----------------------|--------------|-------------------|-------------------|----------|--------------|
| Contract:      | <u>PORT06</u> |                      |              |                   |                   |          |              |
| Lab Code:      | <u>CHEM</u>   | Case No.:            | <u>Q2198</u> | SAS No.:          | <u>Q2198</u>      | SDG NO.: | <u>Q2198</u> |
| Instrument ID: | <u>ECD_P</u>  | Calibration Date(s): |              | <u>05/19/2025</u> | <u>05/19/2025</u> |          |              |
|                |               | Calibration Times:   |              | <u>09:42</u>      | <u>18:25</u>      |          |              |

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP072164.D RT 750 = PP072165.D  
RT 500 = PP072166.D RT 250 = PP072167.D RT 050 = PP072168.D

[illegible]

[illegible]