

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO111468.D Time Analyzed: 15:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	523.360	500.000	4.7
Aroclor-1016-2	4.786	4.692	4.892	535.210	500.000	7.0
Aroclor-1016-3	4.843	4.749	4.949	546.050	500.000	9.2
Aroclor-1016-4	4.964	4.869	5.069	538.370	500.000	7.7
Aroclor-1016-5	5.221	5.126	5.326	540.470	500.000	8.1
Aroclor-1260-1	6.260	6.166	6.366	544.420	500.000	8.9
Aroclor-1260-2	6.449	6.355	6.555	563.730	500.000	12.7
Aroclor-1260-3	6.816	6.722	6.922	555.280	500.000	11.1
Aroclor-1260-4	7.076	6.983	7.183	519.810	500.000	4.0
Aroclor-1260-5	7.318	7.225	7.425	519.560	500.000	3.9
Decachlorobiphenyl	8.715	8.624	8.824	47.370	50.000	-5.3
Tetrachloro-m-xylene	3.679	3.582	3.782	51.580	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO111468.D Time Analyzed: 15:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.753	4.658	4.858	490.320	500.000	-1.9
Aroclor-1016-2	4.772	4.677	4.877	491.420	500.000	-1.7
Aroclor-1016-3	4.947	4.852	5.052	487.880	500.000	-2.4
Aroclor-1016-4	4.990	4.894	5.094	482.620	500.000	-3.5
Aroclor-1016-5	5.202	5.107	5.307	500.070	500.000	0.0
Aroclor-1260-1	6.232	6.138	6.338	486.300	500.000	-2.7
Aroclor-1260-2	6.420	6.326	6.526	482.110	500.000	-3.6
Aroclor-1260-3	6.572	6.479	6.679	449.800	500.000	-10.0
Aroclor-1260-4	7.042	6.949	7.149	460.990	500.000	-7.8
Aroclor-1260-5	7.284	7.191	7.391	448.300	500.000	-10.3
Decachlorobiphenyl	8.665	8.574	8.774	44.590	50.000	-10.8
Tetrachloro-m-xylene	3.675	3.579	3.779	51.160	50.000	2.3