

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP072647.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	428.270	500.000	-14.3
Aroclor-1016-2	5.675	5.574	5.774	474.780	500.000	-5.0
Aroclor-1016-3	5.737	5.636	5.836	471.220	500.000	-5.8
Aroclor-1016-4	5.833	5.734	5.934	485.870	500.000	-2.8
Aroclor-1016-5	6.127	6.027	6.227	479.940	500.000	-4.0
Aroclor-1260-1	7.243	7.147	7.347	483.520	500.000	-3.3
Aroclor-1260-2	7.497	7.401	7.601	468.460	500.000	-6.3
Aroclor-1260-3	7.855	7.760	7.960	469.730	500.000	-6.1
Aroclor-1260-4	8.079	7.985	8.185	456.510	500.000	-8.7
Aroclor-1260-5	8.397	8.303	8.503	477.030	500.000	-4.6
Decachlorobiphenyl	10.199	10.113	10.313	49.390	50.000	-1.2
Tetrachloro-m-xylene	4.501	4.396	4.596	48.870	50.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP072647.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.777	4.977	494.320	500.000	-1.1
Aroclor-1016-2	4.893	4.795	4.995	509.090	500.000	1.8
Aroclor-1016-3	5.069	4.973	5.173	510.790	500.000	2.2
Aroclor-1016-4	5.111	5.015	5.215	512.100	500.000	2.4
Aroclor-1016-5	5.325	5.229	5.429	546.930	500.000	9.4
Aroclor-1260-1	6.356	6.264	6.464	538.020	500.000	7.6
Aroclor-1260-2	6.545	6.453	6.653	507.010	500.000	1.4
Aroclor-1260-3	6.697	6.606	6.806	552.270	500.000	10.5
Aroclor-1260-4	7.167	7.077	7.277	535.910	500.000	7.2
Aroclor-1260-5	7.409	7.319	7.519	522.040	500.000	4.4
Decachlorobiphenyl	8.807	8.723	8.923	54.390	50.000	8.8
Tetrachloro-m-xylene	3.794	3.691	3.891	51.560	50.000	3.1