

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 06/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP072662.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	428.920	500.000	-14.2
Aroclor-1016-2	5.674	5.574	5.774	465.870	500.000	-6.8
Aroclor-1016-3	5.737	5.636	5.836	462.450	500.000	-7.5
Aroclor-1016-4	5.834	5.734	5.934	479.270	500.000	-4.1
Aroclor-1016-5	6.126	6.027	6.227	475.120	500.000	-5.0
Aroclor-1260-1	7.244	7.147	7.347	491.580	500.000	-1.7
Aroclor-1260-2	7.497	7.401	7.601	476.450	500.000	-4.7
Aroclor-1260-3	7.856	7.760	7.960	488.700	500.000	-2.3
Aroclor-1260-4	8.080	7.985	8.185	488.300	500.000	-2.3
Aroclor-1260-5	8.397	8.303	8.503	492.150	500.000	-1.6
Decachlorobiphenyl	10.201	10.113	10.313	50.470	50.000	0.9
Tetrachloro-m-xylene	4.501	4.396	4.596	47.830	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 06/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP072662.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.877	4.777	4.977	497.640	500.000	-0.5
Aroclor-1016-2	4.895	4.795	4.995	514.780	500.000	3.0
Aroclor-1016-3	5.072	4.973	5.173	512.490	500.000	2.5
Aroclor-1016-4	5.114	5.015	5.215	510.570	500.000	2.1
Aroclor-1016-5	5.327	5.229	5.429	567.690	500.000	13.5
Aroclor-1260-1	6.360	6.264	6.464	541.730	500.000	8.3
Aroclor-1260-2	6.547	6.453	6.653	505.820	500.000	1.2
Aroclor-1260-3	6.699	6.606	6.806	538.290	500.000	7.7
Aroclor-1260-4	7.170	7.077	7.277	540.420	500.000	8.1
Aroclor-1260-5	7.413	7.319	7.519	540.930	500.000	8.2
Decachlorobiphenyl	8.811	8.723	8.923	55.730	50.000	11.5
Tetrachloro-m-xylene	3.796	3.691	3.891	51.930	50.000	3.9