

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** PORT06

**Lab Code:** CHEM      **Case No.:** Q2198      **SAS No.:** Q2198      **SDG NO.:** Q2198

**Continuing Calib Date:** 06/06/2025      **Initial Calibration Date(s):** 06/04/2025      06/04/2025

**Continuing Calib Time:** 23:12      **Initial Calibration Time(s):** 11:19      12:55

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.36       | 7.37      | 7.27      | 7.47 | 0.01       |
| 2,4-D            | 8.50       | 8.51      | 8.41      | 8.61 | 0.01       |
| 2,4,5-TP(Silvex) | 9.39       | 9.40      | 9.30      | 9.50 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** PORT06

**Lab Code:** CHEM      **Case No.:** Q2198      **SAS No.:** Q2198      **SDG NO.:** Q2198

**Continuing Calib Date:** 06/06/2025      **Initial Calibration Date(s):** 06/04/2025      06/04/2025

**Continuing Calib Time:** 23:12      **Initial Calibration Time(s):** 11:19      12:55

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| 2,4-DCAA         | 7.77       | 7.77      | 7.67      | 7.87  | 0.00       |
| 2,4-D            | 9.03       | 9.03      | 8.93      | 9.13  | 0.00       |
| 2,4,5-TP(Silvex) | 9.94       | 9.94      | 9.84      | 10.04 | 0.00       |