

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030535.D Time Analyzed: 12:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.394	9.304	9.504	753.080	712.500	5.7
2,4-D	8.503	8.412	8.612	714.810	705.000	1.4
2,4-DCAA	7.363	7.269	7.469	763.970	750.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030535.D Time Analyzed: 12:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.938	9.842	10.042	736.850	712.500	3.4
2,4-D	9.030	8.934	9.134	743.830	705.000	5.5
2,4-DCAA	7.771	7.674	7.874	734.990	750.000	-2.0