

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030563.D Time Analyzed: 12:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.391	9.304	9.504	699.960	712.500	-1.8
2,4-D	8.502	8.412	8.612	652.640	705.000	-7.4
2,4-DCAA	7.361	7.269	7.469	717.390	750.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030563.D Time Analyzed: 12:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.936	9.842	10.042	692.700	712.500	-2.8
2,4-D	9.029	8.934	9.134	671.090	705.000	-4.8
2,4-DCAA	7.770	7.674	7.874	688.320	750.000	-8.2