

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030575.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.390	9.304	9.504	721.350	712.500	1.2
2,4-D	8.500	8.412	8.612	682.250	705.000	-3.2
2,4-DCAA	7.360	7.269	7.469	740.450	750.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030575.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.935	9.842	10.042	715.680	712.500	0.4
2,4-D	9.028	8.934	9.134	699.240	705.000	-0.8
2,4-DCAA	7.769	7.674	7.874	720.580	750.000	-3.9