

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060925\
 Data File : PS030563.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2025 12:17
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 07:27:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060425.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 04 13:21:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.361	7.770	2708.2E6	740.4E6	717.386m	688.322
Target Compounds							
1) T	Dalapon	2.713	2.712	3826.6E6	1772.6E6	677.636	646.945
2) T	3,5-DICHL...	6.518	6.715	3720.9E6	1041.7E6	685.716	665.963
3) T	4-Nitroph...	7.161	7.301	1156.9E6	1048.1E6	630.871	668.728
5) T	DICAMBA	7.553	7.973	10900.8E6	4389.1E6	706.810	681.367
6) T	MCPD	7.734	8.073	697.9E6	156.5E6	70.455m	67.681
7) T	MCPA	7.885	8.322	839.7E6	195.9E6	66.174	63.470
8) T	DICHLORPROP	8.269	8.693	2466.6E6	990.6E6	677.987	667.410
9) T	2,4-D	8.502	9.029	2378.0E6	1055.2E6	652.639	671.095
10) T	Pentachlo...	8.810	9.556	38420.2E6	25785.3E6	756.510	713.524
11) T	2,4,5-TP ...	9.391	9.936	14464.7E6	9501.5E6	699.960	692.697
12) T	2,4,5-T	9.686	10.362	12136.6E6	8928.3E6	658.488	683.786
13) T	2,4-DB	10.265	10.929	1858.8E6	755.8E6	612.179m	657.494
14) T	DINOSEB	11.488	11.315	9872.0E6	6938.4E6	686.748	690.384
15) T	Picloram	11.294	12.423	11514.4E6	14358.7E6	575.584m	650.189
16) T	DCPA	11.784	12.359	18859.8E6	14555.7E6	713.303	711.321m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060925\
Data File : PS030563.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2025 12:17
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 07:27:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060425.M
Quant Title : 8080.M
QLast Update : Wed Jun 04 13:21:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

