

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/09/2025

Lab Sample No.: PSTDCCC050 Data File : PD088841.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.074	8.977	9.177	52.910	50.000	5.8
Endrin	6.574	6.477	6.677	54.770	50.000	9.5
gamma-BHC (Lindane)	4.329	4.232	4.432	54.420	50.000	8.8
Heptachlor	4.929	4.832	5.032	54.980	50.000	10.0
Heptachlor epoxide	5.690	5.593	5.793	54.130	50.000	8.3
Methoxychlor	7.494	7.395	7.595	53.630	50.000	7.3
Tetrachloro-m-xylene	3.549	3.451	3.651	58.230	50.000	16.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/09/2025

Lab Sample No.: PSTDCCC050 Data File : PD088841.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.072	7.977	8.177	50.760	50.000	1.5
Endrin	5.789	5.692	5.892	48.240	50.000	-3.5
gamma-BHC (Lindane)	3.729	3.631	3.831	48.620	50.000	-2.8
Heptachlor	4.082	3.985	4.185	48.030	50.000	-3.9
Heptachlor epoxide	4.872	4.775	4.975	47.980	50.000	-4.0
Methoxychlor	6.754	6.658	6.858	48.510	50.000	-3.0
Tetrachloro-m-xylene	2.879	2.782	2.982	48.610	50.000	-2.8