

## Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/01/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/03/25
Client Sample ID:	B-207-SB02MS	SDG No.:	Q2198
Lab Sample ID:	Q2198-04MS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088838.D	1	06/06/25 12:32	06/09/25 14:39	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
58-89-9	gamma-BHC (Lindane)	6.30		0.037	0.50	ug/L
76-44-8	Heptachlor	5.40		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	6.10		0.096	0.50	ug/L
72-20-8	Endrin	6.30		0.032	0.50	ug/L
72-43-5	Methoxychlor	6.20		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
<b>SURROGATES</b>						
2051-24-3	Decachlorobiphenyl	25.9		30 (57) - 150 (171)	129%	SPK: 20
877-09-8	Tetrachloro-m-xylene	24.2		30 (61) - 150 (148)	121%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit