

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2198</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Portal Partners Tri-Venture</u>	Datafile :	<u>PD088833.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168330BS (Column 1)	gamma-BHC (Lindane)	0.5	0.53	ug/L	105				40 (82)	140 (129)		
	Heptachlor	0.5	0.54	ug/L	108				40 (79)	140 (127)		
	Heptachlor epoxide	0.5	0.53	ug/L	106				40 (81)	140 (124)		
	Endrin	0.5	0.54	ug/L	107				40 (81)	140 (128)		
	Methoxychlor	0.5	0.54	ug/L	107				40 (78)	140 (108)		
PB168330BS (Column 2)	gamma-BHC (Lindane)	0.5	0.49	ug/L	98				40 (82)	140 (129)		
	Heptachlor	0.5	0.48	ug/L	97				40 (79)	140 (127)		
	Heptachlor epoxide	0.5	0.49	ug/L	98				40 (81)	140 (124)		
	Endrin	0.5	0.48	ug/L	97				40 (81)	140 (128)		
	Methoxychlor	0.5	0.49	ug/L	97				40 (78)	140 (108)		