

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: PORT06
 ProjectID: Amtrak Sawtooth Bridges 2025
 Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG No.: Q2198

Calibration Sequence : FG052225		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	205023420	120602	FG015886.D
850	105950670	124648	FG015887.D
340	42697248	125580	FG015888.D
170	22436219	131978	FG015889.D
85	10374809	122057	FG015890.D
AVG RF : 124973		% RSD : 3.514	AVG RT : 14.9848