

Analytical Sequence

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

Instrument ID: FID_G

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 14.9848					
EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	10 Jun 2025 14:37	FG016009.D	14.973	
50 PPM TRPH STD	50 PPM TRPH STD	10 Jun 2025 15:06	FG016010.D	14.976	
PB168382BL	PB168382BL	10 Jun 2025 15:36	FG016011.D	14.974	
PB168382BS	PB168382BS	10 Jun 2025 16:05	FG016012.D	14.973	
B-202-SB02	Q2198-01	10 Jun 2025 18:03	FG016016.D	14.973	
B-207-SB02	Q2198-03	10 Jun 2025 18:32	FG016017.D	14.971	
BU-03-060525MS	Q2246-01MS	10 Jun 2025 19:31	FG016019.D	14.973	
BU-03-060525MSD	Q2246-01MSD	10 Jun 2025 20:01	FG016020.D	14.973	
PIBLK02	LBLK02	10 Jun 2025 20:30	FG016021.D	14.974	
50 PPM TRPH STD	50 PPM TRPH STD	10 Jun 2025 21:00	FG016022.D	14.976	