

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG No.: Q2198
Instrument ID: BNA_F Calibration Date(s): 06/10/2025 06/10/2025
Calibration Time(s): 16:54 20:19

LAB FILE ID:		RRF2.5 = BF142712.D			RRF005 = BF142713.D		RRF010 = BF142714.D	
		RRF020 = BF142715.D			RRF040 = BF142716.D		RRF050 = BF142717.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.238	1.170	1.199	1.161	1.220	1.174	4.5
Benzaldehyde			0.943	0.950	0.839	0.839	0.856	11.5
Phenol-d6		1.434	1.339	1.400	1.376	1.446	1.382	3.7
Phenol		1.550	1.503	1.577	1.522	1.607	1.536	3.4
bis(2-Chloroethyl) ether		1.222	1.118	1.149	1.130	1.202	1.147	4.3
2-Chlorophenol		1.288	1.254	1.300	1.290	1.347	1.283	3.1
2-Methylphenol		0.978	0.950	0.995	0.992	1.043	0.984	3.6
2,2-oxybis(1-Chloropropane)		1.957	1.812	1.840	1.806	1.875	1.806	6.0
Acetophenone		0.475	0.451	0.458	0.435	0.454	0.445	4.8
3+4-Methylphenols			1.261	1.285	1.246	1.299	1.240	4.8
n-Nitroso-di-n-propylamine	0.885	0.912	0.870	0.886	0.878	0.921	0.880	3.4
Nitrobenzene-d5		0.381	0.363	0.369	0.358	0.378	0.365	3.2
Hexachloroethane		0.562	0.521	0.545	0.524	0.552	0.530	4.7
Nitrobenzene		0.338	0.313	0.326	0.320	0.335	0.324	3.1
Isophorone		0.657	0.621	0.628	0.603	0.630	0.619	3.8
2-Nitrophenol		0.172	0.174	0.183	0.181	0.192	0.181	3.9
2,4-Dimethylphenol		0.318	0.303	0.311	0.304	0.321	0.308	3.1
bis(2-Chloroethoxy) methane		0.408	0.381	0.392	0.377	0.397	0.384	4.3
2,4-Dichlorophenol		0.284	0.281	0.292	0.280	0.300	0.284	3.5
Naphthalene		1.065	0.997	1.021	0.972	1.008	0.989	5.2
4-Chloroaniline		0.429	0.389	0.399	0.399	0.408	0.397	4.7
Hexachlorobutadiene		0.210	0.204	0.206	0.197	0.205	0.200	4.4
Caprolactam			0.077	0.079	0.075	0.079	0.077	2.8
4-Chloro-3-methylphenol		0.317	0.299	0.303	0.289	0.301	0.296	4.8
2-Methylnaphthalene		0.674	0.649	0.641	0.618	0.636	0.626	5.7
Hexachlorocyclopentadiene			0.297	0.348	0.375	0.414	0.372	11.6
2,4,6-Trichlorophenol		0.349	0.373	0.383	0.373	0.405	0.375	4.8
2-Fluorobiphenyl		1.690	1.610	1.569	1.444	1.535	1.505	9.0
2,4,5-Trichlorophenol		0.404	0.404	0.413	0.400	0.431	0.405	3.6
1,1-Biphenyl		1.630	1.617	1.587	1.506	1.622	1.553	5.4
2-Chloronaphthalene		1.190	1.172	1.178	1.117	1.197	1.144	4.8
2-Nitroaniline		0.320	0.321	0.333	0.323	0.345	0.325	3.4
Dimethylphthalate		1.444	1.368	1.359	1.291	1.379	1.336	5.4
Acenaphthylene		2.053	1.986	2.003	1.881	1.991	1.929	5.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG No.: Q2198
Instrument ID: BNA_F Calibration Date(s): 06/10/2025 06/10/2025
Calibration Time(s): 16:54 20:19

LAB FILE ID:		RRF2.5 = BF142712.D			RRF005 = BF142713.D		RRF010 = BF142714.D	
		RRF020 = BF142715.D			RRF040 = BF142716.D		RRF050 = BF142717.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.307	0.283	0.295	0.285	0.299	0.289	4.5
3-Nitroaniline		0.334	0.313	0.316	0.307	0.322	0.313	4.1
Acenaphthene		1.266	1.222	1.224	1.170	1.237	1.196	4.8
2,4-Dinitrophenol			0.123	0.154	0.157	0.176	0.158	12.2
4-Nitrophenol			0.203	0.219	0.210	0.222	0.212	3.4
Dibenzofuran		1.855	1.777	1.783	1.643	1.741	1.703	6.9
2,4-Dinitrotoluene		0.396	0.383	0.399	0.377	0.396	0.382	4.7
Diethylphthalate		1.508	1.392	1.372	1.256	1.332	1.324	8.6
4-Chlorophenyl-phenylether		0.748	0.692	0.680	0.633	0.663	0.657	8.8
Fluorene		1.547	1.429	1.397	1.279	1.357	1.345	9.5
4-Nitroaniline		0.297	0.281	0.294	0.270	0.283	0.280	4.6
4,6-Dinitro-2-methylphenol			0.104	0.123	0.125	0.137	0.125	9.2
n-Nitrosodiphenylamine		0.710	0.679	0.693	0.672	0.722	0.687	3.4
2,4,6-Tribromophenol		0.236	0.223	0.224	0.213	0.226	0.219	5.3
4-Bromophenyl-phenylether		0.240	0.229	0.238	0.231	0.252	0.236	3.5
Hexachlorobenzene		0.270	0.261	0.263	0.255	0.275	0.262	3.2
Atrazine		0.185	0.174	0.188	0.179	0.191	0.183	3.3
Pentachlorophenol			0.118	0.133	0.139	0.148	0.137	7.8
Phenanthrene		1.165	1.100	1.094	1.036	1.091	1.071	5.6
Anthracene		1.203	1.145	1.140	1.064	1.132	1.108	6.0
Carbazole		1.003	0.960	0.968	0.898	0.931	0.928	6.1
Di-n-butylphthalate		1.105	1.048	1.072	1.005	1.053	1.036	4.8
Fluoranthene		1.184	1.083	1.081	0.965	0.988	1.019	10.1
Pyrene		2.014	1.918	1.928	1.883	1.878	1.859	6.7
Terphenyl-d14		1.609	1.562	1.556	1.462	1.430	1.456	9.1
Butylbenzylphthalate		0.435	0.486	0.526	0.581	0.626	0.550	12.7
3,3-Dichlorobenzidine			0.370	0.403	0.442	0.473	0.427	8.9
Benzo (a) anthracene		1.367	1.273	1.284	1.350	1.400	1.325	3.9
Chrysene		1.298	1.208	1.245	1.172	1.250	1.224	3.7
Bis (2-ethylhexyl) phthalate		0.616	0.709	0.756	0.895	0.977	0.825	16.0
Di-n-octyl phthalate			1.284	1.385	1.677	1.816	1.582	12.8
Benzo (b) fluoranthene		1.256	1.094	1.112	1.162	1.230	1.178	5.7
Benzo (k) fluoranthene		1.236	1.178	1.245	1.076	1.189	1.143	7.9
Benzo (a) pyrene		1.164	1.085	1.122	1.104	1.184	1.120	3.7
Indeno (1,2,3-cd) pyrene		1.438	1.406	1.458	1.498	1.602	1.482	4.3

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG No.: Q2198
Instrument ID: BNA_F Calibration Date(s): 06/10/2025 06/10/2025
Calibration Time(s): 16:54 20:19

LAB FILE ID:		RRF2.5 = BF142712.D		RRF005 = BF142713.D		RRF010 = BF142714.D		
		RRF020 = BF142715.D		RRF040 = BF142716.D		RRF050 = BF142717.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.134	1.176	1.214	1.236	1.318	1.210	4.9
Benzo (g,h,i) perylene		1.201	1.164	1.178	1.216	1.295	1.203	3.8
1,2,4,5-Tetrachlorobenzene		0.585	0.592	0.589	0.568	0.618	0.579	4.6
1,4-Dioxane		0.499	0.448	0.472	0.455	0.481	0.465	4.4
2,3,4,6-Tetrachlorophenol		0.364	0.339	0.357	0.327	0.354	0.340	5.6