

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB168285BS	SDG No.:	Q2198
Lab Sample ID:	PB168285BS	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142727.D	1	06/04/25 11:45	06/11/25 11:21	PB168285

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	32.2		3.90	10.0	ug/L
108-95-2	Phenol	44.1		0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	44.2		0.81	5.00	ug/L
95-57-8	2-Chlorophenol	45.9		0.58	5.00	ug/L
95-48-7	2-Methylphenol	45.4		1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	43.1		1.30	5.00	ug/L
98-86-2	Acetophenone	44.3		0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	45.6		1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	43.7		1.40	2.50	ug/L
67-72-1	Hexachloroethane	43.4		0.65	5.00	ug/L
98-95-3	Nitrobenzene	44.6		0.76	5.00	ug/L
78-59-1	Isophorone	43.8		0.75	5.00	ug/L
88-75-5	2-Nitrophenol	46.1		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	45.3		1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	44.3		0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	45.4		0.52	5.00	ug/L
91-20-3	Naphthalene	43.8		0.50	5.00	ug/L
106-47-8	4-Chloroaniline	14.3		0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	43.1		0.54	5.00	ug/L
105-60-2	Caprolactam	50.0		1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	44.9		0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	43.3		0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	90.6	E	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	46.7		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	44.3		0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	44.3		0.61	5.00	ug/L
88-74-4	2-Nitroaniline	45.3		1.30	5.00	ug/L
131-11-3	Dimethylphthalate	45.8		0.61	5.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	
Client Sample ID:	PB168285BS		SDG No.:	Q2198
Lab Sample ID:	PB168285BS		Matrix:	Water
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142727.D	1	06/04/25 11:45	06/11/25 11:21	PB168285

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	44.7		0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	46.3		0.92	5.00	ug/L
99-09-2	3-Nitroaniline	26.2		1.10	5.00	ug/L
83-32-9	Acenaphthene	50.3		0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	110	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	98.9	E	2.40	10.0	ug/L
132-64-9	Dibenzofuran	44.4		0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	48.0		1.20	5.00	ug/L
84-66-2	Diethylphthalate	46.4		0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	44.1		0.68	5.00	ug/L
86-73-7	Fluorene	43.9		0.63	5.00	ug/L
100-01-6	4-Nitroaniline	47.5		1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	48.0		2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	44.4		0.58	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	43.9		0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	44.1		0.52	5.00	ug/L
1912-24-9	Atrazine	50.6		1.00	5.00	ug/L
87-86-5	Pentachlorophenol	93.6	E	1.60	10.0	ug/L
85-01-8	Phenanthrene	44.5		0.50	5.00	ug/L
120-12-7	Anthracene	44.2		0.61	5.00	ug/L
86-74-8	Carbazole	45.7		0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	49.3		1.20	5.00	ug/L
206-44-0	Fluoranthene	45.7		0.82	5.00	ug/L
129-00-0	Pyrene	46.5		0.50	5.00	ug/L
85-68-7	Butylbenzylphthalate	50.4		1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	21.6		0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	45.7		0.45	5.00	ug/L
218-01-9	Chrysene	46.7		0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	46.3		1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	44.5		2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	48.8		0.49	5.00	ug/L

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB168285BS	SDG No.:	Q2198
Lab Sample ID:	PB168285BS	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142727.D	1	06/04/25 11:45	06/11/25 11:21	PB168285

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	44.1		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	46.7		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	45.0		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	45.1		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	44.1		0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	44.7		0.52	5.00	ug/L
123-91-1	1,4-Dioxane	36.8		1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	46.1		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	126		15 (23) - 110 (138)	84%	SPK: 150
13127-88-3	Phenol-d6	127		15 (10) - 110 (134)	85%	SPK: 150
4165-60-0	Nitrobenzene-d5	77.8		30 (67) - 130 (132)	78%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.2		30 (52) - 130 (132)	76%	SPK: 100
118-79-6	2,4,6-Tribromophenol	129		15 (44) - 110 (137)	86%	SPK: 150
1718-51-0	Terphenyl-d14	81.8		30 (42) - 130 (152)	82%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	75600	6.892			
1146-65-2	Naphthalene-d8	295000	8.175			
15067-26-2	Acenaphthene-d10	163000	9.933			
1517-22-2	Phenanthrene-d10	278000	11.427			
1719-03-5	Chrysene-d12	148000	14.068			
1520-96-3	Perylene-d12	157000	15.562			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products