

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/05/2025

Lab File ID: BF142625.D Time Analyzed: 09:37

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	122240	6.898	479356	8.18	254347	9.94
UPPER LIMIT	244480	7.398	958712	8.681	508694	10.439
LOWER LIMIT	61120	6.398	239678	7.681	127174	9.439
EPA SAMPLE NO.						
01 PB168285BL	122182	6.89	466755	8.18	253840	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.