

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/06/2025

Lab File ID: BF142640.D Time Analyzed: 11:36

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	375492	11.428	207998	14.069	270482	15.563
UPPER LIMIT	750984	11.928	415996	14.569	540964	16.063
LOWER LIMIT	187746	10.928	103999	13.569	135241	15.063
EPA SAMPLE NO.						
01 BU-703-COMP-05MS	319672	11.42	220105	14.07	274187	15.56
02 BU-703-COMP-05MSD	350996	11.43	216229	14.07	267697	15.56
03 B-202-SB02	371551	11.42	222853	14.06	271822	15.56
04 B-202-GW01	369246	11.42	237734	14.06	277909	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.