

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q2198 SDG NO.: Q2198

Lab File ID: BF142710.D

DFTPP Injection Date: 06/10/2025

Instrument ID: BNA_F

DFTPP Injection Time: 15:42

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36.0
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	33.2
70	Less than 2.0% of mass 69	0.1 (0.2) 1
127	10.0 - 80.0% of mass 198	45.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.7
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	18.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142712.D	06/10/2025	16:54
SSTDICC005	SSTDICC005	BF142713.D	06/10/2025	17:24
SSTDICC010	SSTDICC010	BF142714.D	06/10/2025	17:53
SSTDICC020	SSTDICC020	BF142715.D	06/10/2025	18:22
SSTDICCC040	SSTDICCC040	BF142716.D	06/10/2025	18:52
SSTDICC050	SSTDICC050	BF142717.D	06/10/2025	19:21
SSTDICC060	SSTDICC060	BF142718.D	06/10/2025	19:50
SSTDICC080	SSTDICC080	BF142719.D	06/10/2025	20:19