

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/06/2025

Lab File ID: BM050206.D Time Analyzed: 09:29

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	295242	7.787	1183110	10.58	689122	14.42
UPPER LIMIT	590484	8.287	2366220	11.075	1378240	14.916
LOWER LIMIT	147621	7.287	591555	10.075	344561	13.916
EPA SAMPLE NO.						
01 COMP-12MS	348615	7.79	1377430	10.58	768114	14.42
02 COMP-12MSD	346679	7.79	1352850	10.58	758249	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.