

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q2198

SDG NO.: Q2198

Lab File ID: BM050205.D

DFTPP Injection Date: 06/06/2025

Instrument ID: BNA_M

DFTPP Injection Time: 08:50

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	19.3
68	Less than 2.0% of mass 69	0.5 (1.5) 1
69	Mass 69 relative abundance	34
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	44.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12.9
442	Greater than 50% of mass 198	83.8
443	15.0 - 24.0% of mass 442	16.2 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050206.D	06/06/2025	09:29
COMP-12MS	Q2194-02MS	BM050213.D	06/06/2025	14:06
COMP-12MSD	Q2194-02MSD	BM050214.D	06/06/2025	14:46