

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2198

**Client:** Portal Partners Tri-Venture

**Analytical Method:** SW8260-Low

**Datafile :** VN086902.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|----------------------|-------|--------|------|-----|-----|------|---------|----------------|---------|
| VN0609WBSD01  | Vinyl chloride       | 20    | 18.2   | ug/L | 91  | 1   |      | 70 (65) | 130 (117)      | 20 (20) |
|               | 1,1-Dichloroethene   | 20    | 19.5   | ug/L | 98  | 3   |      | 70 (74) | 130 (110)      | 20 (20) |
|               | 2-Butanone           | 100   | 82.3   | ug/L | 82  | 20  |      | 40 (65) | 160 (122)      | 20 (20) |
|               | Carbon Tetrachloride | 20    | 19.4   | ug/L | 97  | 0   |      | 70 (77) | 130 (113)      | 20 (20) |
|               | Chloroform           | 20    | 18.4   | ug/L | 92  | 8   |      | 70 (79) | 130 (113)      | 20 (20) |
|               | Benzene              | 20    | 19.0   | ug/L | 95  | 6   |      | 70 (82) | 130 (109)      | 20 (20) |
|               | 1,2-Dichloroethane   | 20    | 18.8   | ug/L | 94  | 13  |      | 70 (80) | 130 (115)      | 20 (20) |
|               | Trichloroethene      | 20    | 20.4   | ug/L | 102 | 2   |      | 70 (77) | 130 (113)      | 20 (20) |
|               | Tetrachloroethene    | 20    | 19.2   | ug/L | 96  | 2   |      | 70 (67) | 130 (123)      | 20 (20) |
|               | Chlorobenzene        | 20    | 20.3   | ug/L | 102 | 2   |      | 70 (82) | 130 (109)      | 20 (20) |