

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134124
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046487.D	04 Jun 2025 09:43		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046488.D	04 Jun 2025 10:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0604MBL01	VX0604MBL01	VX046489.D	04 Jun 2025 10:40		JC/MD	Ok
4	VX0604WBL01	VX0604WBL01	VX046490.D	04 Jun 2025 11:04		JC/MD	Ok
5	VX0604WBS01	VX0604WBS01	VX046491.D	04 Jun 2025 11:27	BS failed low for comp. #17	JC/MD	Ok,M
6	Q2169-03DL	303-PPR-2DL	VX046492.D	04 Jun 2025 11:55	vial B pH<2	JC/MD	Ok
7	Q2168-08DL	B3DL	VX046493.D	04 Jun 2025 12:18	vial B pH#5.0	JC/MD	Ok
8	Q2168-12DL	C2DL	VX046494.D	04 Jun 2025 12:41	vial B pH#5.0	JC/MD	Ok
9	Q2169-01	303-PPR-1	VX046495.D	04 Jun 2025 13:05	vial B pH<2;not req	JC/MD	Not Ok
10	Q2175-05	52525-B	VX046496.D	04 Jun 2025 13:28	vial A pH<2 foamy sample	JC/MD	Ok
11	VX0604WBSD01	VX0604WBSD01	VX046497.D	04 Jun 2025 13:52		JC/MD	Ok,M
12	Q2200-01DL	RMW-02B-66-060325D	VX046498.D	04 Jun 2025 14:15	vial A pH<2	JC/MD	Ok
13	Q2200-02	RMW-03B-90-060325	VX046499.D	04 Jun 2025 14:39	vial A pH<2	JC/MD	Ok
14	Q2200-05	MW-11B-37.5-060325	VX046500.D	04 Jun 2025 15:02	vial A pH<2 Need 200X	JC/MD	Dilution
15	Q2175-06	EGR-LIQUID	VX046501.D	04 Jun 2025 15:26	vial A pH<2 Need 2000X	JC/MD	Dilution
16	IBLK	IBLK	VX046502.D	04 Jun 2025 15:50		JC/MD	Ok
17	IBLK	IBLK	VX046503.D	04 Jun 2025 16:13		JC/MD	Ok
18	Q2200-03	EB01-060325	VX046504.D	04 Jun 2025 16:37	vial A pH<2 EB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Maresh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

19	Q2201-01	MW-01-6.5-060325	VX046505.D	04 Jun 2025 17:01	vial A pH<2	JC/MD	Ok
20	Q2198-05	B-202-GW01	VX046506.D	04 Jun 2025 17:25	vial A pH<2	JC/MD	Ok
21	IBLK	IBLK	VX046507.D	04 Jun 2025 17:49		JC/MD	Ok
22	Q2200-01	RMW-02B-66-060325	VX046508.D	04 Jun 2025 18:12	vial B pH<2 Need 100X	JC/MD	Dilution
23	Q2200-05DL	MW-11B-37.5-060325	VX046509.D	04 Jun 2025 18:36	vial B pH<2	JC/MD	Ok
24	Q2200-06	TB-01-060325	VX046510.D	04 Jun 2025 19:00	vial A pH<2 TB	JC/MD	Ok
25	Q2175-06DL	EGR-LIQUIDDL	VX046511.D	04 Jun 2025 19:24	vial B pH<2	JC/MD	Ok,M
26	VX0604MBS01	VX0604MBS01	VX046512.D	04 Jun 2025 19:47		JC/MD	Ok,M
27	Q2168-11ME	C2ME	VX046513.D	04 Jun 2025 20:11	Need 5X	JC/MD	Dilution
28	Q2168-07ME	B3ME	VX046514.D	04 Jun 2025 20:35		JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX046515.D	04 Jun 2025 20:59		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y
		HP Processing Method	82Y060225S.M

STD. NAME	STD REF.#
Tune/Reschk	VP134084
Initial Calibration Stds	VP134085,VP134086,VP134087,VP134088,VP134089,VP134090
CCC	VP134092,VP134093
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134091
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022488.D	02 Jun 2025 08:31		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022489.D	02 Jun 2025 09:08	Need ICAL	SY/MD	Not Ok
3	VY0602SBL01	VY0602SBL01	VY022490.D	02 Jun 2025 10:38		SY/MD	Not Ok
4	VSTDICC005	VSTDICC005	VY022491.D	02 Jun 2025 11:46		SY/MD	Ok,M
5	VSTDICC010	VSTDICC010	VY022492.D	02 Jun 2025 12:09		SY/MD	Ok,M
6	VSTDICC020	VSTDICC020	VY022493.D	02 Jun 2025 12:32		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY022494.D	02 Jun 2025 12:54		SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY022495.D	02 Jun 2025 13:17		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY022496.D	02 Jun 2025 13:39		SY/MD	Ok,M
10	VIBLK	VIBLK	VY022497.D	02 Jun 2025 14:03		SY/MD	Ok
11	VSTDICV050	ICVVY060225	VY022498.D	02 Jun 2025 14:59	Icv Fail For DOD For Com.# 18	SY/MD	Ok,M
12	VY0602SBL02	VY0602SBL02	VY022499.D	02 Jun 2025 16:00		SY/MD	Ok
13	VY0602SBS01	VY0602SBS01	VY022500.D	02 Jun 2025 16:24		SY/MD	Ok,M
14	VY0602SBSD01	VY0602SBSD01	VY022501.D	02 Jun 2025 16:47		SY/MD	Ok,M
15	Q2160-02	TP04-MHG-VOC	VY022502.D	02 Jun 2025 17:10	vial A	SY/MD	Ok
16	Q2152-01	OK-02-05292025	VY022503.D	02 Jun 2025 17:34	Not Purged,vial B	SY/MD	Not Ok
17	Q2153-01RE	TR-04-0592025RE	VY022504.D	02 Jun 2025 17:57	Surrogate Fail;Internal Standard Fail, vial B	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M
STD. NAME		STD REF.#			
Tune/Reschk		VP134084			
Initial Calibration Stds		VP134085,VP134086,VP134087,VP134088,VP134089,VP134090			
CCC		VP134092,VP134093			
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134091			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2147-02	VOC	VY022505.D	02 Jun 2025 18:21	Not purge,vial B	SY/MD	Not Ok
19	Q2150-02	TP-42	VY022506.D	02 Jun 2025 18:44	vial B	SY/MD	Ok
20	Q2150-05RE	TP-47RE	VY022507.D	02 Jun 2025 19:08	Internal Standard Fail,vial B	SY/MD	Confirms
21	Q2161-01	B27-SOIL-SAMPLE	VY022508.D	02 Jun 2025 19:31	vial-B,Internal Standard Fail; Surrogate Fail	SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022509.D	02 Jun 2025 19:54		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060425

Review By	Maresh Dadoda	Review On	6/5/2025 4:43:35 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:47:03 PM
SubDirectory	VY060425	HP Acquire Method	MSVOA_Y HP Processing Method 82y060225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134109
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134110,VP134111,MDL VP134112,VP134113 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022537.D	04 Jun 2025 08:48		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022538.D	04 Jun 2025 09:54		SY/MD	Ok,M
3	VY0604SBL01	VY0604SBL01	VY022539.D	04 Jun 2025 10:24		SY/MD	Ok
4	VY0604SBS01	VY0604SBS01	VY022540.D	04 Jun 2025 10:54		SY/MD	Ok,M
5	VY0604SBSD01	VY0604SBSD01	VY022541.D	04 Jun 2025 11:16		SY/MD	Ok,M
6	Q2126-03	MDL-SOIL-03-QT2-202	VY022542.D	04 Jun 2025 11:53		SY/MD	Ok,M
7	Q2126-03	MDL-SOIL-03-QT2-202	VY022543.D	04 Jun 2025 12:16		SY/MD	Ok,M
8	Q2168-03RE	A3RE	VY022544.D	04 Jun 2025 12:39	vial-B Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
9	Q2182-01RE	OR-03-06022025RE	VY022545.D	04 Jun 2025 13:03	vial-B Internal Standard Fail	SY/MD	Confirms
10	Q2195-01	OK-01-060325	VY022546.D	04 Jun 2025 13:26	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q2194-01	COMP-12	VY022547.D	04 Jun 2025 13:50	vial-A	SY/MD	Ok
12	Q2194-03	COMP-13	VY022548.D	04 Jun 2025 14:13	vial-A Internal Standard Fail	SY/MD	ReRun
13	Q2199-02	ETGI-343	VY022549.D	04 Jun 2025 14:37	vial-A Internal Standard Fail	SY/MD	ReRun
14	Q2199-04	VNJ-231	VY022550.D	04 Jun 2025 15:00	vial-A Not purge	SY/MD	Not Ok
15	Q2199-06	72-11978	VY022551.D	04 Jun 2025 15:24	vial-A	SY/MD	Ok
16	Q2198-01	B-202-SB02	VY022552.D	04 Jun 2025 15:47	vial-A	SY/MD	Ok
17	Q2198-03	B-207-SB02	VY022553.D	04 Jun 2025 16:11	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:43:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:47:03 PM		
SubDirectory	VY060425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y060225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134109			
CCC		VP134110,VP134111,MDL VP134112,VP134113			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2177-02	B-187-SB01	VY022554.D	04 Jun 2025 16:34	vial-A	SY/MD	Ok
19	Q2177-04	B-187-SB02	VY022555.D	04 Jun 2025 16:58	vial-A	SY/MD	Ok
20	Q2177-06	B-202-SB01	VY022556.D	04 Jun 2025 17:21	vial-A Internal Standard Fail	SY/MD	ReRun
21	VIBLK	VIBLK	VY022557.D	04 Jun 2025 17:45		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022558.D	04 Jun 2025 18:30		SY/MD	Ok,M

M : Manual Integration