

Surrogate Summary

SDG No.: **Q2198**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2198-01	B-202-SB02	1,2-Dichloroethane-d4	50	44.8	90	70 (63)	130 (155)
		Dibromofluoromethane	50	48.7	97	70 (70)	130 (134)
		Toluene-d8	50	49.1	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.0	80	70 (17)	130 (146)
Q2198-03	B-207-SB02	1,2-Dichloroethane-d4	50	51.4	103	70 (63)	130 (155)
		Dibromofluoromethane	50	50.7	101	70 (70)	130 (134)
		Toluene-d8	50	49.5	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.6	75	70 (17)	130 (146)
VY0604SBL01	VY0604SBL01	1,2-Dichloroethane-d4	50	52.7	105	70 (63)	130 (155)
		Dibromofluoromethane	50	50.7	101	70 (70)	130 (134)
		Toluene-d8	50	49.3	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.7	83	70 (17)	130 (146)
VY0604SBS01	VY0604SBS01	1,2-Dichloroethane-d4	50	53.9	108	70 (63)	130 (155)
		Dibromofluoromethane	50	53.3	107	70 (70)	130 (134)
		Toluene-d8	50	53.4	107	70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.0	104	70 (17)	130 (146)
VY0604SBSD01	VY0604SBSD01	1,2-Dichloroethane-d4	50	52.1	104	70 (63)	130 (155)
		Dibromofluoromethane	50	50.5	101	70 (70)	130 (134)
		Toluene-d8	50	50.4	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.0	98	70 (17)	130 (146)

Surrogate Summary

SDG No.: **Q2198**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2198-05	B-202-GW01	1,2-Dichloroethane-d4	50	52.2	104	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	50.2	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.3	105	70 (77)	130 (121)
VX0604WBL01	VX0604WBL01	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.1	106	70 (77)	130 (121)
VX0604WBS01	VX0604WBS01	1,2-Dichloroethane-d4	50	50.1	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103	70 (75)	130 (124)
		Toluene-d8	50	47.3	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98	70 (77)	130 (121)
VX0604WBSD0	VX0604WBSD01	1,2-Dichloroethane-d4	50	51.7	103	70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104	70 (75)	130 (124)
		Toluene-d8	50	49.0	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)

