

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2198 SAS No.: Q2198 SDG NO.: Q2198
 Lab File ID: VY022538.D Date Analyzed: 06/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	125910	13.346				
UPPER LIMIT	251820	13.846				
LOWER LIMIT	62955	12.846				
EPA SAMPLE NO.						
B-202-SB02	132509	13.34				
B-207-SB02	101553	13.34				
VY0604SBL01	137326	13.35				
VY0604SBS01	119650	13.35				
VY0604SBSD01	119455	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.