

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2200 SAS No.: Q2200 SDG No.: Q2200

Instrument ID: BNA_N Calibration Date/Time: 06/05/2025 09:42

Lab File ID: BN037172.D Init. Calib. Date(s): 06/03/2025 06/03/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:39 15:14

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.557	0.553		-0.7	20.0
Fluoranthene-d10	1.016	0.921		-9.4	20.0
2-Fluorophenol	0.989	0.924		-6.6	20.0
Phenol-d6	1.199	1.117		-6.8	20.0
Nitrobenzene-d5	0.422	0.422		0.0	20.0
2-Fluorobiphenyl	1.705	1.677		-1.6	20.0
2,4,6-Tribromophenol	0.161	0.143		-11.2	20.0
Terphenyl-d14	0.942	0.898		-4.7	20.0
1,4-Dioxane	0.533	0.502		-5.8	20.0

All other compounds must meet a minimum RRF of 0.010.