

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q2200 SAS No.: Q2200 SDG No.: Q2200

Instrument ID: MSVOA_X Calibration Date/Time: 06/04/2025 10:12

Lab File ID: VX046488.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.665		-3.76	20
1,1-Dichloroethene	0.593	0.593		0	20
1,1-Dichloroethane	1.219	1.302	0.1	6.81	20
cis-1,2-Dichloroethene	0.718	0.747		4.04	20
1,1,1-Trichloroethane	1.101	1.161		5.45	20
Benzene	1.417	1.496		5.57	20
1,2-Dichloroethane	0.612	0.651		6.37	20
Trichloroethene	0.341	0.364		6.74	20
1,1,2-Trichloroethane	0.343	0.370		7.87	20
Tetrachloroethene	0.354	0.378		6.78	20
1,2-Dichloroethane-d4	0.932	0.862		-7.51	20
Dibromofluoromethane	0.360	0.356		-1.11	20
Toluene-d8	1.246	1.135		-8.91	20
4-Bromofluorobenzene	0.478	0.469		-1.88	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.