

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133811 VP133832,VP133833,VP133834,VP133835,VP133836,VP133837 VP133838

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134124
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046487.D	04 Jun 2025 09:43		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046488.D	04 Jun 2025 10:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0604MBL01	VX0604MBL01	VX046489.D	04 Jun 2025 10:40		JC/MD	Ok
4	VX0604WBL01	VX0604WBL01	VX046490.D	04 Jun 2025 11:04		JC/MD	Ok
5	VX0604WBS01	VX0604WBS01	VX046491.D	04 Jun 2025 11:27	BS failed low for comp. #17	JC/MD	Ok,M
6	Q2169-03DL	303-PPR-2DL	VX046492.D	04 Jun 2025 11:55	vial B pH<2	JC/MD	Ok
7	Q2168-08DL	B3DL	VX046493.D	04 Jun 2025 12:18	vial B pH#5.0	JC/MD	Ok
8	Q2168-12DL	C2DL	VX046494.D	04 Jun 2025 12:41	vial B pH#5.0	JC/MD	Ok
9	Q2169-01	303-PPR-1	VX046495.D	04 Jun 2025 13:05	vial B pH<2;Need 2X	JC/MD	Dilution
10	Q2175-05	52525-B	VX046496.D	04 Jun 2025 13:28	vial A pH<2 foamy sample	JC/MD	Ok
11	VX0604WBSD01	VX0604WBSD01	VX046497.D	04 Jun 2025 13:52		JC/MD	Ok,M
12	Q2200-01DL	RMW-02B-66-060325D	VX046498.D	04 Jun 2025 14:15		JC/MD	Ok
13	Q2200-02	RMW-03B-90-060325	VX046499.D	04 Jun 2025 14:39		JC/MD	Ok
14	Q2200-05	MW-11B-37.5-060325	VX046500.D	04 Jun 2025 15:02	Need 200X	JC/MD	Dilution
15	Q2175-06	EGR-LIQUID	VX046501.D	04 Jun 2025 15:26	Need 2000X	JC/MD	Dilution
16	IBLK	IBLK	VX046502.D	04 Jun 2025 15:50		JC/MD	Ok
17	IBLK	IBLK	VX046503.D	04 Jun 2025 16:13		JC/MD	Ok
18	Q2200-03	EB01-060325	VX046504.D	04 Jun 2025 16:37	EB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Maresh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

19	Q2201-01	MW-01-6.5-060225	VX046505.D	04 Jun 2025 17:01	vial A pH<2	JC/MD	Ok
20	Q2198-05	B-202-GW01	VX046506.D	04 Jun 2025 17:25	BS failed low for comp. #17	JC/MD	ReRun
21	IBLK	IBLK	VX046507.D	04 Jun 2025 17:49		JC/MD	Ok
22	Q2200-01	RMW-02B-66-060325	VX046508.D	04 Jun 2025 18:12	Need 100X	JC/MD	Dilution
23	Q2200-05DL	MW-11B-37.5-060325D	VX046509.D	04 Jun 2025 18:36		JC/MD	Ok
24	Q2200-06	TB-01-060325	VX046510.D	04 Jun 2025 19:00	TB	JC/MD	Ok
25	Q2175-06DL	EGR-LIQUIDDL	VX046511.D	04 Jun 2025 19:24		JC/MD	Ok,M
26	VX0604MBS01	VX0604MBS01	VX046512.D	04 Jun 2025 19:47		JC/MD	Ok,M
27	Q2168-11ME	C2ME	VX046513.D	04 Jun 2025 20:11	Need 10X	JC/MD	Dilution
28	Q2168-07ME	B3ME	VX046514.D	04 Jun 2025 20:35		JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX046515.D	04 Jun 2025 20:59		JC/MD	Ok,M

M : Manual Integration