

Surrogate Summary

SDG No.: **Q2200**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2200-01	RMW-02B-66-060325	1,2-Dichloroethane-d4	50	51.6	103	70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.8	98	70 (77)	130 (121)
Q2200-01DL	RMW-02B-66-060325DL	1,2-Dichloroethane-d4	50	51.3	103	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.4	107	70 (77)	130 (121)
Q2200-02	RMW-03B-90-060325	1,2-Dichloroethane-d4	50	51.0	102	70 (74)	130 (125)
		Dibromofluoromethane	50	50.2	100	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.6	101	70 (77)	130 (121)
Q2200-03	EB01-060325	1,2-Dichloroethane-d4	50	52.3	105	70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.7	105	70 (77)	130 (121)
Q2200-05	MW-11B-37.5-060325	1,2-Dichloroethane-d4	50	52.1	104	70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100	70 (75)	130 (124)
		Toluene-d8	50	49.8	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.6	101	70 (77)	130 (121)
Q2200-05DL	MW-11B-37.5-060325DL	1,2-Dichloroethane-d4	50	52.7	105	70 (74)	130 (125)
		Dibromofluoromethane	50	49.9	100	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.3	101	70 (77)	130 (121)
Q2200-06	TB-01-060325	1,2-Dichloroethane-d4	50	53.2	106	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.7	101	70 (77)	130 (121)
VX0604WBL01	VX0604WBL01	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.1	106	70 (77)	130 (121)
VX0604WBS01	VX0604WBS01	1,2-Dichloroethane-d4	50	50.1	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103	70 (75)	130 (124)
		Toluene-d8	50	47.3	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98	70 (77)	130 (121)
VX0604WBSD0	VX0604WBSD01	1,2-Dichloroethane-d4	50	51.7	103	70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104	70 (75)	130 (124)
		Toluene-d8	50	49.0	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)

