

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2200

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

**Datafile :** VX046497.D

| Lab Sample ID | Parameter              | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|------------------------|-------|--------|------|-----|-----|------|---------|----------------|---------|
| VX0604WBSD01  | Vinyl chloride         | 20    | 18.8   | ug/L | 94  | 19  |      | 70 (65) | 130 (117)      | 20 (20) |
|               | 1,1-Dichloroethene     | 20    | 20.3   | ug/L | 102 | 17  |      | 70 (74) | 130 (110)      | 20 (20) |
|               | 1,1-Dichloroethane     | 20    | 22.2   | ug/L | 111 | 12  |      | 70 (78) | 130 (112)      | 20 (20) |
|               | cis-1,2-Dichloroethene | 20    | 22.0   | ug/L | 110 | 12  |      | 70 (77) | 130 (110)      | 20 (20) |
|               | 1,1,1-Trichloroethane  | 20    | 22.5   | ug/L | 113 | 15  |      | 70 (80) | 130 (108)      | 20 (20) |
|               | Benzene                | 20    | 21.4   | ug/L | 107 | 14  |      | 70 (82) | 130 (109)      | 20 (20) |
|               | 1,2-Dichloroethane     | 20    | 21.6   | ug/L | 108 | 9   |      | 70 (80) | 130 (115)      | 20 (20) |
|               | Trichloroethene        | 20    | 21.3   | ug/L | 106 | 14  |      | 70 (77) | 130 (113)      | 20 (20) |
|               | 1,1,2-Trichloroethane  | 20    | 22.2   | ug/L | 111 | 7   |      | 70 (83) | 130 (112)      | 20 (20) |
|               | Tetrachloroethene      | 20    | 20.6   | ug/L | 103 | 8   |      | 70 (67) | 130 (123)      | 20 (20) |