

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2200 SAS No.: Q2200 SDG NO.: Q2200
Lab File ID: VX046488.D Date Analyzed: 06/04/2025
Instrument ID: MSVOA_X Time Analyzed: 10:12
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97475	5.54	165033	6.75	141151	10.05
UPPER LIMIT	194950	6.043	330066	7.25	282302	10.549
LOWER LIMIT	48737.5	5.043	82516.5	6.25	70575.5	9.549
EPA SAMPLE NO.						
RMW-02B-66-060325	59475	5.55	118649	6.76	107657	10.06
RMW-02B-66-060325DL	62989	5.55	125027	6.76	118665	10.06
RMW-03B-90-060325	65911	5.55	132066	6.76	121549	10.06
EB01-060325	64028	5.54	128366	6.76	122354	10.06
MW-11B-37.5-060325	61191	5.55	121976	6.76	112952	10.06
MW-11B-37.5-060325DL	63962	5.55	127435	6.76	117554	10.06
TB-01-060325	58762	5.54	118020	6.76	111037	10.06
VX0604WBL01	69580	5.55	139946	6.76	133992	10.05
VX0604WBS01	92897	5.54	164481	6.76	139452	10.05
VX0604WBSD01	84483	5.55	152834	6.76	133225	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.