

DATA PACKAGE VOLATILE ORGANICS

PROJECT NAME : FORMER SCHLUMBERGER STC PTC SITE D3868221

JACOBS ENGINEERING GROUP, INC.

412 Mt. Kemble Ave

Downtown Building

Morristown, NJ - 07960

Phone No: 9732670555

ORDER ID : Q2201

ATTENTION : John Ynfante



Laboratory Certification ID # 20012



1) VOLATILES DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	6
5) Conformance/Non Conformance	7
6) QA Checklist	9
7) Chronicle	10
8) Hit Summary	11
9) QC Data Summary For VOCMS Group3	12
9.1) Deuterated Monitoring Compound Summary	13
9.2) LCS/LCSD Summary	14
9.3) Method Blank Summary	16
9.4) GS/MS Tune Summary	17
9.5) Internal Standard Area and RT Summary	19
10) Sample Data	21
10.1) MW-01-6.5-060325	22
11) Calibration Data Summary	31
11.1) Initial Calibration Data	32
11.1.1) VX050525	32
11.2) Continued Calibration Data	401
11.2.1) VX046488.D	401
12) QC Sample Data	460
12.1) Tune Raw Data	461
12.2) Method Blank Data	463
12.3) LCS Data	470
12.4) LCSD Data	523
13) Manual Integration	576
14) Analytical Runlogs	579
15) Standard Prep Logs	585
16) Shipping Document	680
16.1) Chain Of Custody	681
16.2) Lab Certificate	682
16.3) Internal COC	683

Cover Page

Order ID : Q2201

Project ID : Former Schlumberger STC PTC Site D3868221

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

Q2201-01

Client Sample Number

MW-01-6.5-060325

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/9/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger STC PTC Site D3868221

Project # N/A

Order ID # Q2201

Test Name: VOCMS Group3

A. Number of Samples and Date of Receipt:

1 Water sample was received on 06/03/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group3. This data package contains results for VOCMS Group3.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group3 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2201

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples. The Blank Spike Duplicate met requirements for all samples.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2201

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/09/2025

LAB CHRONICLE

OrderID:	Q2201	OrderDate:	6/3/2025 4:10:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	Location:	VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2201-01	MW-01-6.5-060325	Water	VOCMS Group3	8260-Low	06/03/25		06/04/25	06/03/25

Hit Summary Sheet
SW-846

SDG No.: Q2201
Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	MW-01-6.5-060325							
Q2201-01	MW-01-6.5-060325	Water	cis-1,2-Dichloroethene	4.40		0.19	1.00	ug/L
Q2201-01	MW-01-6.5-060325	Water	Trichloroethene	34.7		0.090	1.00	ug/L
			Total Voc :		39.1			
			Total Concentration:		39.1			



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Surrogate Summary

SDG No.: **Q2201**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2201-01	MW-01-6.5-060325	1,2-Dichloroethane-d4	50	51.5	103	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
VX0604WBL01	VX0604WBL01	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.1	106	70 (77)	130 (121)
VX0604WBS01	VX0604WBS01	1,2-Dichloroethane-d4	50	50.1	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103	70 (75)	130 (124)
		Toluene-d8	50	47.3	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98	70 (77)	130 (121)
VX0604WBSD0	VX0604WBSD01	1,2-Dichloroethane-d4	50	51.7	103	70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104	70 (75)	130 (124)
		Toluene-d8	50	49.0	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2201

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX046491.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0604WBS01	Vinyl chloride	20	15.5	ug/L	78			70 (65)	130 (117)	
	1,1-Dichloroethene	20	17.2	ug/L	86			70 (74)	130 (110)	
	1,1-Dichloroethane	20	19.6	ug/L	98			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	19.6	ug/L	98			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	19.4	ug/L	97			70 (80)	130 (108)	
	Benzene	20	18.6	ug/L	93			70 (82)	130 (109)	
	1,2-Dichloroethane	20	19.8	ug/L	99			70 (80)	130 (115)	
	Trichloroethene	20	18.4	ug/L	92			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	20.5	ug/L	103			70 (83)	130 (112)	
	Tetrachloroethene	20	19.0	ug/L	95			70 (67)	130 (123)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2201

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX046497.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0604WBSD01	Vinyl chloride	20	18.8	ug/L	94	19		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	20.3	ug/L	102	17		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	22.2	ug/L	111	12		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	22.0	ug/L	110	12		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	22.5	ug/L	113	15		70 (80)	130 (108)	20 (20)
	Benzene	20	21.4	ug/L	107	14		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.6	ug/L	108	9		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	21.3	ug/L	106	14		70 (77)	130 (113)	20 (20)
	1,1,2-Trichloroethane	20	22.2	ug/L	111	7		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	20.6	ug/L	103	8		70 (67)	130 (123)	20 (20)

() = LABORATORY INHOUSE LIMIT

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0604WBL01

Lab Name: CHEMTECH

Contract: JACO05

Lab Code: CHEM Case No.: Q2201

SAS No.: Q2201 SDG NO.: Q2201

Lab File ID: VX046490.D

Lab Sample ID: VX0604WBL01

Date Analyzed: 06/04/2025

Time Analyzed: 11:04

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0604WBS01	VX0604WBS01	VX046491.D	06/04/2025
VX0604WBSD01	VX0604WBSD01	VX046497.D	06/04/2025
MW-01-6.5-060325	Q2201-01	VX046505.D	06/04/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>JACO05</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2201</u>
Lab File ID:	<u>VX046038.D</u>	SAS No.:	<u>Q2201</u>
Instrument ID:	<u>MSVOA_X</u>	SDG NO.:	<u>Q2201</u>
GC Column:	<u>DB-624UI</u>	BFB Injection Date:	<u>05/05/2025</u>
ID:	<u>0.18</u>	BFB Injection Time:	<u>09:37</u>
(mm)		Heated Purge:	<u>Y/N</u>
			<u>N</u>

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2201 SAS No.: Q2201 SDG NO.: Q2201
 Lab File ID: VX046487.D BFB Injection Date: 06/04/2025
 Instrument ID: MSVOA_X BFB Injection Time: 09:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22
75	30.0 - 60.0% of mass 95	55.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	68.4
175	5.0 - 9.0% of mass 174	4.8 (7) 1
176	95.0 - 101.0% of mass 174	67.2 (98.3) 1
177	5.0 - 9.0% of mass 176	4.5 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046488.D	06/04/2025	10:12
VX0604WBL01	VX0604WBL01	VX046490.D	06/04/2025	11:04
VX0604WBS01	VX0604WBS01	VX046491.D	06/04/2025	11:27
VX0604WBSD01	VX0604WBSD01	VX046497.D	06/04/2025	13:52
MW-01-6.5-060325	Q2201-01	VX046505.D	06/04/2025	17:01

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2201 SAS No.: Q2201 SDG NO.: Q2201
Lab File ID: VX046488.D Date Analyzed: 06/04/2025
Instrument ID: MSVOA_X Time Analyzed: 10:12
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97475	5.54	165033	6.75	141151	10.05
UPPER LIMIT	194950	6.043	330066	7.25	282302	10.549
LOWER LIMIT	48737.5	5.043	82516.5	6.25	70575.5	9.549
EPA SAMPLE NO.						
MW-01-6.5-060325	62450	5.55	122833	6.76	114448	10.06
VX0604WBL01	69580	5.55	139946	6.76	133992	10.05
VX0604WBS01	92897	5.54	164481	6.76	139452	10.05
VX0604WBSD01	84483	5.55	152834	6.76	133225	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2201 SAS No.: Q2201 SDG NO.: Q2201
Lab File ID: VX046488.D Date Analyzed: 06/04/2025
Instrument ID: MSVOA_X Time Analyzed: 10:12
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69016	12.018				
UPPER LIMIT	138032	12.518				
LOWER LIMIT	34508	11.518				
EPA SAMPLE NO.						
MW-01-6.5-060325	48812	12.02				
VX0604WBL01	59967	12.02				
VX0604WBS01	63937	12.02				
VX0604WBSD01	62838	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/03/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/03/25
Client Sample ID:	MW-01-6.5-060325	SDG No.:	Q2201
Lab Sample ID:	Q2201-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046505.D	1		06/04/25 17:01	VX060425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	4.40		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	34.7		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.5		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.0		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62500	5.55			
540-36-3	1,4-Difluorobenzene	123000	6.763			
3114-55-4	Chlorobenzene-d5	114000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	48800	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046505.D
Acq On : 04 Jun 2025 17:01
Operator : JC/MD
Sample : Q2201-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-060325

Quant Time: Jun 05 02:02:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration


Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	62450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	122833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59937	51.480	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.960%		
35) Dibromofluoromethane	5.385	113	44599	50.421	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.840%		
50) Toluene-d8	8.647	98	153105	50.010	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.020%		
62) 4-Bromofluorobenzene	11.079	95	59280	50.480	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.960%		
Target Compounds						Qvalue
16) Acetone	2.380	43	2867	6.142	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	3959	4.416	ug/l	87
44) Trichloroethene	7.123	130	29049	34.672	ug/l	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046505.D
Acq On : 04 Jun 2025 17:01
Operator : JC/MD
Sample : Q2201-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-060325

Quant Time: Jun 05 02:02:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01




Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-060325


Tgt Ion:168 Resp: 62450
Ion Ratio Lower Upper
168 100
99 72.2 54.9 82.3



#16
Acetone
Concen: 6.142 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01

Tgt Ion: 43 Resp: 2867
Ion Ratio Lower Upper
43 100
58 24.1 21.2 31.8





#27

cis-1,2-Dichloroethene

Concen: 4.416 ug/l

RT: 4.489 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046505.D

Acq: 04 Jun 2025 17:01

Instrument :

MSVOA_X

ClientSampleId :

MW-01-6.5-060325





Tgt Ion: 96 Resp: 3959

Ion Ratio Lower Upper

96 100

61 138.4 0.0 322.8

98 62.8 0.0 129.0



#33

1,2-Dichloroethane-d4

Concen: 51.480 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046505.D




Acq: 04 Jun 2025 17:01


Tgt Ion: 65 Resp: 59937

Ion Ratio Lower Upper

65 100

67 50.0 0.0 99.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. 0.006 min

Lab File: VX046505.D

Acq: 04 Jun 2025 17:01

Instrument :

MSVOA_X

ClientSampleId :

MW-01-6.5-060325





Tgt Ion:114 Resp: 122833

Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 17.7 0.0 33.6



#35

Dibromofluoromethane

Concen: 50.421 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.006 min

Lab File: VX046505.D

Acq: 04 Jun 2025 17:01




Tgt Ion:113 Resp: 44599


Ion Ratio Lower Upper

113 100

111 102.3 83.1 124.7

192 16.6 13.3 19.9









#44
Trichloroethene
Concen: 34.672 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01




Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-060325


Tgt Ion:130 Resp: 29049
Ion Ratio Lower Upper
130 100
95 112.4 0.0 204.2



#50
Toluene-d8
Concen: 50.010 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01

Tgt Ion: 98 Resp: 153105
Ion Ratio Lower Upper
98 100
100 65.4 53.5 80.3









#62
4-Bromofluorobenzene
Concen: 50.480 ug/l
RT: 11.079 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01




Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-060325


Tgt Ion: 95 Resp: 59280
Ion Ratio Lower Upper
95 100
174 68.6 0.0 135.8
176 66.2 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046505.D
Acq: 04 Jun 2025 17:01

Tgt Ion: 117 Resp: 114448
Ion Ratio Lower Upper
117 100
82 60.4 50.6 76.0
119 32.1 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046505.D


Acq: 04 Jun 2025 17:01

Instrument :

MSVOA_X

ClientSampleId :

MW-01-6.5-060325





Tgt Ion:152 Resp: 48812

Ion Ratio Lower Upper

152 100

115 66.0 46.9 140.7

150 159.2 0.0 351.0





CALIBRATION SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2201 SAS No.: Q2201 SDG No.: Q2201
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2	
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9	
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6	
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5	
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6	
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3	
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5	
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3	
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M
 Title : SW846 8260
 Last Update : Tue May 06 07:12:22 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046047.D 5 =VX046046.D 20 =VX046041.D 50 =VX046042.D 100 =VX046043.D 150 =VX046044.D

Compound		1	5	20	50	100	150	Avg	%RSD

1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.658	0.639	0.697	0.864	0.859	0.875	0.765	14.62
3) P	Chloromethane	0.694	0.679	0.727	0.775	0.787	0.791	0.742	6.58
4) C	Vinyl Chloride	0.673	0.619	0.660	0.710	0.727	0.755	0.691	7.17#
5) T	Bromomethane		0.305	0.296	0.326	0.340	0.334	0.320	5.84
6) T	Chloroethane	0.467	0.368	0.354	0.378	0.329	0.317	0.369	14.44
7) T	Trichlorofluor...	1.064	0.990	1.035	1.068	0.983	0.985	1.021	3.92
8) T	Diethyl Ether	0.403	0.311	0.340	0.337	0.338	0.355	0.347	8.83
9) T	1,1,2-Trichlor...	0.633	0.610	0.628	0.641	0.629	0.648	0.632	2.10
10) T	Methyl Iodide		0.608	0.767	0.806	0.793	0.763	0.747	10.68
11) T	Tert butyl alc...		0.114	0.122	0.129	0.144	0.146	0.131	10.45
12) CM	1,1-Dichloroet...	0.594	0.567	0.565	0.601	0.607	0.625	0.593	3.94#
13) T	Acrolein		0.117	0.158	0.152	0.154	0.163	0.149	12.33
14) T	Allyl chloride	1.052	1.058	1.127	1.179	1.187	1.196	1.133	5.75
15) T	Acrylonitrile	0.363	0.345	0.378	0.388	0.381	0.390	0.374	4.52
16) T	Acetone	0.380	0.408	0.361	0.362	0.361	0.370	0.374	4.90
17) T	Carbon Disulfide	1.423	1.141	1.295	1.455	1.522	1.597	1.406	11.68
18) T	Methyl Acetate	1.006	0.816	0.814	0.848	0.845	0.875	0.867	8.27
19) T	Methyl tert-bu...	1.949	1.908	2.044	2.160	2.172	2.239	2.079	6.39
20) T	Methylene Chlo...	0.853	0.689	0.689	0.684	0.691	0.691	0.716	9.39
21) T	trans-1,2-Dich...	0.604	0.557	0.573	0.610	0.612	0.622	0.596	4.27
22) T	Diisopropyl ether	2.095	1.924	2.219	2.278	2.295	2.321	2.189	6.97
23) T	Vinyl Acetate	1.660	1.698	1.928	2.048	2.082	2.134	1.925	10.52
24) P	1,1-Dichloroet...	1.116	1.154	1.233	1.263	1.263	1.286	1.219	5.60
25) T	2-Butanone	0.495	0.539	0.540	0.555	0.558	0.569	0.543	4.77
26) T	2,2-Dichloropr...	0.965	0.850	0.910	0.957	1.003	1.039	0.954	7.03
27) T	cis-1,2-Dichlo...	0.719	0.642	0.716	0.737	0.738	0.755	0.718	5.55
28) T	Bromochloromet...	0.576	0.553	0.628	0.578	0.595	0.590	0.587	4.28
29) T	Tetrahydrofuran	0.318	0.318	0.340	0.350	0.351	0.362	0.340	5.36
30) C	Chloroform	1.265	1.199	1.287	1.296	1.277	1.300	1.271	2.95#
31) T	Cyclohexane		1.059	1.090	1.128	1.128	1.150	1.111	3.26
32) T	1,1,1-Trichlor...	1.015	1.013	1.106	1.131	1.155	1.188	1.101	6.60
33) S	1,2-Dichloroet...		0.935	0.953	0.910	0.930	0.932	0.932	1.65
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.354	0.359	0.355	0.364	0.368	0.360	1.70
36) T	1,1-Dichloropr...	0.493	0.462	0.463	0.495	0.483	0.505	0.484	3.68
37) T	Ethyl Acetate	0.586	0.582	0.569	0.611	0.609	0.631	0.598	3.83
38) T	Carbon Tetrach...	0.541	0.505	0.528	0.558	0.552	0.577	0.544	4.61
39) T	Methylcyclohexane	0.627	0.587	0.596	0.641	0.627	0.658	0.623	4.34
40) TM	Benzene	1.348	1.337	1.426	1.474	1.441	1.477	1.417	4.30
41) T	Methacrylonitrile	0.233	0.288	0.318	0.346	0.343	0.348	0.313	14.50
42) TM	1,2-Dichloroet...	0.579	0.594	0.632	0.627	0.611	0.625	0.612	3.45
43) T	Isopropyl Acetate	0.764	0.826	0.905	0.963	0.982	1.030	0.912	11.05
44) TM	Trichloroethene	0.324	0.315	0.344	0.355	0.345	0.362	0.341	5.28
45) C	1,2-Dichloropr...	0.317	0.324	0.356	0.371	0.368	0.378	0.352	7.37#
46) T	Dibromomethane	0.263	0.262	0.285	0.287	0.280	0.289	0.278	4.35
47) T	Bromodichlorom...	0.485	0.498	0.557	0.577	0.573	0.594	0.547	8.22
48) T	Methyl methacr...	0.370	0.426	0.465	0.502	0.500	0.531	0.466	12.75
49) T	1,4-Dioxane	0.007	0.009	0.009	0.009	0.009	0.010	0.009	8.45
50) S	Toluene-d8		1.221	1.246	1.223	1.266	1.275	1.246	1.95
51) T	4-Methyl-2-Pen...	0.561	0.555	0.620	0.634	0.630	0.631	0.605	6.04
52) CM	Toluene	0.803	0.838	0.884	0.898	0.885	0.904	0.869	4.54#
53) T	t-1,3-Dichloro...	0.371	0.406	0.468	0.528	0.555	0.591	0.487	17.86
54) T	cis-1,3-Dichlo...	0.423	0.469	0.531	0.578	0.602	0.623	0.538	14.61
55) T	1,1,2-Trichlor...	0.308	0.337	0.349	0.354	0.351	0.356	0.343	5.30
56) T	Ethyl methacry...	0.377	0.508	0.540	0.595	0.617	0.639	0.546	17.58

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M

57)	T	1,3-Dichloropr...	0.610	0.601	0.618	0.623	0.613	0.627	0.615	1.53
58)	T	2-Chloroethyl ...	0.230	0.247	0.270	0.307	0.303	0.313	0.278	12.41
59)	T	2-Hexanone	0.385	0.414	0.466	0.473	0.477	0.473	0.448	8.69
60)	T	Dibromochlorom...	0.306	0.326	0.378	0.400	0.415	0.431	0.376	13.28
61)	T	1,2-Dibromoethane	0.322	0.333	0.359	0.373	0.368	0.381	0.356	6.54
62)	S	4-Bromofluorob...		0.464	0.455	0.470	0.500	0.500	0.478	4.41
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.347	0.323	0.390	0.375	0.345	0.344	0.354	6.84
65)	PM	Chlorobenzene	1.131	1.046	1.093	1.098	1.085	1.114	1.094	2.65
66)	T	1,1,1,2-Tetrac...	0.369	0.341	0.365	0.390	0.382	0.395	0.374	5.22
67)	C	Ethyl Benzene	1.803	1.816	1.919	2.022	1.979	2.036	1.929	5.24#
68)	T	m/p-Xylenes	0.648	0.678	0.706	0.740	0.721	0.740	0.706	5.21
69)	T	o-Xylene	0.642	0.639	0.688	0.727	0.706	0.726	0.688	5.75
70)	T	Styrene	0.951	1.012	1.135	1.219	1.214	1.230	1.127	10.56
71)	P	Bromoform	0.234	0.236	0.270	0.304	0.312	0.327	0.281	14.17
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.789	3.562	3.843	4.130	3.876	4.156	3.893	5.72
74)	T	N-amyl acetate	1.715	1.652	1.846	2.067	2.068	2.192	1.924	11.32
75)	P	1,1,2,2-Tetrac...	1.552	1.350	1.315	1.338	1.284	1.345	1.364	6.97
76)	T	1,2,3-Trichlor...	1.405	1.167	1.151	1.187	1.131	1.181	1.204	8.36
77)	T	Bromobenzene	0.926	0.862	0.896	0.928	0.883	0.928	0.904	3.08
78)	T	n-propylbenzene	4.272	4.186	4.394	4.854	4.583	4.868	4.526	6.45
79)	T	2-Chlorotoluene	3.184	2.748	2.832	2.994	2.805	2.953	2.919	5.45
80)	T	1,3,5-Trimethy...	3.036	3.053	3.275	3.487	3.255	3.405	3.252	5.60
81)	T	trans-1,4-Dich...		0.269	0.335	0.385	0.410	0.449	0.370	18.91
82)	T	4-Chlorotoluene	3.226	2.939	3.196	3.430	3.255	3.379	3.238	5.32
83)	T	tert-Butylbenzene	3.341	3.098	3.115	3.435	3.255	3.411	3.276	4.44
84)	T	1,2,4-Trimethy...	3.150	3.034	3.274	3.522	3.335	3.444	3.293	5.52
85)	T	sec-Butylbenzene	3.708	3.767	3.937	4.282	4.095	4.343	4.022	6.55
86)	T	p-Isopropyltol...	3.025	3.084	3.206	3.555	3.450	3.599	3.320	7.44
87)	T	1,3-Dichlorobe...	1.619	1.558	1.633	1.701	1.656	1.729	1.649	3.71
88)	T	1,4-Dichlorobe...	1.817	1.606	1.629	1.693	1.639	1.722	1.684	4.64
89)	T	n-Butylbenzene	2.443	2.650	2.748	3.147	3.139	3.346	2.912	12.00
90)	T	Hexachloroethane	0.511	0.523	0.551	0.622	0.622	0.680	0.585	11.44
91)	T	1,2-Dichlorobe...	1.710	1.577	1.613	1.696	1.634	1.702	1.655	3.34
92)	T	1,2-Dibromo-3-...	0.259	0.248	0.299	0.322	0.329	0.356	0.302	13.89
93)	T	1,2,4-Trichlor...	0.862	0.842	0.861	0.981	1.035	1.123	0.951	12.03
94)	T	Hexachlorobuta...	0.393	0.414	0.394	0.427	0.418	0.445	0.415	4.79
95)	T	Naphthalene	3.499	2.929	3.204	3.613	3.690	3.984	3.487	10.69
96)	T	1,2,3-Trichlor...	0.941	0.846	0.921	1.019	1.051	1.107	0.981	9.74

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	25.640%#		
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	25.600%#		
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	26.400%#		
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	26.480%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	27845m	11.039	ug/l	
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 8367
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1





#3

Chloromethane

Concen: 13.371 ug/l

RT: 1.306 min Scan# 30

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 50 Resp: 24310

Ion Ratio Lower Upper

50 100





52 33.0 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#4

Vinyl Chloride

Concen: 13.868 ug/l

RT: 1.373 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 62 Resp: 22096

Ion Ratio Lower Upper

62 100

64 34.9 25.2 37.8









#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 14.271 ug/l
RT: 1.873 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion: 101 Resp: 3463
Ion Ratio Lower Upper
101 100
103 63.5 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#8
Diethyl Ether
Concen: 14.278 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35



Tgt Ion: 74 Resp: 11392
Ion Ratio Lower Upper
74 100
45 103.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 14.499 ug/l
RT: 2.324 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 2102
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 70.4 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 15.070 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:142 Resp: 25661
Ion Ratio Lower Upper
142 100
127 53.0 41.7 62.5
141 14.7 11.5 17.3





#11

Tert butyl alcohol

Concen: 68.662 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 59 Resp: 20370

Ion Ratio Lower Upper

59 100





57 11.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 13.385 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 96 Resp: 18899


Ion Ratio Lower Upper

96 100

61 179.8 146.2 219.2

98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 2642

Ion Ratio Lower Upper

56 100





55 74.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3





#15

Acrylonitrile

Concen: 71.312 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion:	53	Resp:	63290
Ion	Ratio	Lower	Upper
53	100		
52	83.5	65.3	97.9
51	38.4	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 70.884 ug/l




RT: 2.379 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	43	Resp:	60363
Ion	Ratio	Lower	Upper
43	100		
58	26.0	21.2	31.8





#17

Carbon Disulfide

Concen: 13.251 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 4334

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 13.397 ug/l

RT: 2.702 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 43 Resp: 27234

Ion Ratio Lower Upper

43 100

74 20.7 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 13.890 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 73 Resp: 6840

Ion Ratio Lower Upper

73 100





57 22.2 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 13.362 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion: 84 Resp: 23050




Ion Ratio Lower Upper


84 100

49 146.0 113.9 170.9

51 43.8 33.5 50.3

86 62.2 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 13.294 ug/l

RT: 3.086 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 1916

Ion Ratio Lower Upper

96 100

61 155.6 119.5 179.3





98 63.2 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 14.896 ug/l

RT: 3.763 min Scan# 439

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion: 45 Resp: 74257




Ion Ratio Lower Upper


45 100

43 78.2 66.6 100.0

87 24.0 19.8 29.6

59 11.7 8.6 12.8









#23
Vinyl Acetate
Concen: 73.230 ug/l
RT: 3.720 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 322598
Ion Ratio Lower Upper
43 100
86 9.3 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 14.082 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 63 Resp: 41257
Ion Ratio Lower Upper
63 100
98 7.9 3.6 10.8
100 5.0 2.1 6.3





#25

2-Butanone

Concen: 73.489 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 9033

Ion Ratio Lower Upper

43 100





72 23.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 13.729 ug/l

RT: 4.470 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 77 Resp: 30472

Ion Ratio Lower Upper

77 100

97 20.9 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 13.791 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	96	Resp:	23970
Ion	Ratio	Lower	Upper
96	100		
61	158.3	0.0	322.8
98	63.5	0.0	129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 13.497 ug/l




RT: 4.897 min Scan# 625


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	49	Resp:	21028
Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.0
130	70.9	56.2	84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	42	Resp:	56819
Ion	Ratio	Lower	Upper
42	100		
72	42.3	33.7	50.5
71	39.2	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 14.144 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	83	Resp:	43065
Ion	Ratio	Lower	Upper
83	100		
85	65.7	49.3	73.9





#31

Cyclohexane

Concen: 14.769 ug/l

RT: 5.458 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 36470

Ion Ratio Lower Upper

56 100

69 28.8 24.4 36.6





84 78.5 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 14.092 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 97 Resp: 37003


Ion Ratio Lower Upper

97 100

99 62.8 51.8 77.6

61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100





67 49.9 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2

88 18.0 0.0 33.6





#35

Dibromofluoromethane

Concen: 12.804 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion:	113	Resp:	2111
Ion	Ratio	Lower	Upper
113	100		
111	103.0	83.1	124.7
192	17.0	13.3	19.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 14.231 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	75	Resp:	27242
Ion	Ratio	Lower	Upper
75	100		
110	31.6	16.3	48.9
77	31.9	24.3	36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X





ClientSampleId :

VSTDICC020

Tgt Ion:	43	Resp:	3345
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	10.9	7.9	11.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 14.137 ug/l




RT: 5.671 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	117	Resp:	31091
Ion	Ratio	Lower	Upper
117	100		
119	95.6	75.2	112.8
121	32.1	24.2	36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 83	Resp: 3507
Ion Ratio	Lower Upper
83 100	
55 80.7	64.7 97.1
98 44.8	37.4 56.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 14.140 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion: 78	Resp: 83898
Ion Ratio	Lower Upper
78 100	
77 22.3	19.0 28.4





#41

Methacrylonitrile

Concen: 14.038 ug/l

RT: 4.915 min Scan# 61

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 41 Resp: 1873

Ion Ratio Lower Upper

41 100

39 62.6 47.2 70.8

67 68.3 50.7 76.1





52 32.5 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#42

1,2-Dichloroethane

Concen: 15.165 ug/l

RT: 6.080 min Scan# 819

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 62 Resp: 37203

Ion Ratio Lower Upper

62 100

98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 14.512 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 43	Resp: 5327
Ion Ratio	Lower Upper
43	100
61	17.3 14.3 21.5
87	11.6 9.5 14.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 14.429 ug/l




RT: 7.116 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:130	Resp: 20239
Ion Ratio	Lower Upper
130	100
95	102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100





65 29.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2





#47

Bromodichloromethane

Concen: 14.545 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	83	Resp:	3276
Ion	Ratio	Lower	Upper
83	100		
85	68.4	50.5	75.7
127	8.3	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 14.521 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	41	Resp:	27374
Ion	Ratio	Lower	Upper
41	100		
69	73.5	58.5	87.7
39	64.9	51.7	77.5





#49

1,4-Dioxane

Concen: 282.442 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion:	88	Resp:	1058
Ion	Ratio	Lower	Upper
88	100		
43	45.5	33.4	50.2
58	73.6	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 13.200 ug/l




RT: 8.646 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	98	Resp:	73333
Ion	Ratio	Lower	Upper
98	100		
100	64.8	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 18245
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100





77 31.4 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 14.442 ug/l

RT: 9.152 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 97 Resp: 20530

Ion Ratio Lower Upper

97 100

83 87.0 69.0 103.6

85 58.7 44.3 66.5





99 66.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 14.425 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 69 Resp: 31787


Ion Ratio Lower Upper

69 100

41 78.2 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 3634

Ion Ratio Lower Upper

76 100





78 31.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 74.997 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 137054

Ion Ratio Lower Upper

43 100





58 48.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 14.387 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:129 Resp: 22267

Ion Ratio Lower Upper

129 100

127 78.0 39.3 117.8





#61

1,2-Dibromoethane

Concen: 14.458 ug/l

RT: 9.610 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion:107 Resp: 2111

Ion Ratio Lower Upper

107 100





109 95.3 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 13.236 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 95 Resp: 26760


Ion Ratio Lower Upper

95 100

174 67.4 0.0 135.8

176 67.0 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1470

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:117 Resp: 127829

Ion Ratio Lower Upper

117 100

82 57.7 50.6 76.0





119 32.7 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#64

Tetrachloroethene

Concen: 15.263 ug/l

RT: 9.268 min Scan# 1342

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924




Ion Ratio Lower Upper


164 100

166 127.2 105.0 157.6

129 107.6 83.5 125.3

131 97.7 76.5 114.7





#65

Chlorobenzene

Concen: 14.202 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion:112 Resp: 5586

Ion Ratio Lower Upper

112 100





114 32.0 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 14.231 ug/l

RT: 10.158 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:131 Resp: 18671


Ion Ratio Lower Upper

131 100

133 94.5 47.3 141.9

119 66.7 31.6 95.0





#67

Ethyl Benzene

Concen: 14.631 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 9811

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 29.806 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:106 Resp: 72202

Ion Ratio Lower Upper

106 100

91 216.1 171.2 256.8





#69

o-Xylene

Concen: 14.408 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 35178

Ion Ratio Lower Upper

106 100





91 231.7 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 14.853 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:104 Resp: 58034


Ion Ratio Lower Upper

104 100

78 58.7 45.7 68.5


103 55.6 43.7 65.5





#71
Bromoform
Concen: 14.030 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 13825
Ion Ratio Lower Upper
173 100
175 46.4 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:152 Resp: 60503
Ion Ratio Lower Upper
152 100
115 76.6 46.9 140.7
150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:105 Resp: 9301
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 13.560 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 83 Resp: 31830
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 63.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 11.039 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 27845
Ion Ratio Lower Upper
75 100
77 43.0 20.5 61.5





#77

Bromobenzene

Concen: 14.212 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	156	Resp:	2169
Ion Ratio	Lower	Upper	
156	100		
77	179.9	92.8	278.3
158	93.1	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#78

n-propylbenzene

Concen: 14.376 ug/l




RT: 11.298 min Scan# 1675


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	91	Resp:	106337
Ion Ratio	Lower	Upper	
91	100		
120	22.0	10.8	32.4





#79

2-Chlorotoluene

Concen: 13.748 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 68534

Ion Ratio Lower Upper

91 100





126 31.4 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 14.468 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:105 Resp: 79261

Ion Ratio Lower Upper

105 100

120 45.9 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 13.371 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 8104
Ion Ratio Lower Upper
75 100
53 138.3 100.5 150.7
89 44.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 14.116 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 91 Resp: 77338
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 14.011 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	119	Resp:	7538
Ion	Ratio	Lower	Upper
119	100		
91	66.8	32.9	98.7
134	23.7	11.4	34.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 14.480 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	105	Resp:	79232
Ion	Ratio	Lower	Upper
105	100		
120	42.0	21.2	63.6





#85

sec-Butylbenzene

Concen: 14.263 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion:105 Resp: 9527

Ion Ratio Lower Upper

105 100





134 19.3 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 14.387 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:119 Resp: 77599


Ion Ratio Lower Upper

119 100

134 25.9 12.5 37.5

91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:146	Resp:	39520
Ion Ratio	Lower	Upper
146	100	
111	42.7	22.1 66.3
148	62.9	32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:146	Resp:	39419
Ion Ratio	Lower	Upper
146	100	
111	42.8	21.3 63.9
148	64.2	31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 66500

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7





134 23.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:117 Resp: 13332

Ion Ratio Lower Upper

117 100


201 62.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:146 Resp: 3903
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 13.745 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:180 Resp: 20840
Ion Ratio Lower Upper
180 100
182 96.6 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 13.426 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:225 Resp: 9524
Ion Ratio Lower Upper
225 100
223 65.6 30.8 92.4
227 61.4 30.9 92.7





#95

Naphthalene

Concen: 14.040 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	128	Resp:	7754
Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.4	15.6
129	11.1	8.6	13.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 14.023 ug/l




RT: 13.956 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	180	Resp:	22301
Ion	Ratio	Lower	Upper
180	100		
182	92.1	47.8	143.3
145	36.5	18.1	54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 97070
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1





#3

Chloromethane

Concen: 35.658 ug/l

RT: 1.307 min Scan# 30

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 50 Resp: 75238

Ion Ratio Lower Upper

50 100





52 31.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#4

Vinyl Chloride

Concen: 37.273 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 62 Resp: 68901

Ion Ratio Lower Upper

62 100

64 31.5 25.2 37.8









#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 36.824 ug/l
RT: 1.874 min Scan# 129
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:101 Resp: 10369
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#8
Diethyl Ether
Concen: 35.350 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion: 74 Resp: 32724
Ion Ratio Lower Upper
74 100
45 109.9 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 36.997 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:101	Resp: 6225
Ion Ratio	Lower Upper
101 100	
85 48.3	38.6 58.0
151 69.0	55.2 82.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 39.622 ug/l




RT: 2.447 min Scan# 223


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:142	Resp: 78277
Ion Ratio	Lower Upper
142 100	
127 52.1	41.7 62.5
141 14.4	11.5 17.3





#11

Tert butyl alcohol

Concen: 181.745 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 6255

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 35.610 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 96 Resp: 58334


Ion Ratio Lower Upper

96 100

61 182.7 146.2 219.2

98 63.8 51.0 76.6





#13

Acrolein

Concen: 184.907 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 7392

Ion Ratio Lower Upper

56 100





55 70.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 36.971 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 41 Resp: 114445


Ion Ratio Lower Upper

41 100

39 75.7 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 182.874 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	53	Resp:	188304
Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.3	97.9
51	37.3	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 177.910 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	43	Resp:	175777
Ion	Ratio	Lower	Upper
43	100		
58	26.5	21.2	31.8





#17

Carbon Disulfide

Concen: 37.223 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 14128

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 34.915 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 43 Resp: 82347

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 36.695 ug/l

RT: 3.111 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 20966

Ion Ratio Lower Upper





73	100		
57	22.1	17.7	26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 33.182 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min




Lab File: VX046042.D


Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412

Ion Ratio Lower Upper

84	100		
49	142.4	113.9	170.9
51	41.9	33.5	50.3
86	67.3	53.8	80.8









#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
Vinyl Acetate
Concen: 194.528 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 994230
Ion Ratio Lower Upper
43 100
86 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 36.067 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 63 Resp: 122601
Ion Ratio Lower Upper
63 100
98 7.2 3.6 10.8
100 4.2 2.1 6.3





#25

2-Butanone

Concen: 188.783 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 269224

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 36.059 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8





98 64.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	42	Resp:	170124
Ion	Ratio	Lower	Upper
42	100		
72	42.1	33.7	50.5
71	39.3	31.4	47.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 35.626 ug/l




RT: 5.087 min Scan# 656


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	83	Resp:	125850
Ion	Ratio	Lower	Upper
83	100		
85	61.6	49.3	73.9





#31

Cyclohexane

Concen: 38.205 ug/l

RT: 5.464 min Scan# 718

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 109459

Ion Ratio Lower Upper

56 100

69 30.5 24.4 36.6





84 83.6 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 36.034 ug/l

RT: 5.373 min Scan# 703

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 97 Resp: 109781


Ion Ratio Lower Upper

97 100

99 64.7 51.8 77.6

61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8837

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 31.633 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	113	Resp:	59550
Ion	Ratio	Lower	Upper
113	100		
111	103.9	83.1	124.7
192	16.6	13.3	19.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 38.073 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	75	Resp:	83215
Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.3	48.9
77	30.4	24.3	36.5





#37

Ethyl Acetate

Concen: 37.394 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 10253

Ion Ratio Lower Upper

43 100

61 12.9 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 37.320 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:117 Resp: 93712


Ion Ratio Lower Upper

117 100

119 94.0 75.2 112.8

121 30.3 24.2 36.4





#39

Methylcyclohexane

Concen: 39.307 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 10765

Ion Ratio Lower Upper

83 100

55 80.9 64.7 97.1





98 46.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 36.530 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 78 Resp: 247476

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 612
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 5808
Ion Ratio Lower Upper
41 100
39 59.0 47.2 70.8
67 63.4 50.7 76.1
52 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 105332
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 16178

Ion Ratio Lower Upper

43

100

61

17.9

87

11.9

14.3

9.5

21.5





14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130




100


95

102.1

0.0

204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 6238

Ion Ratio Lower Upper

63 100





65 30.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2





#47

Bromodichloromethane

Concen: 37.685 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	83	Resp:	96910
Ion	Ratio	Lower	Upper
83	100		
85	63.1	50.5	75.7
127	8.1	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 39.157 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	41	Resp:	84277
Ion	Ratio	Lower	Upper
41	100		
69	73.1	58.5	87.7
39	64.6	51.7	77.5





#49

1,4-Dioxane

Concen: 713.477 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	88	Resp:	30529
Ion	Ratio	Lower	Upper
88	100		
43	41.8	33.4	50.2
58	73.3	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 32.395 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	98	Resp:	205479
Ion	Ratio	Lower	Upper
98	100		
100	66.9	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 193.653 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 53238
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 37.530 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 92 Resp: 150757
Ion Ratio Lower Upper
92 100
91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 8865

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031


Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0

39 65.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 36.648 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 5949

Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.4 44.3 66.5





99 59.7 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 39.726 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 69 Resp: 99947


Ion Ratio Lower Upper

69 100

41 76.0 60.8 91.2

39 48.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 104600

Ion Ratio Lower Upper

76 100





78 32.9 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3





#59

2-Hexanone

Concen: 190.168 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 396784

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 38.027 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:129 Resp: 67198

Ion Ratio Lower Upper

129 100

127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 6263
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 14625
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7









#65
Chlorobenzene
Concen: 35.686 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 160600
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 38.017 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:131 Resp: 57066
Ion Ratio Lower Upper
131 100
133 94.6 47.3 141.9
119 63.3 31.6 95.0





#67

Ethyl Benzene

Concen: 38.543 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 29571

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 78.142 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:106 Resp: 216578

Ion Ratio Lower Upper

106 100

91 214.0 171.2 256.8





#69

o-Xylene

Concen: 38.064 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 106333

Ion Ratio Lower Upper

106 100





91 225.4 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 39.888 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:104 Resp: 178313


Ion Ratio Lower Upper

104 100

78 57.1 45.7 68.5


103 54.6 43.7 65.5





#71
Bromoform
Concen: 39.457 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:173 Resp: 44480
Ion Ratio Lower Upper
173 100
175 46.7 23.4 70.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion:152 Resp: 67976
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 34.495 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	90978
Ion	Ratio	Lower	Upper
83	100		
131	9.9	5.0	14.9
85	63.8	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 28.468 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58


Tgt Ion:	75	Resp:	80678
Ion	Ratio	Lower	Upper
75	100		
77	41.0	20.5	61.5





#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 156 Resp: 6305
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4





#79

2-Chlorotoluene

Concen: 36.336 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 203509

Ion Ratio Lower Upper

91 100





126 31.1 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 38.516 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 237066

Ion Ratio Lower Upper

105 100

120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 38.622 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:119 Resp: 233468

Ion Ratio Lower Upper

119 100

91 65.8 32.9 98.7





134 22.7 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 38.946 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 239433

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6





#85

sec-Butylbenzene

Concen: 38.786 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 291074

Ion Ratio Lower Upper

105 100





134 19.4 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 39.879 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:119 Resp: 241658


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 36.199 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 11565

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3





148 64.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 36.763 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:146 Resp: 115097


Ion Ratio Lower Upper

146 100

111 42.6 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 21389

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7





134 23.5 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:117 Resp: 42314

Ion Ratio Lower Upper

117 100

201 63.1 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:146 Resp: 115280
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 39.117 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:180 Resp: 6665

Ion Ratio Lower Upper

180 100

182 95.9 47.9 143.8





145 34.1 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 36.395 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:225 Resp: 29007


Ion Ratio Lower Upper

225 100

223 61.6 30.8 92.4


227 61.8 30.9 92.7





#95
Naphthalene
Concen: 39.575 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:128 Resp: 245568
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.8 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 38.765 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion:180 Resp: 69262
Ion Ratio Lower Upper
180 100
182 95.5 47.8 143.3
145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.060%#	
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 130.020%#	
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 134.060%#	
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 145.640%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	150182m	54.277	ug/l	
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDIC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1









#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 50 Resp: 13693
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8









#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 94 Resp: 59240
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5









#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 17104
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 72.604 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:101 Resp: 109520

Ion Ratio Lower Upper

101 100

85 48.1 38.6 58.0





151 69.9 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 77.943 ug/l

RT: 2.440 min Scan# 222

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:142 Resp: 138045


Ion Ratio Lower Upper

142 100

127 52.0 41.7 62.5

141 14.5 11.5 17.3





#11

Tert butyl alcohol

Concen: 405.394 ug/l

RT: 2.995 min Scan# 310

Delta R.T. 0.018 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 59 Resp: 125094

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 71.884 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 96 Resp: 105566


Ion Ratio Lower Upper

96 100

61 189.9 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 375.132 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 134450

Ion Ratio Lower Upper

56 100





55 71.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 74.419 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 206519


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8

76 31.4 24.9 37.3





#15

Acrylonitrile

Concen: 358.723 ug/l

RT: 3.068 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 53 Resp: 33114

Ion Ratio Lower Upper

53 100

52 84.7 65.3 97.9





51 37.9 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 354.936 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 314383

Ion Ratio Lower Upper

43 100

58 27.6 21.2 31.8





#17

Carbon Disulfide

Concen: 77.867 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 264950

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 69.531 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 147014

Ion Ratio Lower Upper

43 100

74 22.6 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 73.818 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 73 Resp: 37811

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 67.030 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt Ion: 84 Resp: 120270




Ion Ratio Lower Upper


84 100

49 136.4 113.9 170.9

51 41.1 33.5 50.3


86 64.0 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 106600
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8









#23
Vinyl Acetate
Concen: 395.363 ug/l
RT: 3.721 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1811550
Ion Ratio Lower Upper
43 100
86 9.5 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 72.114 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
Ion Ratio Lower Upper
63 100
98 6.8 3.6 10.8
100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 48544

Ion Ratio Lower Upper

43 100





72 23.1 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 71.056 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 96 Resp: 128454

Ion Ratio Lower Upper

96 100

61 160.1 0.0 322.8





98 64.9 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 63.932 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 49 Resp: 103602


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 30575

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 62.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 65 Resp: 16187

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:114 Resp: 152958


Ion Ratio Lower Upper

114 100

63 24.3 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 65.006 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:113 Resp: 11145

Ion Ratio Lower Upper

113 100

111 101.7 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 74.285 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 147873


Ion Ratio Lower Upper

75 100

110 33.5 16.3 48.9

77 31.1 24.3 36.5





#37

Ethyl Acetate

Concen: 74.565 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X





ClientSampleId :

VSTDICC100

Tgt Ion:	43	Resp:	186215
Ion	Ratio	Lower	Upper
43	100		
61	12.4	10.3	15.5
70	9.9	7.9	11.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 73.875 ug/l




RT: 5.672 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt Ion:	117	Resp:	168946
Ion	Ratio	Lower	Upper
117	100		
119	95.0	75.2	112.8
121	29.6	24.2	36.4





#39

Methylcyclohexane

Concen: 76.952 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 19194

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1





98 49.0 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 71.457 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 78 Resp: 440885

Ion Ratio Lower Upper

78 100

77 24.6 19.0 28.4









#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 41 Resp: 10486
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 78.714 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 30049

Ion Ratio Lower Upper

43 100

61 18.0 14.3 21.5





87 11.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 72.420 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:130 Resp: 105628

Ion Ratio Lower Upper

130 100

95 103.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2





#47

Bromodichloromethane

Concen: 74.845 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17530

Ion Ratio Lower Upper

83 100

85 63.6 50.5 75.7





127 7.8 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 78.101 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 153093


Ion Ratio Lower Upper

41 100

69 74.1 58.5 87.7

39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion	Ratio	Lower	Upper
88	100		
43	40.1	33.4	50.2
58	73.3	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 67.031 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt Ion	Ratio	Lower	Upper
98	100		
100	65.2	53.5	80.3





#51

4-Methyl-2-Pentanone

Concen: 384.740 ug/l

RT: 8.573 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 963319

Ion Ratio Lower Upper

43 100





58 36.1 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#52

Toluene

Concen: 74.010 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 92 Resp: 270763

Ion Ratio Lower Upper

92 100

91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 86.372 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 75 Resp: 16986

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 80.264 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 184134


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 59.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 72.543 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 107264

Ion Ratio Lower Upper

97 100

83 86.8 69.0 103.6

85 56.4 44.3 66.5





99 61.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 82.359 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 69 Resp: 188717


Ion Ratio Lower Upper

69 100

41 75.7 60.8 91.2

39 47.7 39.0 58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100





78 32.1 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 383.945 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 729602

Ion Ratio Lower Upper

43 100





58 49.9 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 78.918 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:129 Resp: 127012

Ion Ratio Lower Upper

129 100

127 77.0 39.3 117.8





#61

1,2-Dibromoethane

Concen: 74.200 ug/l

RT: 9.610 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:107 Resp: 112669

Ion Ratio Lower Upper

107 100





109 93.9 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 72.818 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 95 Resp: 153085


Ion Ratio Lower Upper

95 100

174 67.1 0.0 135.8

176 65.3 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion:117	Resp: 135214
Ion Ratio	Lower Upper
117 100	
82 61.7	50.6 76.0
119 30.6	25.8 38.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#64

Tetrachloroethene

Concen: 67.493 ug/l




RT: 9.268 min Scan# 1342


Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt	Ion:164	Resp:	93195
Ion	Ratio	Lower	Upper
164	100		
166	123.4	105.0	157.6
129	101.4	83.5	125.3
131	98.0	76.5	114.7









#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0





#67

Ethyl Benzene

Concen: 75.444 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 53512

Ion Ratio Lower Upper

91 100





106 29.6 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 152.181 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:106 Resp: 389935

Ion Ratio Lower Upper

106 100

91 215.8 171.2 256.8





#69

o-Xylene

Concen: 73.890 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:106 Resp: 190833

Ion Ratio Lower Upper

106 100





91 230.3 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 79.411 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:104 Resp: 328195


Ion Ratio Lower Upper

104 100

78 57.4 45.7 68.5


103 54.8 43.7 65.5





#71
Bromoform
Concen: 80.927 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 8435
Ion Ratio Lower Upper
173 100
175 48.3 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:152 Resp: 66369
Ion Ratio Lower Upper
152 100
115 120.4 46.9 140.7
150 190.4 0.0 351.0









#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 66.207 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17048

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 54.277 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 75 Resp: 150182

Ion Ratio Lower Upper

75 100


77 42.2 20.5 61.5





#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 156 Resp: 11715
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 68.100 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 372392

Ion Ratio Lower Upper

91 100





126 30.9 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 71.903 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 432096

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 73.197 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100




Tgt Ion:119 Resp: 43201
Ion Ratio Lower Upper
119 100
91 64.4 32.9 98.7
134 22.8 11.4 34.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#84
1,2,4-Trimethylbenzene
Concen: 73.758 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Tgt Ion:105 Resp: 442728
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6





#85
sec-Butylbenzene
Concen: 74.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:105 Resp: 54359
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#86
p-Isopropyltoluene
Concen: 77.393 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:119 Resp: 457898
Ion Ratio Lower Upper
119 100
134 24.6 12.5 37.5
91 26.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 70.449 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:146 Resp: 21975

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 71.190 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:146 Resp: 217613


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 41660

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:117 Resp: 82528

Ion Ratio Lower Upper

117 100


201 63.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8





145 34.6 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4

227 64.9 30.9 92.7





#95

Naphthalene

Concen: 80.842 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:128 Resp: 489778

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6





129 10.9 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 79.981 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:180 Resp: 139525

Ion Ratio Lower Upper

180 100

182 93.7 47.8 143.3

145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 187.960%#	
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 197.120%#	
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 202.500%#	
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 218.340%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	213744m	84.960	ug/l	
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025


Quant Time: May 06 06:11:37 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Quant Title : SW846 8260

QLast Update : Tue May 06 06:04:56 2025

Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:168 Resp: 8296
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1









#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 50 Resp: 19684
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8









#5
Bromomethane
Concen: 103.625 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 94 Resp: 83008
Ion Ratio Lower Upper
94 100
96 95.8 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 96.822 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 64 Resp: 78806
Ion Ratio Lower Upper
64 100
66 31.3 24.3 36.5









#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 101
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 616.119 ug/l

RT: 3.002 min Scan# 31

Delta R.T. 0.024 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 59 Resp: 181238

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 111.039 ug/l

RT: 2.307 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 96 Resp: 155451


Ion Ratio Lower Upper

96 100

61 184.5 146.2 219.2

98 64.3 51.0 76.6





#13

Acrolein

Concen: 594.872 ug/l

RT: 2.239 min Scan# 188

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 20325

Ion Ratio Lower Upper

56 100





55 72.1 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 112.561 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 41 Resp: 297776


Ion Ratio Lower Upper

41 100

39 73.7 60.6 90.8

76 31.8 24.9 37.3





#15

Acrylonitrile

Concen: 550.973 ug/l

RT: 3.069 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X





ClientSampleId :

VSTDIC150

Tgt Ion:	53	Resp:	484854
Ion	Ratio	Lower	Upper
53	100		
52	84.2	65.3	97.9
51	37.5	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 545.754 ug/l




RT: 2.392 min Scan# 214


Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45


Tgt Ion:	43	Resp:	460820
Ion	Ratio	Lower	Upper
43	100		
58	27.3	21.2	31.8





#17
Carbon Disulfide
Concen: 122.499 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 76 Resp: 39735
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#18
Methyl Acetate
Concen: 108.099 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 217885
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 114.112 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 73 Resp: 557210

Ion Ratio Lower Upper

73 100





57 22.0 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 100.579 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion: 84 Resp: 172035




Ion Ratio Lower Upper


84 100

49 136.0 113.9 170.9

51 41.5 33.5 50.3


86 63.7 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150




Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45



Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8





#23

Vinyl Acetate

Concen: 608.005 ug/l

RT: 3.721 min Scan# 41

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 265575

Ion Ratio Lower Upper

43 100





86 9.6 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 110.182 ug/l

RT: 3.605 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 63 Resp: 320083


Ion Ratio Lower Upper

63 100

98 7.0 3.6 10.8

100 4.7 2.1 6.3





#25

2-Butanone

Concen: 581.292 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 70846

Ion Ratio Lower Upper

43 100





72 24.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 117.548 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 77 Resp: 258698

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 18783

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8





98 64.8 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 450450

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100

85 65.1 49.3 73.9





#31

Cyclohexane

Concen: 116.934 ug/l

RT: 5.458 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 28631

Ion Ratio Lower Upper

56 100

69 31.2 24.4 36.6





84 82.8 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 113.572 ug/l

RT: 5.373 min Scan# 703

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 97 Resp: 295702


Ion Ratio Lower Upper

97 100

99 64.3 51.8 77.6

61 48.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 93.983 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 65 Resp: 231952

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:114 Resp: 144975


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:113 Resp: 160168

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7





192 16.9 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5





#37

Ethyl Acetate

Concen: 115.889 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 274312

Ion Ratio Lower Upper

43 100

61 12.5 10.3 15.5





70 10.2 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 115.760 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:117 Resp: 250916


Ion Ratio Lower Upper

117 100

119 94.3 75.2 112.8

121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 28631

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1





98 47.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4





#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 41 Resp: 151424
Ion Ratio Lower Upper
41 100
39 60.0 47.2 70.8
67 65.9 50.7 76.1
52 30.9 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 123.780 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion: 43	Resp: 44787
Ion Ratio	Lower Upper
43	100
61	18.1 14.3 21.5
87	11.7 9.5 14.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 114.032 ug/l




RT: 7.117 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion:130	Resp: 157641
Ion Ratio	Lower Upper
130	100
95	105.9 0.0 204.2





#45

1,2-Dichloropropane

Concen: 113.217 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 16450

Ion Ratio Lower Upper

63 100





65 30.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 109.869 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 93 Resp: 125846


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2





#47

Bromodichloromethane

Concen: 116.365 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 83 Resp: 258320

Ion Ratio Lower Upper

83 100

85 64.8 50.5 75.7





127 8.0 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 124.356 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 41 Resp: 231040


Ion Ratio Lower Upper

41 100

69 73.0 58.5 87.7

39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2





58 74.8 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3





#51

4-Methyl-2-Pentanone

Concen: 577.831 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 137127

Ion Ratio Lower Upper

43 100





58 36.2 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#52

Toluene

Concen: 113.430 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 92 Resp: 393323

Ion Ratio Lower Upper

92 100

91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 138.001 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 75 Resp: 25723

Ion Ratio Lower Upper

75 100





77 31.7 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 124.658 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 271054


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0

39 60.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 110.551 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 97 Resp: 154932

Ion Ratio Lower Upper

97 100

83 88.5 69.0 103.6

85 56.0 44.3 66.5





99 64.5 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 127.968 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 69 Resp: 277921


Ion Ratio Lower Upper

69 100

41 75.4 60.8 91.2

39 47.2 39.0 58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 571.094 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 102859

Ion Ratio Lower Upper

43 100





58 50.2 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 122.915 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:129 Resp: 187497

Ion Ratio Lower Upper

129 100

127 77.1 39.3 117.8





#61

1,2-Dibromoethane

Concen: 115.087 ug/l

RT: 9.604 min Scan# 1397

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion:107 Resp: 165634

Ion Ratio Lower Upper

107 100





109 94.6 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 109.174 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 95 Resp: 217536


Ion Ratio Lower Upper

95 100

174 67.5 0.0 135.8


176 64.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:117 Resp: 12855
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7





#65

Chlorobenzene

Concen: 108.615 ug/l

RT: 10.080 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion:112 Resp: 429650

Ion Ratio Lower Upper

112 100





114 31.5 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 115.333 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:131 Resp: 152173


Ion Ratio Lower Upper

131 100

133 95.7 47.3 141.9

119 64.0 31.6 95.0









#67
Ethyl Benzene
Concen: 116.421 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 91 Resp: 78512
Ion Ratio Lower Upper
91 100
106 29.9 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 234.349 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:106 Resp: 570913
Ion Ratio Lower Upper
106 100
91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:106 Resp: 279900
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 127.128 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 125980
Ion Ratio Lower Upper
173 100
175 47.3 23.4 70.0
254 0.1 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:152 Resp: 60345
Ion Ratio Lower Upper
152 100
115 150.7 46.9 140.7#
150 211.6 0.0 351.0









#73
Isopropylbenzene
Concen: 114.312 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 75233
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 120.944 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 22.5 18.7 28.1
61 22.5 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 104.018 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 83 Resp: 24353
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 64.4 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 84.960 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 75 Resp: 213744
Ion Ratio Lower Upper
75 100
77 42.4 20.5 61.5





#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 107.514 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 534560

Ion Ratio Lower Upper

91 100





126 31.0 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 112.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 616477

Ion Ratio Lower Upper

105 100

120 46.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 134.609 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 75 Resp: 81372
Ion Ratio Lower Upper
75 100
53 114.9 100.5 150.7
89 46.5 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 111.948 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 611736
Ion Ratio Lower Upper
91 100
126 26.7 13.3 39.8









#83
tert-Butylbenzene
Concen: 115.065 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:119 Resp: 61747
Ion Ratio Lower Upper
119 100
91 65.6 32.9 98.7
134 22.6 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 114.236 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:105 Resp: 623456
Ion Ratio Lower Upper
105 100
120 43.1 21.2 63.6









#85
sec-Butylbenzene
Concen: 118.023 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 78629
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 121.113 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:119 Resp: 651531
Ion Ratio Lower Upper
119 100
134 25.4 12.5 37.5
91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3





148 64.0 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 60578

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:117 Resp: 123188

Ion Ratio Lower Upper

117 100

201 63.2 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 111.506 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:146 Resp: 308039
Ion Ratio Lower Upper
146 100
111 46.3 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 120.915 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
Ion Ratio Lower Upper
75 100
155 69.6 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:180	Resp: 20325
Ion Ratio	Lower Upper
180 100	
182 94.1	47.9 143.8
145 34.5	17.1 51.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 113.857 ug/l




RT: 13.719 min Scan# 2072


Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion:225	Resp: 80558
Ion Ratio	Lower Upper
225 100	
223 62.5	30.8 92.4
227 63.1	30.9 92.7





#95

Naphthalene

Concen: 130.929 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:128 Resp: 721230

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6





129 11.0 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 126.331 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:180 Resp: 200379

Ion Ratio Lower Upper

180 100

182 96.5 47.8 143.3

145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	6.280%#
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	6.320%#
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	6.460%#
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	6.760%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9806m	3.565	ug/l	
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	7928m	2.799	ug/l	
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:168 Resp: 9696
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 658
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8









#5
Bromomethane
Concen: 3.164 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 94 Resp: 2962
Ion Ratio Lower Upper
94 100
96 102.2 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 3.750 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 64 Resp: 3567
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 3.413 ug/l
RT: 1.874 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005




Tgt Ion:101 Resp: 959
Ion Ratio Lower Upper
101 100
103 54.0 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#8
Diethyl Ether
Concen: 3.266 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion: 74 Resp: 3020
Ion Ratio Lower Upper
74 100
45 115.1 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 3.517 ug/l

RT: 2.325 min Scan# 20

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	101	Resp:	591
Ion	Ratio	Lower	Upper
101	100		
85	46.3	38.6	58.0
151	68.8	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 2.990 ug/l




RT: 2.441 min Scan# 222


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	142	Resp:	5900
Ion	Ratio	Lower	Upper
142	100		
127	52.8	41.7	62.5
141	14.3	11.5	17.3





#11

Tert butyl alcohol

Concen: 16.111 ug/l

RT: 2.965 min Scan# 30

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 5539

Ion Ratio Lower Upper

59 100





57 10.0 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 3.359 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 96 Resp: 5496


Ion Ratio Lower Upper

96 100

61 178.7 146.2 219.2

98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 567

Ion Ratio Lower Upper

56 100





55 74.5 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3





#15

Acrylonitrile

Concen: 16.282 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion:	53	Resp:	16740
Ion	Ratio	Lower	Upper
53	100		
52	84.3	65.3	97.9
51	38.5	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 20.036 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	43	Resp:	19773
Ion	Ratio	Lower	Upper
43	100		
58	27.7	21.2	31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 1106

Ion Ratio Lower Upper

76 100





78 10.2 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 3.241 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 1849

Ion Ratio Lower Upper

73 100





57 20.5 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 3.341 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680




Ion Ratio Lower Upper


84 100

49 137.0 113.9 170.9

51 43.8 33.5 50.3

86 67.6 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 3.234 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 540

Ion Ratio Lower Upper

96 100

61 149.9 119.5 179.3





98 62.9 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 3.230 ug/l

RT: 3.751 min Scan# 437

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657




Ion Ratio Lower Upper


45 100

43 128.0 66.6 100.0#

87 24.7 19.8 29.6

59 10.7 8.6 12.8









#23
Vinyl Acetate
Concen: 16.127 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 82328
Ion Ratio Lower Upper
43 100
86 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 3.297 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
Ion Ratio Lower Upper
63 100
98 6.9 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 2612

Ion Ratio Lower Upper

43 100





72 21.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 3.089 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 96 Resp: 622

Ion Ratio Lower Upper

96 100

61 168.5 0.0 322.8





98 67.6 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 2.969 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 49 Resp: 5360


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 42 Resp: 1543

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5





71 37.4 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9





#31

Cyclohexane

Concen: 3.588 ug/l

RT: 5.458 min Scan# 717

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 10269

Ion Ratio Lower Upper

56 100

69 31.6 24.4 36.6





84 79.1 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 3.229 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 97 Resp: 9827


Ion Ratio Lower Upper

97 100

99 62.6 51.8 77.6

61 51.8 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 3.143 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 906

Ion Ratio Lower Upper

65 100





67 50.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:114 Resp: 168484


Ion Ratio Lower Upper

114 100

63 24.5 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 3.161 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	113	Resp:	596
Ion	Ratio	Lower	Upper
113	100		
111	104.4	83.1	124.7
192	17.5	13.3	19.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 3.551 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	75	Resp:	7787
Ion	Ratio	Lower	Upper
75	100		
110	30.6	16.3	48.9
77	30.5	24.3	36.5





#37

Ethyl Acetate

Concen: 3.565 ug/l m

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	43	Resp:	9800
Ion	Ratio	Lower	Upper
43	100		
61	11.0	10.3	15.5
70	8.4	7.9	11.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 3.376 ug/l




RT: 5.659 min Scan# 750


Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	117	Resp:	8505
Ion	Ratio	Lower	Upper
117	100		
119	90.6	75.2	112.8
121	29.5	24.2	36.4





#39

Methylcyclohexane

Concen: 3.597 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	988
Ion	Ratio	Lower	Upper
83	100		
55	73.3	64.7	97.1
98	50.3	37.4	56.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 3.315 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	78	Resp:	22528
Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.4





#41

Methacrylonitrile

Concen: 3.174 ug/l

RT: 4.916 min Scan# 61

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 41 Resp: 485

Ion Ratio Lower Upper

41 100

39 59.7 47.2 70.8

67 0.0 50.7 76.1





52 28.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#42

1,2-Dichloroethane

Concen: 3.563 ug/l

RT: 6.080 min Scan# 819

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 62 Resp: 10011

Ion Ratio Lower Upper

62 100

98 7.8 0.0 15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 43	Resp: 13914
Ion Ratio	Lower Upper
43	100
61	16.2 14.3 21.5
87	10.6 9.5 14.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 3.307 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:130	Resp: 5313
Ion Ratio	Lower Upper
130	100
95	108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 545

Ion Ratio Lower Upper

63 100





65 36.2 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2





#47

Bromodichloromethane

Concen: 3.254 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	8394
Ion	Ratio	Lower	Upper
83	100		
85	61.9	50.5	75.7
127	9.0	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 3.321 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	41	Resp:	7171
Ion	Ratio	Lower	Upper
41	100		
69	68.4	58.5	87.7
39	65.1	51.7	77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	88	Resp:	290
Ion	Ratio	Lower	Upper
88	100		
43	45.3	33.4	50.2
58	71.0	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 3.232 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	98	Resp:	20566
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100





77 31.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 3.490 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 568

Ion Ratio Lower Upper

97 100

83 88.0 69.0 103.6

85 54.4 44.3 66.5





99 62.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 3.390 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 69 Resp: 8555


Ion Ratio Lower Upper

69 100

41 73.7 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 3.508 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 1012

Ion Ratio Lower Upper

76 100





78 29.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 18.346 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 63 Resp: 20809

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 16.651 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 3485

Ion Ratio Lower Upper

43 100





58 47.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 3.102 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:129 Resp: 5500

Ion Ratio Lower Upper

129 100

127 80.4 39.3 117.8





#61

1,2-Dibromoethane

Concen: 3.356 ug/l

RT: 9.604 min Scan# 1397

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 3.378 ug/l




RT: 11.079 min Scan# 1639


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:117 Resp: 14716
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7





#65

Chlorobenzene

Concen: 3.399 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:112 Resp: 1539

Ion Ratio Lower Upper

112 100





114 32.6 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 3.326 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:131 Resp: 5024


Ion Ratio Lower Upper

131 100

133 95.0 47.3 141.9

119 64.8 31.6 95.0





#67

Ethyl Benzene

Concen: 3.463 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 2673

Ion Ratio Lower Upper

91 100





106 29.9 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 7.156 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:106 Resp: 19957

Ion Ratio Lower Upper

106 100

91 210.1 171.2 256.8





#69

o-Xylene

Concen: 3.343 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 9398

Ion Ratio Lower Upper

106 100





91 233.6 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 3.312 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:104 Resp: 14900


Ion Ratio Lower Upper

104 100

78 57.9 45.7 68.5


103 55.6 43.7 65.5





#71
Bromoform
Concen: 3.059 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 3470
Ion Ratio Lower Upper
173 100
175 45.8 23.4 70.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion:152 Resp: 67939
Ion Ratio Lower Upper
152 100
115 66.8 46.9 140.7
150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 105 Resp: 2420
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 3.479 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	9170
Ion	Ratio	Lower	Upper
83	100		
131	8.3	5.0	14.9
85	61.7	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 2.799 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	75	Resp:	7928
Ion	Ratio	Lower	Upper
75	100		
77	40.6	20.5	61.5





#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 156 Resp: 585
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 3.336 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 18672

Ion Ratio Lower Upper

91 100





126 29.7 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 3.372 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20742

Ion Ratio Lower Upper

105 100


120 45.9 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 2.686 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 1828
Ion Ratio Lower Upper
75 100
53 125.8 100.5 150.7
89 47.9 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 3.245 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
Ion Ratio Lower Upper
91 100
126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	119	Resp:	21049
Ion	Ratio	Lower	Upper
119	100		
91	67.0	32.9	98.7
134	23.8	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	105	Resp:	20613
Ion	Ratio	Lower	Upper
105	100		
120	44.5	21.2	63.6





#85

sec-Butylbenzene

Concen: 3.412 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:105 Resp: 25590

Ion Ratio Lower Upper

105 100





134 19.1 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 3.460 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:119 Resp: 20955


Ion Ratio Lower Upper

119 100

134 25.6 12.5 37.5

91 28.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 3.315 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:146	Resp:	1058
Ion Ratio	Lower	Upper
146	100	
111	46.2	22.1 66.3
148	62.4	32.1 96.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 3.486 ug/l




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:146	Resp:	10908
Ion Ratio	Lower	Upper
146	100	
111	52.1	21.3 63.9
148	69.4	31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	91	Resp:	1800
Ion	Ratio	Lower	Upper
91	100		
92	49.5	26.9	80.7
134	21.2	11.8	35.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 3.153 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	117	Resp:	3553
Ion	Ratio	Lower	Upper
117	100		
201	58.5	31.6	94.7





#91
1,2-Dichlorobenzene
Concen: 3.444 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 146 Resp: 1071
Ion Ratio Lower Upper
146 100
111 44.7 22.4 67.3
148 64.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 2.812 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion: 75 Resp: 1687
Ion Ratio Lower Upper
75 100
155 71.7 34.9 104.8
157 87.0 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 3.359 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:180	Resp:	5720
Ion Ratio	Lower	Upper
180	100	
182	91.6	47.9 143.8
145	35.7	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 3.535 ug/l




RT: 13.719 min Scan# 2072


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:225	Resp:	2816
Ion Ratio	Lower	Upper
225	100	
223	65.1	30.8 92.4
227	58.8	30.9 92.7





#95

Naphthalene

Concen: 3.209 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	128	Resp:	1990
Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.4	15.6
129	11.5	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 3.217 ug/l




RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	180	Resp:	5745
Ion	Ratio	Lower	Upper
180	100		
182	99.1	47.8	143.3
145	34.8	18.1	54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion:168 Resp: 92040
Ion Ratio Lower Upper
168 100
99 71.5 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 0.596 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 85 Resp: 1211
Ion Ratio Lower Upper
85 100
87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#









#6
Chloroethane
Concen: 0.951 ug/l
RT: 1.678 min Scan# 91
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 64 Resp: 859
Ion Ratio Lower Upper
64 100
66 35.2 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#7
Trichlorofluoromethane
Concen: 0.734 ug/l
RT: 1.880 min Scan# 130
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1959
Ion Ratio Lower Upper
101 100
103 67.8 51.0 76.4









#8
Diethyl Ether
Concen: 0.845 ug/l
RT: 2.130 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 74 Resp: 74
Ion Ratio Lower Upper
74 100
45 123.6 54.9 164.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.731 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1166
Ion Ratio Lower Upper
101 100
85 47.3 38.6 58.0
151 73.1 55.2 82.8





#12

1,1-Dichloroethene

Concen: 0.704 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001





Tgt Ion:	96	Resp:	109
Ion	Ratio	Lower	Upper
96	100		
61	199.6	146.2	219.2
98	49.6	51.0	76.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 0.660 ug/l




RT: 2.654 min Scan# 257


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1936
Ion	Ratio	Lower	Upper
41	100		
39	85.5	60.6	90.8
76	26.4	24.9	37.3





#15

Acrylonitrile

Concen: 3.424 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 53	Resp: 334
Ion Ratio	Lower Upper
53 100	
52 82.9	65.3 97.9
51 34.4	29.8 44.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 3.736 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 43	Resp: 3500
Ion Ratio	Lower Upper
43 100	
58 28.1	21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100





78 8.0 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 0.662 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 358

Ion Ratio Lower Upper

73 100





57 20.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 0.828 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 84 Resp: 1571




Ion Ratio Lower Upper


84 100

49 140.0 113.9 170.9

51 51.3 33.5 50.3#

86 85.2 53.8 80.8#





#21

trans-1,2-Dichloroethene

Concen: 0.701 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 96 Resp: 111

Ion Ratio Lower Upper

96 100

61 119.0 119.5 179.3





98 30.1 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 0.703 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 45 Resp: 3857




Ion Ratio Lower Upper


45 100

43 176.2 66.6 100.0#

87 28.9 19.8 29.6

59 9.1 8.6 12.8





#23

Vinyl Acetate

Concen: 3.154 ug/l

RT: 3.727 min Scan# 413

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 1528

Ion Ratio Lower Upper

43

100

86

10.8

7.5





11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 0.637 ug/l

RT: 3.605 min Scan# 413

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054

Ion Ratio Lower Upper

63

100

98

6.3

3.6




10.8


100

5.8

2.1

6.3





#25

2-Butanone

Concen: 3.371 ug/l

RT: 4.593 min Scan# 51

Delta R.T. 0.037 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 455

Ion Ratio Lower Upper

43 100





72 19.2 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 0.728 ug/l

RT: 4.458 min Scan# 553

Delta R.T. -0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 77 Resp: 1777

Ion Ratio Lower Upper

77 100

97 13.7 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 0.693 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X





ClientSampleId :

VSTDICC001

Tgt Ion: 96	Resp: 1324
Ion Ratio	Lower Upper
96 100	
61 145.5	0.0 322.8
98 63.1	0.0 129.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 0.619 ug/l m




RT: 4.910 min Scan# 627


Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 49	Resp: 1061
Ion Ratio	Lower Upper
49 100	
129 0.0	0.0 4.0
130 65.3	56.2 84.2





#29

Tetrahydrofuran

Concen: 3.338 ug/l

RT: 5.019 min Scan# 64

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X





ClientSampleId :

VSTDIC001

Tgt Ion:	42	Resp:	293
Ion	Ratio	Lower	Upper
42	100		
72	41.9	33.7	50.5
71	37.7	31.4	47.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 0.695 ug/l




RT: 5.093 min Scan# 657


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	83	Resp:	2328
Ion	Ratio	Lower	Upper
83	100		
85	60.4	49.3	73.9





#32

1,1,1-Trichloroethane

Concen: 0.647 ug/l

RT: 5.379 min Scan# 70

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 186

Ion Ratio Lower Upper

97 100

99 0.0 51.8 77.6





61 55.0 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 5.95 min

Lab File: VX046047.D


Acq: 05 May 2025 16:27


Tgt Ion: 65

Sig Exp Ratio

65 100

67 49.5





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:114 Resp: 16702

Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2





88 17.1 0.0 33.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 5.38 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.9

192 16.6





#36

1,1-Dichloropropene

Concen: 0.758 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	164
Ion	Ratio	Lower	Upper
75	100		
110	28.2	16.3	48.9
77	15.8	24.3	36.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#37

Ethyl Acetate

Concen: 0.717 ug/l m




RT: 4.733 min Scan# 598


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	43	Resp:	1956
Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.3	15.5#
70	1.0	7.9	11.9#





#38

Carbon Tetrachloride

Concen: 0.724 ug/l

RT: 5.678 min Scan# 752

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 117 Resp: 180

Ion Ratio Lower Upper

117 100

119 67.0 75.2 112.8





121 28.5 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#39

Methylcyclohexane

Concen: 0.770 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 83 Resp: 2096


Ion Ratio Lower Upper

83 100

55 80.1 64.7 97.1

98 41.9 37.4 56.2





#40

Benzene

Concen: 0.668 ug/l

RT: 6.044 min Scan# 811

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 450

Ion Ratio Lower Upper

78 100





77 29.1 19.0 28.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#41

Methacrylonitrile

Concen: 0.513 ug/l

RT: 4.946 min Scan# 633

Delta R.T. 0.030 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 41 Resp: 778




Ion Ratio Lower Upper


41 100

39 51.7 47.2 70.8

67 0.0 50.7 76.1#

52 13.6 24.6 37.0#





#42

1,2-Dichloroethane

Concen: 0.694 ug/l

RT: 6.092 min Scan# 819

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 62 Resp: 1934
Ion Ratio Lower Upper
62 100
98 1.1 0.0 15.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#43

Isopropyl Acetate

Concen: 0.612 ug/l




RT: 6.354 min Scan# 864


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 2553
Ion Ratio Lower Upper
43 100
61 10.7 14.3 21.5#
87 3.2 9.5 14.3#





#44

Trichloroethene

Concen: 0.680 ug/l

RT: 7.135 min Scan# 990

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:130 Resp: 1083

Ion Ratio Lower Upper

130 100





95 117.3 0.0 204.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#45

1,2-Dichloropropane

Concen: 0.633 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 63 Resp: 1059

Ion Ratio Lower Upper

63 100


65 36.1 24.1 36.1





#46
Dibromomethane
Concen: 0.667 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion: 93 Resp: 880
Ion Ratio Lower Upper
93 100
95 84.5 65.6 98.4
174 88.5 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#47
Bromodichloromethane
Concen: 0.633 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 83 Resp: 1620
Ion Ratio Lower Upper
83 100
85 57.3 50.5 75.7
127 0.0 6.5 9.7#





#48

Methyl methacrylate

Concen: 0.577 ug/l

RT: 7.714 min Scan# 1087

Delta R.T. 0.024 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	41	Resp:	123
Ion	Ratio	Lower	Upper
41	100		
69	70.6	58.5	87.7
39	84.8	51.7	77.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#49

1,4-Dioxane

Concen: 11.703 ug/l




RT: 7.677 min Scan# 1081


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	88	Resp:	498
Ion	Ratio	Lower	Upper
88	100		
43	44.4	33.4	50.2
58	30.3	58.6	88.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.9

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3









#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1251
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#53
t-1,3-Dichloropropene
Concen: 0.576 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 1238
Ion Ratio Lower Upper
75 100
77 42.7 25.0 37.6#





#54

cis-1,3-Dichloropropene

Concen: 0.564 ug/l

RT: 8.379 min Scan# 1193

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 75	Resp: 1414
Ion Ratio	Lower Upper
75	100
77	44.2 25.4 38.0
39	59.0 52.2 78.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#55

1,1,2-Trichloroethane

Concen: 0.637 ug/l




RT: 9.153 min Scan# 1323


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 97	Resp: 1029
Ion Ratio	Lower Upper
97	100
83	74.8 69.0 103.6
85	59.5 44.3 66.5
99	63.0 47.8 71.6





#56

Ethyl methacrylate

Concen: 0.504 ug/l

RT: 9.128 min Scan# 1319

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 69 Resp: 1260

Ion Ratio Lower Upper

69 100

41 86.6 60.8 91.2





39 78.0 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5





#58

2-Chloroethyl Vinyl ether

Concen: 3.420 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 63 Resp: 3840

Ion Ratio Lower Upper

63 100





106 24.0 21.5 32.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#59

2-Hexanone

Concen: 3.096 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 6424

Ion Ratio Lower Upper

43 100

58 41.6 24.9 74.6





#60

Dibromochloromethane

Concen: 0.582 ug/l

RT: 9.525 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion:129 Resp: 102

Ion Ratio Lower Upper

129 100





127 72.5 39.3 117.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#61

1,2-Dibromoethane

Concen: 0.650 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:107 Resp: 1077

Ion Ratio Lower Upper

107 100

109 94.0 73.3 109.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 95
Sig Exp Ratio
95 100
174 67.9
176 65.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6





#64
Tetrachloroethene
Concen: 0.679 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:164 Resp: 972
Ion Ratio Lower Upper
164 100
166 119.1 105.0 157.6
129 103.4 83.5 125.3
131 95.4 76.5 114.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#65
Chlorobenzene
Concen: 0.735 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion:112 Resp: 3170
Ion Ratio Lower Upper
112 100
114 30.6 25.4 38.2









#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#67
Ethyl Benzene
Concen: 0.687 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
106 32.5 23.4 35.2









#68
m/p-Xylenes
Concen: 1.367 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion:106 Resp: 3630
Ion Ratio Lower Upper
106 100
91 209.8 171.2 256.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1





#70

Styrene

Concen: 0.622 ug/l

RT: 10.659 min Scan# 1569

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:104	Resp:	266
Ion Ratio	Lower	Upper
104	100	
78	54.3	45.7 68.5
103	53.9	43.7 65.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#71

Bromoform

Concen: 0.608 ug/l




RT: 10.799 min Scan# 1593


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:173	Resp:	657
Ion Ratio	Lower	Upper
173	100	
175	46.6	23.4 70.0
254	0.0	0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion:152 Resp: 5912

Ion Ratio Lower Upper

152 100

115 66.6 46.9 140.7





150 156.3 0.0 351.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#73

Isopropylbenzene

Concen: 0.695 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:105 Resp: 4480

Ion Ratio Lower Upper

105 100


120 23.7 12.8 38.4





#74
N-amy1 acetate
Concen: 0.631 ug/l
RT: 10.854 min Scan# 1602
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion	Ratio	Lower	Upper
43	100		
70	31.5	30.9	46.3
55	24.5	18.7	28.1
61	15.4	17.1	25.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#75
1,1,2,2-Tetrachloroethane
Concen: 0.800 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion	Ratio	Lower	Upper
83	100		
131	7.5	5.0	14.9
85	58.7	31.9	95.7









#76
1,2,3-Trichloropropane
Concen: 0.674 ug/l m
RT: 11.244 min Scan# 1665
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 75 Resp: 166
Ion Ratio Lower Upper
75 100
77 29.7 20.5 61.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3





#78

n-propylbenzene

Concen: 0.699 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 505

Ion Ratio Lower Upper

91 100





120 20.1 10.8 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#79

2-Chlorotoluene

Concen: 0.773 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3765

Ion Ratio Lower Upper

91 100

126 28.0 15.6 46.7





#80

1,3,5-Trimethylbenzene

Concen: 0.671 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 3590

Ion Ratio Lower Upper

105 100





120 43.4 23.1 69.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#82

4-Chlorotoluene

Concen: 0.713 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3815

Ion Ratio Lower Upper

91 100

126 24.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 0.751 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	119	Resp:	395
Ion	Ratio	Lower	Upper
119	100		
91	63.9	32.9	98.7
134	25.2	11.4	34.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.697 ug/l




RT: 11.756 min Scan# 1750


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	105	Resp:	3725
Ion	Ratio	Lower	Upper
105	100		
120	40.3	21.2	63.6





#85

sec-Butylbenzene

Concen: 0.672 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 438

Ion Ratio Lower Upper

105 100





134 19.2 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 0.679 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:119 Resp: 3577


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 33.1 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 0.689 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:146	Resp:	1914
Ion Ratio	Lower	Upper
146	100	
111	43.9	22.1 66.3
148	65.3	32.1 96.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 0.789 ug/l m




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:146	Resp:	2149
Ion Ratio	Lower	Upper
146	100	
111	39.1	21.3 63.9
148	58.2	31.9 95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC001

Tgt Ion: 91	Resp: 288
Ion Ratio	Lower Upper
91	100
92	48.1 26.9 80.7
134	20.1 11.8 35.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 0.616 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 117	Resp: 604
Ion Ratio	Lower Upper
117	100
201	64.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion: 146 Resp: 202
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8

Manual Integrations
APPROVED



Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180	Resp:	1019
Ion Ratio	Lower	Upper
180	100	
182	100.6	47.9 143.8
145	29.5	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 0.671 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion:225	Resp:	465
Ion Ratio	Lower	Upper
225	100	
223	69.7	30.8 92.4
227	70.5	30.9 92.7





#95
Naphthalene
Concen: 0.767 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:128 Resp: 4138
Ion Ratio Lower Upper
128 100
127 12.4 10.4 15.6
129 8.9 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.716 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion:180 Resp: 1113
Ion Ratio Lower Upper
180 100
182 92.5 47.8 143.3
145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.020%
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79763m	48.568	ug/l	
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:168 Resp: 9258

Ion Ratio Lower Upper

168 100





99 74.5 54.9 82.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#2

Dichlorodifluoromethane

Concen: 55.594 ug/l

RT: 1.167 min Scan# 13

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 85 Resp: 78783

Ion Ratio Lower Upper

85 100

87 32.2 15.7 47.1









#3
Chloromethane
Concen: 52.592 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 50 Resp: 72275
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 51.598 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 62 Resp: 65994
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8









#5
Bromomethane
Concen: 50.627 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 94 Resp: 3003
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 52.271 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 64 Resp: 35690
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5









#7
Trichlorofluoromethane
Concen: 52.560 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:101 Resp: 9935
Ion Ratio Lower Upper
101 100
103 62.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 49.536 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
Ion Ratio Lower Upper
74 100
45 107.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.822 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:101 Resp: 5945
Ion Ratio Lower Upper
101 100
85 47.8 38.6 58.0
151 69.4 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 53.419 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:142 Resp: 73940
Ion Ratio Lower Upper
142 100
127 52.9 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 252.025 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 59 Resp: 6106

Ion Ratio Lower Upper

59 100





57 11.1 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 51.097 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 96 Resp: 56097


Ion Ratio Lower Upper

96 100

61 179.7 146.2 219.2

98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 56 Resp: 6415

Ion Ratio Lower Upper

56 100





55 70.8 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 262.412 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 53 Resp: 181808

Ion Ratio Lower Upper

53 100

52 82.9 65.3 97.9





51 36.5 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 245.518 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 169923

Ion Ratio Lower Upper

43 100

58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 50.981 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 76 Resp: 13269

Ion Ratio Lower Upper

76 100





78 9.1 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 49.968 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 80251

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 50.777 ug/l

RT: 3.111 min Scan# 332

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 73 Resp: 19544

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 47.779 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369




Ion Ratio Lower Upper


84 100

49 142.7 113.9 170.9

51 43.3 33.5 50.3

86 66.8 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 50.513 ug/l

RT: 3.087 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 96 Resp: 55770

Ion Ratio Lower Upper

96 100

61 149.3 119.5 179.3





98 64.6 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 51.800 ug/l

RT: 3.757 min Scan# 438

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946




Ion Ratio Lower Upper


45 100

43 86.9 66.6 100.0

87 24.8 19.8 29.6

59 10.8 8.6 12.8









#23
Vinyl Acetate
Concen: 265.585 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 43 Resp: 946740
Ion Ratio Lower Upper
43 100
86 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 50.825 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 63 Resp: 114733
Ion Ratio Lower Upper
63 100
98 6.4 3.6 10.8
100 4.8 2.1 6.3





#25

2-Butanone

Concen: 257.110 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 258339

Ion Ratio Lower Upper

43 100





72 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 77 Resp: 88229

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.956 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 96 Resp: 6772

Ion Ratio Lower Upper

96 100

61 160.7 0.0 322.8





98 64.4 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 50.975 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 49 Resp: 55390


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 68.7 56.2 84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	42	Resp:	16367
Ion	Ratio	Lower	Upper
42	100		
72	42.0	33.7	50.5
71	38.5	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 51.332 ug/l




RT: 5.093 min Scan# 657


Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50


Tgt Ion:	83	Resp:	120782
Ion	Ratio	Lower	Upper
83	100		
85	63.6	49.3	73.9





#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 65 Resp: 8373

Ion Ratio Lower Upper

65 100





67 50.0 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.683 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion:	113	Resp:	5819
Ion	Ratio	Lower	Upper
113	100		
111	102.1	83.1	124.7
192	16.2	13.3	19.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 49.521 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	75	Resp:	77931
Ion	Ratio	Lower	Upper
75	100		
110	32.7	16.3	48.9
77	31.1	24.3	36.5





#37

Ethyl Acetate

Concen: 50.714 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion:	43	Resp:	98604
Ion Ratio	Lower	Upper	
43	100		
61	12.6	10.3	15.5
70	10.0	7.9	11.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 49.825 ug/l




RT: 5.666 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	117	Resp:	88100
Ion Ratio	Lower	Upper	
117	100		
119	90.7	75.2	112.8
121	29.4	24.2	36.4





#39

Methylcyclohexane

Concen: 49.757 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 83 Resp: 100808

Ion Ratio Lower Upper

83 100

55 82.7 64.7 97.1





98 48.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 50.759 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 78 Resp: 233977

Ion Ratio Lower Upper

78 100


77 24.0 19.0 28.4





#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion: 41 Resp: 5601
Ion Ratio Lower Upper
41 100
39 58.7 47.2 70.8
67 63.7 50.7 76.1
52 30.5 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 62 Resp: 101185
Ion Ratio Lower Upper
62 100
98 7.1 0.0 15.2





#43

Isopropyl Acetate

Concen: 52.195 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 154818

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5





87 11.4 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 50.018 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:130 Resp: 55491

Ion Ratio Lower Upper

130 100

95 102.6 0.0 204.2





#45

1,2-Dichloropropane

Concen: 51.983 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 63 Resp: 5958

Ion Ratio Lower Upper

63 100





65 31.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 49.978 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 93 Resp: 45181


Ion Ratio Lower Upper

93 100

95 83.0 65.6 98.4

174 84.8 68.2 102.2





#47

Bromodichloromethane

Concen: 51.889 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:	83	Resp:	9238
Ion	Ratio	Lower	Upper
83	100		
85	63.0	50.5	75.7
127	8.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 54.226 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	41	Resp:	82144
Ion	Ratio	Lower	Upper
41	100		
69	70.4	58.5	87.7
39	64.2	51.7	77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVX050525

Tgt Ion:	88	Resp:	3047
Ion	Ratio	Lower	Upper
88	100		
43	39.9	33.4	50.2
58	76.2	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 48.984 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	98	Resp:	198575
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 43 Resp: 510200
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.920 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 75 Resp: 8375

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 52.892 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 92517


Ion Ratio Lower Upper

75 100

77 31.5 25.4 38.0

39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion	Ratio	Lower	Upper
97	100		
83	89.8	69.0	103.6
85	58.4	44.3	66.5
99	63.9	47.8	71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 53.918 ug/l




RT: 9.116 min Scan# 1317


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion	Ratio	Lower	Upper
69	100		
41	76.1	60.8	91.2
39	47.4	39.0	58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 76 Resp: 9966

Ion Ratio Lower Upper

76 100





78 31.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3





#59

2-Hexanone

Concen: 268.218 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 39070

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 52.075 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:129 Resp: 63738

Ion Ratio Lower Upper

129 100

127 77.9 39.3 117.8





#61

1,2-Dibromoethane

Concen: 51.050 ug/l

RT: 9.610 min Scan# 1397

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVX050525

Tgt Ion: 107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 48.596 ug/l




RT: 11.079 min Scan# 1639


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion:	117	Resp:	145078
Ion	Ratio	Lower	Upper
117	100		
82	62.1	50.6	76.0
119	32.9	25.8	38.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#64

Tetrachloroethene

Concen: 46.795 ug/l




RT: 9.269 min Scan# 1342


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50


Tgt Ion:	164	Resp:	48034
Ion	Ratio	Lower	Upper
164	100		
166	129.9	105.0	157.6
129	105.3	83.5	125.3
131	98.8	76.5	114.7





#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67

Ethyl Benzene

Concen: 50.245 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 91 Resp: 281239

Ion Ratio Lower Upper

91 100





106 29.1 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 101.022 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:106 Resp: 206813

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8









#69
o-Xylene
Concen: 50.841 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:106 Resp: 101469
Ion Ratio Lower Upper
106 100
91 227.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 51.907 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:104 Resp: 169703
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 49.583 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:173 Resp: 40364
Ion Ratio Lower Upper
173 100
175 48.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:152 Resp: 68229
Ion Ratio Lower Upper
152 100
115 93.3 46.9 140.7
150 175.3 0.0 351.0









#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 47.827 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	83	Resp:	8902
Ion	Ratio	Lower	Upper
83	100		
131	9.7	5.0	14.9
85	65.5	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 48.568 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50


Tgt Ion:	75	Resp:	79763
Ion	Ratio	Lower	Upper
75	100		
77	40.8	20.5	61.5





#77
Bromobenzene
Concen: 48.549 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525




Tgt Ion:156 Resp: 5987
Ion Ratio Lower Upper
156 100
77 189.0 92.8 278.3
158 98.9 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 50.562 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50



Tgt Ion: 91 Resp: 312285
Ion Ratio Lower Upper
91 100
120 21.4 10.8 32.4





#79

2-Chlorotoluene

Concen: 49.107 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 91 Resp: 195630

Ion Ratio Lower Upper

91 100





126 30.8 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 51.500 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 228538

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 75 Resp: 2485
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.511 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVXX050525

Tgt Ion:119 Resp: 22578

Ion Ratio Lower Upper

119 100

91 64.9 32.9 98.7





134 22.6 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 51.485 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371

Ion Ratio Lower Upper

105 100

120 42.3 21.2 63.6









#85
sec-Butylbenzene
Concen: 51.122 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:105 Resp: 280578
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 51.198 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:119 Resp: 231942
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.5 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 50.272 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:146 Resp: 113140

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 48.368 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:146 Resp: 111173


Ion Ratio Lower Upper

146 100

111 43.9 21.3 63.9

148 63.3 31.9 95.5





#89

n-Butylbenzene

Concen: 51.804 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 91 Resp: 20585

Ion Ratio Lower Upper

91 100

92 53.7 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 49.895 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:117 Resp: 39823

Ion Ratio Lower Upper

117 100


201 62.4 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 49.361 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525




Tgt Ion:146 Resp: 11148
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 64.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 51.458 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 21220
Ion Ratio Lower Upper
75 100
155 68.6 34.9 104.8
157 86.7 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 49.662 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:180 Resp: 6442
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.4 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 48.408 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:225 Resp: 27425
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.4
227 63.7 30.9 92.7





#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:128 Resp: 23980
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.765	0.851	-11.2	94	0.00
3 P	Chloromethane	0.742	0.781	-5.3	96	0.00
4 C	Vinyl Chloride	0.691	0.713	-3.2#	96	0.00
5 T	Bromomethane	0.320	0.324	-1.3	95	0.00
6 T	Chloroethane	0.369	0.385	-4.3	97	0.00
7 T	Trichlorofluoromethane	1.021	1.073	-5.1	96	0.00
8 T	Diethyl Ether	0.347	0.344	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.642	-1.6	96	0.00
10 T	Methyl Iodide	0.747	0.799	-7.0	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	98	0.00
12 CM	1,1-Dichloroethene	0.593	0.606	-2.2#	96	0.00
13 T	Acrolein	0.149	0.139	6.7	87	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	95	0.00
15 T	Acrylonitrile	0.374	0.393	-5.1	97	0.00
16 T	Acetone	0.374	0.367	1.9	97	0.00
17 T	Carbon Disulfide	1.406	1.433	-1.9	94	0.00
18 T	Methyl Acetate	0.867	0.867	0.0	97	0.00
19 T	Methyl tert-butyl Ether	2.079	2.111	-1.5	93	0.00
20 T	Methylene Chloride	0.716	0.684	4.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	94	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	95	0.00
23 T	Vinyl Acetate	1.925	2.045	-6.2	95	0.00
24 P	1,1-Dichloroethane	1.219	1.239	-1.6	94	0.00
25 T	2-Butanone	0.543	0.558	-2.8	96	0.00
26 T	2,2-Dichloropropane	0.954	0.953	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.731	-1.8	95	0.00
28 T	Bromochloromethane	0.587	0.598	-1.9	99	0.00
29 T	Tetrahydrofuran	0.340	0.354	-4.1	96	0.00
30 C	Chloroform	1.271	1.305	-2.7#	96	0.00
31 T	Cyclohexane	1.111	1.114	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	1.101	1.126	-2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.904	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.360	0.358	0.6	98	0.00
36 T	1,1-Dichloropropene	0.484	0.479	1.0	94	0.00
37 T	Ethyl Acetate	0.598	0.606	-1.3	96	0.00
38 T	Carbon Tetrachloride	0.544	0.542	0.4	94	0.00
39 T	Methylcyclohexane	0.623	0.620	0.5	94	0.00
40 TM	Benzene	1.417	1.439	-1.6	95	0.00
41 T	Methacrylonitrile	0.313	0.344	-9.9	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.622	-1.6	96	0.00
43 T	Isopropyl Acetate	0.912	0.952	-4.4	96	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	93	0.00
45 C	1,2-Dichloropropane	0.352	0.366	-4.0#	96	0.00
46 T	Dibromomethane	0.278	0.278	0.0	94	0.00
47 T	Bromodichloromethane	0.547	0.568	-3.8	95	0.00
48 T	Methyl methacrylate	0.466	0.505	-8.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.246	1.221	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.627	-3.6	96	0.00
52 CM	Toluene	0.869	0.889	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.515	-5.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.569	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.343	0.349	-1.7	95	0.00
56 T	Ethyl methacrylate	0.546	0.589	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.613	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.296	-6.5	93	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	98	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	95	0.00
61 T	1,2-Dibromoethane	0.356	0.364	-2.2	94	0.00
62 S	4-Bromofluorobenzene	0.478	0.465	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.331	6.5	88	0.00
65 PM	Chlorobenzene	1.094	1.066	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.370	1.1	94	0.00
67 C	Ethyl Benzene	1.929	1.939	-0.5#	95	0.00
68 T	m/p-Xylenes	0.706	0.713	-1.0	95	0.00
69 T	o-Xylene	0.688	0.699	-1.6	95	0.00
70 T	Styrene	1.127	1.170	-3.8	95	0.00
71 P	Bromoform	0.281	0.278	1.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.893	3.950	-1.5	96	0.00
74 T	N-amyl acetate	1.924	2.016	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.305	4.3	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.169	2.9	99	0.00
77 T	Bromobenzene	0.904	0.878	2.9	95	0.00
78 T	n-propylbenzene	4.526	4.577	-1.1	95	0.00
79 T	2-Chlorotoluene	2.919	2.867	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.350	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.364	1.6	95	0.00
82 T	4-Chlorotoluene	3.238	3.280	-1.3	96	0.00
83 T	tert-Butylbenzene	3.276	3.309	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.391	-3.0	97	0.00
85 T	sec-Butylbenzene	4.022	4.112	-2.2	96	0.00
86 T	p-Isopropyltoluene	3.320	3.399	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	1.649	1.658	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	1.684	1.629	3.3	97	0.00
89 T	n-Butylbenzene	2.912	3.017	-3.6	96	0.00
90 T	Hexachloroethane	0.585	0.584	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.655	1.634	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.311	-3.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.944	0.7	97	0.00
94 T	Hexachlorobutadiene	0.415	0.402	3.1	95	0.00
95 T	Naphthalene	3.487	3.515	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	0.981	0.952	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-amyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2201 SAS No.: Q2201 SDG No.: Q2201

Instrument ID: MSVOA_X Calibration Date/Time: 06/04/2025 10:12

Lab File ID: VX046488.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.665		-3.76	20
1,1-Dichloroethene	0.593	0.593		0	20
1,1-Dichloroethane	1.219	1.302	0.1	6.81	20
cis-1,2-Dichloroethene	0.718	0.747		4.04	20
1,1,1-Trichloroethane	1.101	1.161		5.45	20
Benzene	1.417	1.496		5.57	20
1,2-Dichloroethane	0.612	0.651		6.37	20
Trichloroethene	0.341	0.364		6.74	20
1,1,2-Trichloroethane	0.343	0.370		7.87	20
Tetrachloroethene	0.354	0.378		6.78	20
1,2-Dichloroethane-d4	0.932	0.862		-7.51	20
Dibromofluoromethane	0.360	0.356		-1.11	20
Toluene-d8	1.246	1.135		-8.91	20
4-Bromofluorobenzene	0.478	0.469		-1.88	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046488.D
 Acq On : 04 Jun 2025 10:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.543	168	97475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	165033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	141151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83988	46.217	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.440%		
35) Dibromofluoromethane	5.373	113	58786	49.466	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.940%		
50) Toluene-d8	8.646	98	187253	45.524	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.040%		
62) 4-Bromofluorobenzene	11.079	95	77443	49.083	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.160%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	73169	49.043	ug/l	96
3) Chloromethane	1.306	50	69991	48.377	ug/l	98
4) Vinyl Chloride	1.373	62	64836	48.151	ug/l	100
5) Bromomethane	1.599	94	27003	43.237	ug/l	91
6) Chloroethane	1.672	64	36727	51.093	ug/l	98
7) Trichlorofluoromethane	1.879	101	103198	51.858	ug/l	99
8) Diethyl Ether	2.135	74	34758	51.308	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.318	101	65312	53.034	ug/l	99
10) Methyl Iodide	2.446	142	70274	48.225	ug/l	99
11) Tert butyl alcohol	2.971	59	71909	281.901	ug/l	99
12) 1,1-Dichloroethene	2.312	96	57779	49.990	ug/l	97
13) Acrolein	2.233	56	72503	249.577	ug/l	99
14) Allyl chloride	2.660	41	120766	54.671	ug/l	97
15) Acrylonitrile	3.062	53	196655	269.611	ug/l	98
16) Acetone	2.379	43	215178	295.318	ug/l	99
17) Carbon Disulfide	2.501	76	126198	46.054	ug/l	100
18) Methyl Acetate	2.702	43	116961	69.175	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	223672	55.198	ug/l	99
20) Methylene Chloride	2.782	84	67521	48.357	ug/l	95
21) trans-1,2-Dichloroethene	3.086	96	58349	50.199	ug/l	100
22) Diisopropyl ether	3.757	45	234578	54.976	ug/l	96
23) Vinyl Acetate	3.714	43	983901	262.170	ug/l	100
24) 1,1-Dichloroethane	3.605	63	126924	53.406	ug/l	100
25) 2-Butanone	4.550	43	295497	279.346	ug/l	99
26) 2,2-Dichloropropane	4.464	77	103262	55.512	ug/l	99
27) cis-1,2-Dichloroethene	4.476	96	72840	52.056	ug/l	98
28) Bromochloromethane	4.885	49	53251	46.549	ug/l	97
29) Tetrahydrofuran	5.001	42	181777	274.234	ug/l	98
30) Chloroform	5.086	83	130154	52.542	ug/l	96
31) Cyclohexane	5.458	56	107767	49.761	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	113144	52.691	ug/l	99
36) 1,1-Dichloropropene	5.677	75	83522	52.307	ug/l	99
37) Ethyl Acetate	4.708	43	105142	53.296	ug/l	99
38) Carbon Tetrachloride	5.671	117	96961	54.045	ug/l	99
39) Methylcyclohexane	7.372	83	107311	52.202	ug/l	97
40) Benzene	6.031	78	246896	52.789	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046488.D
 Acq On : 04 Jun 2025 10:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:34:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.909	41	61057m	59.164	ug/l	
42) 1,2-Dichloroethane	6.080	62	107364	53.188	ug/l	99
43) Isopropyl Acetate	6.336	43	172534	57.328	ug/l	100
44) Trichloroethene	7.116	130	60052	53.347	ug/l	93
45) 1,2-Dichloropropane	7.421	63	63728	54.797	ug/l	100
46) Dibromomethane	7.573	93	48150	52.494	ug/l	99
47) Bromodichloromethane	7.817	83	99979	55.341	ug/l	99
48) Methyl methacrylate	7.689	41	88716	57.719	ug/l	99
49) 1,4-Dioxane	7.653	88	31277	1071.717	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	559427	280.031	ug/l	100
52) Toluene	8.713	92	148931	51.931	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	91872	57.213	ug/l	98
54) cis-1,3-Dichloropropene	8.360	75	101352	57.106	ug/l	95
55) 1,1,2-Trichloroethane	9.146	97	61136	54.064	ug/l	98
56) Ethyl methacrylate	9.110	69	105633	58.611	ug/l	98
57) 1,3-Dichloropropane	9.305	76	107331	52.850	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	252660	274.981	ug/l	100
59) 2-Hexanone	9.427	43	423096	286.265	ug/l	100
60) Dibromochloromethane	9.518	129	69918	56.300	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62700	53.347	ug/l	97
64) Tetrachloroethene	9.268	164	53293	53.362	ug/l	93
65) Chlorobenzene	10.073	112	164933	53.386	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.158	131	57986	54.966	ug/l	99
67) Ethyl Benzene	10.189	91	298815	54.870	ug/l	100
68) m/p-Xylenes	10.299	106	215687	108.288	ug/l	97
69) o-Xylene	10.640	106	106763	54.982	ug/l	97
70) Styrene	10.652	104	181074	56.926	ug/l	100
71) Bromoform	10.798	173	44870	56.651	ug/l #	98
73) Isopropylbenzene	10.957	105	288460	53.686	ug/l	100
74) N-amyl acetate	10.841	43	147771	55.656	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	95562	50.752	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	86142m	51.854	ug/l	
77) Bromobenzene	11.195	156	65831	52.773	ug/l	98
78) n-propylbenzene	11.298	91	338759	54.223	ug/l	100
79) 2-Chlorotoluene	11.359	91	207728	51.549	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	241792	53.865	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	28928	56.693	ug/l	96
82) 4-Chlorotoluene	11.451	91	240706	53.864	ug/l	99
83) tert-Butylbenzene	11.713	119	240573	53.206	ug/l	100
84) 1,2,4-Trimethylbenzene	11.749	105	244850	53.863	ug/l	99
85) sec-Butylbenzene	11.890	105	304738	54.891	ug/l	99
86) p-Isopropyltoluene	12.006	119	251225	54.822	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	119595	52.532	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	120210	51.703	ug/l	99
89) n-Butylbenzene	12.329	91	228216	56.775	ug/l	100
90) Hexachloroethane	12.536	117	44135	54.668	ug/l	98
91) 1,2-Dichlorobenzene	12.329	146	120071	52.558	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.938	75	22836	54.746	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	71028	54.131	ug/l	99
94) Hexachlorobutadiene	13.719	225	31559	55.070	ug/l	99
95) Naphthalene	13.774	128	258228	53.657	ug/l	100
96) 1,2,3-Trichlorobenzene	13.956	180	73266	54.114	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 97475
Ion Ratio Lower Upper
168 100
99 71.2 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#2
Dichlorodifluoromethane
Concen: 49.043 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 85 Resp: 73169
Ion Ratio Lower Upper
85 100
87 33.4 15.7 47.1









#3
Chloromethane
Concen: 48.377 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 6999
Ion Ratio Lower Upper
50 100
52 32.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#4
Vinyl Chloride
Concen: 48.151 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 62 Resp: 64836
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8









#5
Bromomethane
Concen: 43.237 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 2700
Ion Ratio Lower Upper
94 100
96 99.6 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#6
Chloroethane
Concen: 51.093 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion: 64 Resp: 36727
Ion Ratio Lower Upper
64 100
66 31.5 24.3 36.5





#7
Trichlorofluoromethane
Concen: 51.858 ug/l
RT: 1.879 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 101 Resp: 103198
Ion Ratio Lower Upper
101 100
103 64.4 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#8
Diethyl Ether
Concen: 51.308 ug/l
RT: 2.135 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12



Tgt Ion: 74 Resp: 34758
Ion Ratio Lower Upper
74 100
45 107.9 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.034 ug/l

RT: 2.318 min Scan# 20

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:101 Resp: 6531

Ion Ratio Lower Upper

101 100

85 46.2 38.6 58.0





151 69.1 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#10

Methyl Iodide

Concen: 48.225 ug/l

RT: 2.446 min Scan# 223

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion:142 Resp: 70274


Ion Ratio Lower Upper

142 100

127 51.6 41.7 62.5

141 14.5 11.5 17.3





#11

Tert butyl alcohol

Concen: 281.901 ug/l

RT: 2.971 min Scan# 30

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 71909

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#12

1,1-Dichloroethene

Concen: 49.990 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 96 Resp: 57779


Ion Ratio Lower Upper

96 100

61 187.2 146.2 219.2

98 65.4 51.0 76.6





#13

Acrolein

Concen: 249.577 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 7250

Ion Ratio Lower Upper

56 100





55 71.0 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#14

Allyl chloride

Concen: 54.671 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 41 Resp: 120766


Ion Ratio Lower Upper

41 100

39 73.0 60.6 90.8

76 29.8 24.9 37.3





#15

Acrylonitrile

Concen: 269.611 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 53 Resp: 19665

Ion Ratio Lower Upper

53 100

52 83.7 65.3 97.9





51 37.4 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#16

Acetone

Concen: 295.318 ug/l

RT: 2.379 min Scan# 212

Delta R.T. -0.007 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion: 43 Resp: 215178

Ion Ratio Lower Upper

43 100

58 27.0 21.2 31.8





#17

Carbon Disulfide

Concen: 46.054 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 126198

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#18

Methyl Acetate

Concen: 69.175 ug/l

RT: 2.702 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion: 43 Resp: 116961

Ion Ratio Lower Upper

43 100

74 21.5 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 55.198 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 223672

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#20

Methylene Chloride

Concen: 48.357 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion: 84 Resp: 67521




Ion Ratio Lower Upper


84 100

49 134.2 113.9 170.9

51 42.4 33.5 50.3

86 64.9 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 50.199 ug/l

RT: 3.086 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 58349

Ion Ratio Lower Upper

96 100

61 149.6 119.5 179.3





98 62.4 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#22

Diisopropyl ether

Concen: 54.976 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion: 45 Resp: 234578




Ion Ratio Lower Upper


45 100

43 78.9 66.6 100.0

87 24.6 19.8 29.6

59 10.8 8.6 12.8





#23

Vinyl Acetate

Concen: 262.170 ug/l

RT: 3.714 min Scan# 41

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDC050

Tgt Ion: 43 Resp: 98390

Ion Ratio Lower Upper

43 100





86 9.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#24

1,1-Dichloroethane

Concen: 53.406 ug/l

RT: 3.605 min Scan# 413

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 63 Resp: 126924


Ion Ratio Lower Upper

63 100

98 7.0 3.6 10.8

100 4.3 2.1 6.3





#25

2-Butanone

Concen: 279.346 ug/l

RT: 4.550 min Scan# 5

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 29549

Ion Ratio Lower Upper

43

100

72

23.6

18.4





27.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#26

2,2-Dichloropropane

Concen: 55.512 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion: 77 Resp: 103262

Ion Ratio Lower Upper

77




100


97

20.5

10.5

31.5





#27

cis-1,2-Dichloroethene

Concen: 52.056 ug/l

RT: 4.476 min Scan# 557

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 72840

Ion Ratio Lower Upper





96 100

61 164.0 0.0 322.8

98 64.8 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#28

Bromochloromethane

Concen: 46.549 ug/l

RT: 4.885 min Scan# 623

Delta R.T. -0.013 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 49 Resp: 53251


Ion Ratio Lower Upper

49 100

129 1.6 0.0 4.0

130 67.3 56.2 84.2





#29

Tetrahydrofuran

Concen: 274.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 18177

Ion Ratio Lower Upper

42 100

72 41.3 33.7 50.5





71 38.1 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#30

Chloroform

Concen: 52.542 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion: 83 Resp: 130154

Ion Ratio Lower Upper

83 100

85 64.4 49.3 73.9





#31

Cyclohexane

Concen: 49.761 ug/l

RT: 5.458 min Scan# 717

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDC050

Tgt Ion: 56 Resp: 10776

Ion Ratio Lower Upper

56 100

69 31.3 24.4 36.6





84 84.7 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#32

1,1,1-Trichloroethane

Concen: 52.691 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 97 Resp: 113144


Ion Ratio Lower Upper

97 100

99 63.9 51.8 77.6

61 49.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 46.217 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 8398





Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.750 min Scan# 929

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion:114 Resp: 165033


Ion Ratio Lower Upper

114 100

63 24.4 0.0 49.2

88 18.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.466 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 5878

Ion Ratio Lower Upper

113 100

111 104.7 83.1 124.7





192 17.4 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#36

1,1-Dichloropropene

Concen: 52.307 ug/l

RT: 5.677 min Scan# 753

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 75 Resp: 83522

Ion Ratio Lower Upper


75 100

110 32.3 16.3 48.9

77 31.1 24.3 36.5



Abundance Scan 595 (4.715 min): VX046042.D\data.ms (-58



#37

Ethyl Acetate

Concen: 53.296 ug/l

RT: 4.708 min Scan# 594

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 10514

Ion Ratio Lower Upper

43 100

61 12.7 10.3 15.5


70 9.7 7.9 11.9

Manual Integrations



APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025


Supervised By :Semsettin Yesilyurt 06/05/2025



Abundance Scan 594 (4.708 min): VX046488.D\data.ms (-54



Abundance Scan 752 (5.672 min): VX046042.D\data.ms (-74



#38

Carbon Tetrachloride

Concen: 54.045 ug/l

RT: 5.671 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12


Tgt Ion:117 Resp: 96961

Ion Ratio Lower Upper



117 100


119 93.0 75.2 112.8

121 30.9 24.2 36.4



Abundance Scan 752 (5.671 min): VX046488.D\data.ms (-70





#39

Methylcyclohexane

Concen: 52.202 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 10731

Ion Ratio Lower Upper

83 100

55 83.4 64.7 97.1





98 49.4 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#40

Benzene

Concen: 52.789 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion: 78 Resp: 246896

Ion Ratio Lower Upper

78 100


77 23.9 19.0 28.4





#41
Methacrylonitrile
Concen: 59.164 ug/l m
RT: 4.909 min Scan# 610
Delta R.T. -0.007 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 6105
Ion Ratio Lower Upper
41 100
39 58.5 47.2 70.8
67 64.4 50.7 76.1
52 30.6 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#42
1,2-Dichloroethane
Concen: 53.188 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 62 Resp: 107364
Ion Ratio Lower Upper
62 100
98 7.3 0.0 15.2





#43

Isopropyl Acetate

Concen: 57.328 ug/l

RT: 6.336 min Scan# 861

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 172534

Ion Ratio Lower Upper

43

100

61

87

17.8

11.7

14.3

9.5

21.5





14.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#44

Trichloroethene

Concen: 53.347 ug/l

RT: 7.116 min Scan# 989

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion:130 Resp: 60052

Ion Ratio Lower Upper

130




100


95

109.6

0.0

204.2





#45

1,2-Dichloropropane

Concen: 54.797 ug/l

RT: 7.421 min Scan# 1039

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 63728

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#46

Dibromomethane

Concen: 52.494 ug/l

RT: 7.573 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 93 Resp: 48150


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 86.7 68.2 102.2





#47

Bromodichloromethane

Concen: 55.341 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDC050

Tgt Ion:	83	Resp:	99979
Ion	Ratio	Lower	Upper
83	100		
85	63.8	50.5	75.7
127	7.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#48

Methyl methacrylate

Concen: 57.719 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion:	41	Resp:	88716
Ion	Ratio	Lower	Upper
41	100		
69	72.6	58.5	87.7
39	62.9	51.7	77.5





#49

1,4-Dioxane

Concen: 1071.717 ug/l

RT: 7.653 min Scan# 1077

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tgt Ion:	88	Resp:	3127
Ion	Ratio	Lower	Upper
88	100		
43	41.5	33.4	50.2
58	75.7	58.6	88.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#50

Toluene-d8

Concen: 45.524 ug/l




RT: 8.646 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Tgt Ion:	98	Resp:	187253
Ion	Ratio	Lower	Upper
98	100		
100	66.0	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 280.031 ug/l
RT: 8.567 min Scan# 1227
Delta R.T. -0.007 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 55942
Ion Ratio Lower Upper
43 100
58 35.9 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#52
Toluene
Concen: 51.931 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 92 Resp: 148931
Ion Ratio Lower Upper
92 100
91 172.6 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 57.213 ug/l

RT: 8.976 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 91872

Ion Ratio Lower Upper

75 100





77 32.2 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#54

cis-1,3-Dichloropropene

Concen: 57.106 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 75 Resp: 101352


Ion Ratio Lower Upper

75 100

77 30.2 25.4 38.0

39 60.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 54.064 ug/l

RT: 9.146 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 97 Resp: 61130

Ion Ratio Lower Upper

97 100

83 86.2 69.0 103.6

85 54.9 44.3 66.5





99 63.3 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#56

Ethyl methacrylate

Concen: 58.611 ug/l

RT: 9.110 min Scan# 1316

Delta R.T. -0.007 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion: 69 Resp: 105633


Ion Ratio Lower Upper

69 100

41 77.1 60.8 91.2

39 46.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 52.850 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 10733

Ion Ratio Lower Upper

76 100





78 32.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#58

2-Chloroethyl Vinyl ether

Concen: 274.981 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion: 63 Resp: 252660

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3





#59

2-Hexanone

Concen: 286.265 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 423090

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#60

Dibromochloromethane

Concen: 56.300 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion:129 Resp: 69918

Ion Ratio Lower Upper

129 100

127 78.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 53.347 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 62700
Ion Ratio Lower Upper
107 100
109 94.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#62
4-Bromofluorobenzene
Concen: 49.083 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion: 95 Resp: 77443
Ion Ratio Lower Upper
95 100
174 66.9 0.0 135.8
176 65.0 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 14115
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 31.6 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#64
Tetrachloroethene
Concen: 53.362 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion:164 Resp: 53293
Ion Ratio Lower Upper
164 100
166 132.3 105.0 157.6
129 115.5 83.5 125.3
131 107.1 76.5 114.7





#65
Chlorobenzene
Concen: 53.386 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 16493
Ion Ratio Lower Upper
112 100
114 31.4 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 54.966 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion:131 Resp: 57986
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 64.5 31.6 95.0









#67
Ethyl Benzene
Concen: 54.870 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 298815
Ion Ratio Lower Upper
91 100
106 29.0 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#68
m/p-Xylenes
Concen: 108.288 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion:106 Resp: 215687
Ion Ratio Lower Upper
106 100
91 218.1 171.2 256.8









#69
o-Xylene
Concen: 54.982 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 106763
Ion Ratio Lower Upper
106 100
91 229.8 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#70
Styrene
Concen: 56.926 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion:104 Resp: 181074
Ion Ratio Lower Upper
104 100
78 57.5 45.7 68.5
103 54.9 43.7 65.5





#71
Bromoform
Concen: 56.651 ug/l
RT: 10.798 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 44870
Ion Ratio Lower Upper
173 100
175 48.2 23.4 70.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12



Tgt Ion:152 Resp: 69016
Ion Ratio Lower Upper
152 100
115 94.5 46.9 140.7
150 172.5 0.0 351.0









#73
Isopropylbenzene
Concen: 53.686 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 288460
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#74
N-ethyl acetate
Concen: 55.656 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 43 Resp: 147771
Ion Ratio Lower Upper
43 100
70 38.3 30.9 46.3
55 22.8 18.7 28.1
61 21.9 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 50.752 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 9556
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 64.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#76
1,2,3-Trichloropropane
Concen: 51.854 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion: 75 Resp: 86142
Ion Ratio Lower Upper
75 100
77 40.0 20.5 61.5





#77
Bromobenzene
Concen: 52.773 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 6583
Ion Ratio Lower Upper
156 100
77 181.5 92.8 278.3
158 96.5 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#78
n-propylbenzene
Concen: 54.223 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 91 Resp: 338759
Ion Ratio Lower Upper
91 100
120 21.5 10.8 32.4





#79

2-Chlorotoluene

Concen: 51.549 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 207728





Ion Ratio Lower Upper

91 100

126 31.1 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#80

1,3,5-Trimethylbenzene

Concen: 53.865 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion:105 Resp: 241792

Ion Ratio Lower Upper

105 100

120 46.4 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 56.693 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 28928
Ion Ratio Lower Upper
75 100
53 119.3 100.5 150.7
89 45.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#82
4-Chlorotoluene
Concen: 53.864 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12

Tgt Ion: 91 Resp: 240706
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 53.206 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 24057

Ion Ratio Lower Upper

119 100

91 65.6 32.9 98.7





134 22.7 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#84

1,2,4-Trimethylbenzene

Concen: 53.863 ug/l

RT: 11.749 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion:105 Resp: 244850

Ion Ratio Lower Upper

105 100


120 42.9 21.2 63.6





#85
sec-Butylbenzene
Concen: 54.891 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 304738
Ion Ratio Lower Upper
105 100
134 19.0 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#86
p-Isopropyltoluene
Concen: 54.822 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12



Tgt Ion:119 Resp: 251225
Ion Ratio Lower Upper
119 100
134 25.6 12.5 37.5
91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 52.532 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 11959

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.1 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#88

1,4-Dichlorobenzene

Concen: 51.703 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12




Tgt Ion:146 Resp: 120210


Ion Ratio Lower Upper

146 100

111 43.8 21.3 63.9

148 64.0 31.9 95.5





#89

n-Butylbenzene

Concen: 56.775 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046488.D

Acq: 04 Jun 2025 10:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 228210

Ion Ratio Lower Upper

91 100

92 53.9 26.9 80.7





134 24.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#90

Hexachloroethane

Concen: 54.668 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046488.D




Acq: 04 Jun 2025 10:12


Tgt Ion:117 Resp: 44135

Ion Ratio Lower Upper

117 100


201 64.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 52.558 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.007 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 12007
Ion Ratio Lower Upper
146 100
111 46.4 22.4 67.3
148 63.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 54.746 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion: 75 Resp: 22836
Ion Ratio Lower Upper
75 100
155 70.2 34.9 104.8
157 90.3 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 54.131 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:180 Resp: 71028
Ion Ratio Lower Upper
180 100
182 94.8 47.9 143.8
145 35.3 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#94
Hexachlorobutadiene
Concen: 55.070 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.007 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12


Tgt Ion:225 Resp: 31559
Ion Ratio Lower Upper
225 100
223 62.5 30.8 92.4
227 62.4 30.9 92.7





#95
Naphthalene
Concen: 53.657 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 258228
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 10.7 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#96
1,2,3-Trichlorobenzene
Concen: 54.114 ug/l
RT: 13.956 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046488.D
Acq: 04 Jun 2025 10:12



Tgt Ion:180 Resp: 73266
Ion Ratio Lower Upper
180 100
182 93.2 47.8 143.3
145 36.0 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.765	0.751	1.8	87	0.00
3 P	Chloromethane	0.742	0.718	3.2	93	0.00
4 C	Vinyl Chloride	0.691	0.665	3.8#	94	0.00
5 T	Bromomethane	0.320	0.277	13.4	85	0.00
6 T	Chloroethane	0.369	0.377	-2.2	100	0.00
7 T	Trichlorofluoromethane	1.021	1.059	-3.7	100	0.00
8 T	Diethyl Ether	0.347	0.357	-2.9	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.670	-6.0	105	0.00
10 T	Methyl Iodide	0.747	0.721	3.5	90	0.00
11 T	Tert butyl alcohol	0.131	0.148	-13.0	115	0.00
12 CM	1,1-Dichloroethene	0.593	0.593	0.0	99	0.00
13 T	Acrolein	0.149	0.149	0.0	98	0.00
14 T	Allyl chloride	1.133	1.239	-9.4	106	0.00
15 T	Acrylonitrile	0.374	0.403	-7.8	104	0.00
16 T	Acetone	0.374	0.442	-18.2	122	0.00
17 T	Carbon Disulfide	1.406	1.295	7.9	89	0.00
18 T	Methyl Acetate	0.867	1.200	-38.4#	142	0.00
19 T	Methyl tert-butyl Ether	2.079	2.295	-10.4	107	0.00
20 T	Methylene Chloride	0.716	0.693	3.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.599	-0.5	99	0.00
22 T	Diisopropyl ether	2.189	2.407	-10.0	106	0.00
23 T	Vinyl Acetate	1.925	2.019	-4.9	99	0.00
24 P	1,1-Dichloroethane	1.219	1.302	-6.8	104	0.00
25 T	2-Butanone	0.543	0.606	-11.6	110	0.00
26 T	2,2-Dichloropropane	0.954	1.059	-11.0	111	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.747	-4.0	102	0.00
28 T	Bromochloromethane	0.587	0.546	7.0	95	-0.01
29 T	Tetrahydrofuran	0.340	0.373	-9.7	107	0.00
30 C	Chloroform	1.271	1.335	-5.0#	103	0.00
31 T	Cyclohexane	1.111	1.106	0.5	98	0.00
32 T	1,1,1-Trichloroethane	1.101	1.161	-5.4	103	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.862	7.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.360	0.356	1.1	99	0.00
36 T	1,1-Dichloropropene	0.484	0.506	-4.5	100	0.00
37 T	Ethyl Acetate	0.598	0.637	-6.5	103	0.00
38 T	Carbon Tetrachloride	0.544	0.588	-8.1	103	0.00
39 T	Methylcyclohexane	0.623	0.650	-4.3	100	0.00
40 TM	Benzene	1.417	1.496	-5.6	100	0.00
41 T	Methacrylonitrile	0.313	0.370	-18.2	105	0.00
42 TM	1,2-Dichloroethane	0.612	0.651	-6.4	102	0.00
43 T	Isopropyl Acetate	0.912	1.045	-14.6	107	0.00
44 TM	Trichloroethene	0.341	0.364	-6.7	101	0.00
45 C	1,2-Dichloropropane	0.352	0.386	-9.7#	102	0.00
46 T	Dibromomethane	0.278	0.292	-5.0	100	0.00
47 T	Bromodichloromethane	0.547	0.606	-10.8	103	0.00
48 T	Methyl methacrylate	0.466	0.538	-15.5	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	102	0.00
50 S	Toluene-d8	1.246	1.135	8.9	91	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.678	-12.1	105	0.00
52 CM	Toluene	0.869	0.902	-3.8#	99	0.00
53 T	t-1,3-Dichloropropene	0.487	0.557	-14.4	104	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.614	-14.1	104	0.00
55 T	1,1,2-Trichloroethane	0.343	0.370	-7.9	103	0.00
56 T	Ethyl methacrylate	0.546	0.640	-17.2	106	0.00
57 T	1,3-Dichloropropane	0.615	0.650	-5.7	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.306	-10.1	98	0.00
59 T	2-Hexanone	0.448	0.513	-14.5	107	0.00
60 T	Dibromochloromethane	0.376	0.424	-12.8	104	0.00
61 T	1,2-Dibromoethane	0.356	0.380	-6.7	100	0.00
62 S	4-Bromofluorobenzene	0.478	0.469	1.9	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.354	0.378	-6.8	97	0.00
65 PM	Chlorobenzene	1.094	1.168	-6.8	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.411	-9.9	102	0.00
67 C	Ethyl Benzene	1.929	2.117	-9.7#	101	0.00
68 T	m/p-Xylenes	0.706	0.764	-8.2	100	0.00
69 T	o-Xylene	0.688	0.756	-9.9	100	0.00
70 T	Styrene	1.127	1.283	-13.8	102	0.00
71 P	Bromoform	0.281	0.318	-13.2	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.893	4.180	-7.4	103	0.00
74 T	N-amyl acetate	1.924	2.141	-11.3	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.385	-1.5	105	0.00
76 T	1,2,3-Trichloropropane	1.204	1.248	-3.7	107	0.00
77 T	Bromobenzene	0.904	0.954	-5.5	104	0.00
78 T	n-propylbenzene	4.526	4.908	-8.4	103	0.00
79 T	2-Chlorotoluene	2.919	3.010	-3.1	102	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.503	-7.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.419	-13.2	111	0.00
82 T	4-Chlorotoluene	3.238	3.488	-7.7	103	0.00
83 T	tert-Butylbenzene	3.276	3.486	-6.4	103	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.548	-7.7	102	0.00
85 T	sec-Butylbenzene	4.022	4.415	-9.8	105	0.00
86 T	p-Isopropyltoluene	3.320	3.640	-9.6	104	0.00
87 T	1,3-Dichlorobenzene	1.649	1.733	-5.1	103	0.00
88 T	1,4-Dichlorobenzene	1.684	1.742	-3.4	104	0.00
89 T	n-Butylbenzene	2.912	3.307	-13.6	107	0.00
90 T	Hexachloroethane	0.585	0.639	-9.2	104	0.00
91 T	1,2-Dichlorobenzene	1.655	1.740	-5.1	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.331	-9.6	104	0.00
93 T	1,2,4-Trichlorobenzene	0.951	1.029	-8.2	107	0.00
94 T	Hexachlorobutadiene	0.415	0.457	-10.1	109	0.00
95 T	Naphthalene	3.487	3.742	-7.3	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	1.062	-8.3	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	50.000	49.043	1.9	87	0.00
3 P	Chloromethane	50.000	48.377	3.2	93	0.00
4 C	Vinyl Chloride	50.000	48.151	3.7#	94	0.00
5 T	Bromomethane	50.000	43.237	13.5	85	0.00
6 T	Chloroethane	50.000	51.093	-2.2	100	0.00
7 T	Trichlorofluoromethane	50.000	51.858	-3.7	100	0.00
8 T	Diethyl Ether	50.000	51.308	-2.6	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.034	-6.1	105	0.00
10 T	Methyl Iodide	50.000	48.225	3.5	90	0.00
11 T	Tert butyl alcohol	250.000	281.901	-12.8	115	0.00
12 CM	1,1-Dichloroethene	50.000	49.990	0.0#	99	0.00
13 T	Acrolein	250.000	249.577	0.2	98	0.00
14 T	Allyl chloride	50.000	54.671	-9.3	106	0.00
15 T	Acrylonitrile	250.000	269.611	-7.8	104	0.00
16 T	Acetone	250.000	295.318	-18.1	122	0.00
17 T	Carbon Disulfide	50.000	46.054	7.9	89	0.00
18 T	Methyl Acetate	50.000	69.175	-38.3#	142	0.00
19 T	Methyl tert-butyl Ether	50.000	55.198	-10.4	107	0.00
20 T	Methylene Chloride	50.000	48.357	3.3	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.199	-0.4	99	0.00
22 T	Diisopropyl ether	50.000	54.976	-10.0	106	0.00
23 T	Vinyl Acetate	250.000	262.170	-4.9	99	0.00
24 P	1,1-Dichloroethane	50.000	53.406	-6.8	104	0.00
25 T	2-Butanone	250.000	279.346	-11.7	110	0.00
26 T	2,2-Dichloropropane	50.000	55.512	-11.0	111	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.056	-4.1	102	0.00
28 T	Bromochloromethane	50.000	46.549	6.9	95	-0.01
29 T	Tetrahydrofuran	250.000	274.234	-9.7	107	0.00
30 C	Chloroform	50.000	52.542	-5.1#	103	0.00
31 T	Cyclohexane	50.000	49.761	0.5	98	0.00
32 T	1,1,1-Trichloroethane	50.000	52.691	-5.4	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.217	7.6	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	49.466	1.1	99	0.00
36 T	1,1-Dichloropropene	50.000	52.307	-4.6	100	0.00
37 T	Ethyl Acetate	50.000	53.296	-6.6	103	0.00
38 T	Carbon Tetrachloride	50.000	54.045	-8.1	103	0.00
39 T	Methylcyclohexane	50.000	52.202	-4.4	100	0.00
40 TM	Benzene	50.000	52.789	-5.6	100	0.00
41 T	Methacrylonitrile	50.000	59.164	-18.3	105	0.00
42 TM	1,2-Dichloroethane	50.000	53.188	-6.4	102	0.00
43 T	Isopropyl Acetate	50.000	57.328	-14.7	107	0.00
44 TM	Trichloroethene	50.000	53.347	-6.7	101	0.00
45 C	1,2-Dichloropropane	50.000	54.797	-9.6#	102	0.00
46 T	Dibromomethane	50.000	52.494	-5.0	100	0.00
47 T	Bromodichloromethane	50.000	55.341	-10.7	103	0.00
48 T	Methyl methacrylate	50.000	57.719	-15.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound		Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1071.717	-7.2	102	0.00
50 S	Toluene-d8	50.000	45.524	9.0	91	0.00
51 T	4-Methyl-2-Pentanone	250.000	280.031	-12.0	105	0.00
52 CM	Toluene	50.000	51.931	-3.9#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	57.213	-14.4	104	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.106	-14.2	104	0.00
55 T	1,1,2-Trichloroethane	50.000	54.064	-8.1	103	0.00
56 T	Ethyl methacrylate	50.000	58.611	-17.2	106	0.00
57 T	1,3-Dichloropropane	50.000	52.850	-5.7	103	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	274.981	-10.0	98	0.00
59 T	2-Hexanone	250.000	286.265	-14.5	107	0.00
60 T	Dibromochloromethane	50.000	56.300	-12.6	104	0.00
61 T	1,2-Dibromoethane	50.000	53.347	-6.7	100	0.00
62 S	4-Bromofluorobenzene	50.000	49.083	1.8	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	53.362	-6.7	97	0.00
65 PM	Chlorobenzene	50.000	53.386	-6.8	103	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.966	-9.9	102	0.00
67 C	Ethyl Benzene	50.000	54.870	-9.7#	101	0.00
68 T	m/p-Xylenes	100.000	108.288	-8.3	100	0.00
69 T	o-Xylene	50.000	54.982	-10.0	100	0.00
70 T	Styrene	50.000	56.926	-13.9	102	0.00
71 P	Bromoform	50.000	56.651	-13.3	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	53.686	-7.4	103	0.00
74 T	N-ethyl acetate	50.000	55.656	-11.3	105	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.752	-1.5	105	0.00
76 T	1,2,3-Trichloropropane	50.000	51.854	-3.7	107	0.00
77 T	Bromobenzene	50.000	52.773	-5.5	104	0.00
78 T	n-propylbenzene	50.000	54.223	-8.4	103	0.00
79 T	2-Chlorotoluene	50.000	51.549	-3.1	102	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.865	-7.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	56.693	-13.4	111	0.00
82 T	4-Chlorotoluene	50.000	53.864	-7.7	103	0.00
83 T	tert-Butylbenzene	50.000	53.206	-6.4	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.863	-7.7	102	0.00
85 T	sec-Butylbenzene	50.000	54.891	-9.8	105	0.00
86 T	p-Isopropyltoluene	50.000	54.822	-9.6	104	0.00
87 T	1,3-Dichlorobenzene	50.000	52.532	-5.1	103	0.00
88 T	1,4-Dichlorobenzene	50.000	51.703	-3.4	104	0.00
89 T	n-Butylbenzene	50.000	56.775	-13.5	107	0.00
90 T	Hexachloroethane	50.000	54.668	-9.3	104	0.00
91 T	1,2-Dichlorobenzene	50.000	52.558	-5.1	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.746	-9.5	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.131	-8.3	107	0.00
94 T	Hexachlorobutadiene	50.000	55.070	-10.1	109	0.00
95 T	Naphthalene	50.000	53.657	-7.3	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046488.D
Acq On : 04 Jun 2025 10:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 01:34:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.114	-8.2	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046038.D
Acq On : 05 May 2025 09:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.1	8676	PASS
75	95	30	60	56.2	22052	PASS
95	95	100	100	100.0	39213	PASS
96	95	5	9	6.4	2505	PASS
173	174	0.00	2	0.7	185	PASS
174	95	50	100	68.8	26963	PASS
175	174	5	9	7.3	1980	PASS
176	174	95	101	97.0	26147	PASS
177	176	5	9	6.9	1802	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046487.D
Acq On : 04 Jun 2025 09:43
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.0	8608	PASS
75	95	30	60	55.2	21617	PASS
95	95	100	100	100.0	39181	PASS
96	95	5	9	6.9	2684	PASS
173	174	0.00	2	1.2	316	PASS
174	95	50	100	68.4	26784	PASS
175	174	5	9	7.0	1878	PASS
176	174	95	101	98.3	26320	PASS
177	176	5	9	6.8	1780	PASS

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0604WBL01	SDG No.:	Q2201
Lab Sample ID:	VX0604WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046490.D	1		06/04/25 11:04	VX060425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.4		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.1		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	69600	5.55			
540-36-3	1,4-Difluorobenzene	140000	6.757			
3114-55-4	Chlorobenzene-d5	134000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	60000	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046490.D
Acq On : 04 Jun 2025 11:04
Operator : JC/MD
Sample : VX0604WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBL01

Quant Time: Jun 05 01:39:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	69580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139946	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	133992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59967	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69212	53.355	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.720%
35) Dibromofluoromethane	5.379	113	50780	50.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.780%
50) Toluene-d8	8.647	98	175770	50.393	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.780%
62) 4-Bromofluorobenzene	11.079	95	71016	53.079	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.160%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046490.D
Acq On : 04 Jun 2025 11:04
Operator : JC/MD
Sample : VX0604WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBL01

Quant Time: Jun 05 01:39:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.550 min Scan# 71

Delta R.T. 0.006 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBL01

Tgt Ion:168 Resp: 69580

Ion Ratio Lower Upper

168 100

99 73.9 54.9 82.3



#33

1,2-Dichloroethane-d4

Concen: 53.355 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046490.D




Acq: 04 Jun 2025 11:04


Tgt Ion: 65 Resp: 69212

Ion Ratio Lower Upper

65 100

67 48.1 0.0 99.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBL01





Tgt Ion:114 Resp: 139946

Ion Ratio Lower Upper

114 100

63 24.8 0.0 49.2

88 18.0 0.0 33.6



#35

Dibromofluoromethane

Concen: 50.389 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04




Tgt Ion:113 Resp: 50780


Ion Ratio Lower Upper

113 100

111 102.0 83.1 124.7

192 16.4 13.3 19.9





#50

Toluene-d8

Concen: 50.393 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBL01

Tgt Ion: 98 Resp: 175770

Ion Ratio Lower Upper

98 100

100 65.3 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 53.079 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04




Tgt Ion: 95 Resp: 71016


Ion Ratio Lower Upper

95 100

174 67.4 0.0 135.8

176 65.5 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBL01





Tgt Ion:117 Resp: 133992

Ion Ratio Lower Upper

117 100

82 63.0 50.6 76.0

119 31.0 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046490.D

Acq: 04 Jun 2025 11:04




Tgt Ion:152 Resp: 59967

Ion Ratio Lower Upper

152 100

115 66.8 46.9 140.7

150 160.7 0.0 351.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0604WBS01	SDG No.:	Q2201
Lab Sample ID:	VX0604WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046491.D	1		06/04/25 11:27	VX060425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	15.5		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.2		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.6		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.4		0.20	1.00	ug/L
71-43-2	Benzene	18.6		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.8		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.4		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.5		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	19.0		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.1		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	47.3		70 (86) - 130 (113)	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.1		70 (77) - 130 (121)	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	92900	5.544			
540-36-3	1,4-Difluorobenzene	164000	6.757			
3114-55-4	Chlorobenzene-d5	139000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	63900	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046491.D
 Acq On : 04 Jun 2025 11:27
 Operator : JC/MD
 Sample : VX0604WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:39:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	139452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63937	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86820	50.130	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.260%			
35) Dibromofluoromethane	5.379	113	60723	51.267	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.540%			
50) Toluene-d8	8.647	98	193713	47.253	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.500%			
62) 4-Bromofluorobenzene	11.079	95	77214	49.102	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.200%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	22743	15.995	ug/l	99
3) Chloromethane	1.307	50	20855	15.125	ug/l	98
4) Vinyl Chloride	1.374	62	19920	15.523	ug/l	94
5) Bromomethane	1.599	94	9608	16.142	ug/l	100
6) Chloroethane	1.673	64	11730	17.122	ug/l	96
7) Trichlorofluoromethane	1.880	101	34089	17.974	ug/l	95
8) Diethyl Ether	2.130	74	11717	18.148	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	21471	18.294	ug/l	99
10) Methyl Iodide	2.447	142	20760	14.948	ug/l	99
11) Tert butyl alcohol	2.971	59	28971	119.170	ug/l	99
12) 1,1-Dichloroethene	2.313	96	18932	17.187	ug/l	97
13) Acrolein	2.233	56	29271	105.725	ug/l	98
14) Allyl chloride	2.660	41	40362	19.172	ug/l	96
15) Acrylonitrile	3.063	53	74090	106.582	ug/l	98
16) Acetone	2.380	43	73753	106.210	ug/l	99
17) Carbon Disulfide	2.508	76	33026	12.646	ug/l #	95
18) Methyl Acetate	2.703	43	44506	27.620	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	79431	20.568	ug/l	99
20) Methylene Chloride	2.782	84	23740	17.840	ug/l	91
21) trans-1,2-Dichloroethene	3.087	96	19363	17.479	ug/l	95
22) Diisopropyl ether	3.758	45	83789	20.604	ug/l	90
23) Vinyl Acetate	3.721	43	346082	96.761	ug/l	99
24) 1,1-Dichloroethane	3.605	63	44360	19.585	ug/l	99
25) 2-Butanone	4.556	43	111365	110.466	ug/l	97
26) 2,2-Dichloropropane	4.471	77	34526	19.475	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	26144	19.605	ug/l	97
28) Bromochloromethane	4.898	49	23833	21.860	ug/l	97
29) Tetrahydrofuran	5.007	42	69831	110.541	ug/l	99
30) Chloroform	5.093	83	47227	20.005	ug/l	97
31) Cyclohexane	5.465	56	34657	16.791	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	39783	19.440	ug/l	100
36) 1,1-Dichloropropene	5.690	75	27602	17.344	ug/l	98
37) Ethyl Acetate	4.721	43	39738	20.211	ug/l	99
38) Carbon Tetrachloride	5.672	117	32904	18.402	ug/l	95
39) Methylcyclohexane	7.373	83	33061	16.137	ug/l	98
40) Benzene	6.031	78	86835	18.629	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046491.D
 Acq On : 04 Jun 2025 11:27
 Operator : JC/MD
 Sample : VX0604WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:39:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	23318	22.671	ug/l	98
42) 1,2-Dichloroethane	6.086	62	39852	19.809	ug/l	99
43) Isopropyl Acetate	6.342	43	63564	21.191	ug/l	99
44) Trichloroethene	7.123	130	20599	18.361	ug/l	97
45) 1,2-Dichloropropane	7.428	63	22961	19.810	ug/l	96
46) Dibromomethane	7.580	93	17290	18.913	ug/l	98
47) Bromodichloromethane	7.818	83	35372	19.645	ug/l	99
48) Methyl methacrylate	7.696	41	32412	21.158	ug/l	98
49) 1,4-Dioxane	7.659	88	13462	462.828	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	215594	108.282	ug/l	100
52) Toluene	8.714	92	54715	19.143	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	30645	19.148	ug/l	95
54) cis-1,3-Dichloropropene	8.366	75	34352	19.420	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	23114	20.509	ug/l	98
56) Ethyl methacrylate	9.116	69	37067	20.636	ug/l	97
57) 1,3-Dichloropropane	9.305	76	39627	19.578	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	97808	106.806	ug/l	99
59) 2-Hexanone	9.427	43	163451	110.961	ug/l	100
60) Dibromochloromethane	9.519	129	25342	20.474	ug/l	98
61) 1,2-Dibromoethane	9.610	107	23528	20.086	ug/l	100
64) Tetrachloroethene	9.269	164	18715	18.968	ug/l	95
65) Chlorobenzene	10.080	112	59399	19.461	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21063	20.209	ug/l	98
67) Ethyl Benzene	10.189	91	105678	19.642	ug/l	99
68) m/p-Xylenes	10.299	106	78322	39.802	ug/l	99
69) o-Xylene	10.640	106	39544	20.613	ug/l	99
70) Styrene	10.653	104	64399	20.492	ug/l	99
71) Bromoform	10.799	173	15687	20.047	ug/l #	97
73) Isopropylbenzene	10.957	105	105573	21.209	ug/l	99
74) N-amyl acetate	10.842	43	53645	21.810	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	37218	21.336	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	32568m	21.162	ug/l	
77) Bromobenzene	11.195	156	23492	20.328	ug/l	99
78) n-propylbenzene	11.299	91	117360	20.277	ug/l	98
79) 2-Chlorotoluene	11.360	91	75768	20.296	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	85839	20.642	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	9522	20.143	ug/l	99
82) 4-Chlorotoluene	11.451	91	83569	20.186	ug/l	98
83) tert-Butylbenzene	11.713	119	87673	20.930	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	86879	20.630	ug/l	100
85) sec-Butylbenzene	11.890	105	108045	21.008	ug/l	100
86) p-Isopropyltoluene	12.006	119	87255	20.553	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	42116	19.969	ug/l	98
88) 1,4-Dichlorobenzene	12.037	146	44193	20.518	ug/l	98
89) n-Butylbenzene	12.329	91	75030	20.149	ug/l	99
90) Hexachloroethane	12.536	117	14661	19.602	ug/l	96
91) 1,2-Dichlorobenzene	12.329	146	44415	20.986	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	9196	23.797	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	25497	20.975	ug/l	97
94) Hexachlorobutadiene	13.719	225	10415	19.618	ug/l	96
95) Naphthalene	13.774	128	94780	21.259	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	25741	20.522	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046491.D
Acq On : 04 Jun 2025 11:27
Operator : JC/MD
Sample : VX0604WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:39:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046491.D
 Acq On : 04 Jun 2025 11:27
 Operator : JC/MD
 Sample : VX0604WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:39:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:168 Resp: 9289
Ion Ratio Lower Upper
168 100
99 70.2 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#2
Dichlorodifluoromethane
Concen: 15.995 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 85 Resp: 22743
Ion Ratio Lower Upper
85 100
87 31.8 15.7 47.1









#3
Chloromethane
Concen: 15.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion: 50 Resp: 2085
Ion Ratio Lower Upper
50 100
52 32.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#4
Vinyl Chloride
Concen: 15.523 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27


Tgt Ion: 62 Resp: 19920
Ion Ratio Lower Upper
62 100
64 28.0 25.2 37.8





#5
Bromomethane
Concen: 16.142 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01



Tgt Ion: 94 Resp: 9608
Ion Ratio Lower Upper
94 100
96 90.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#6
Chloroethane
Concen: 17.122 ug/l
RT: 1.673 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 64 Resp: 11730
Ion Ratio Lower Upper
64 100
66 32.8 24.3 36.5









#7
Trichlorofluoromethane
Concen: 17.974 ug/l
RT: 1.880 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:101 Resp: 34089
Ion Ratio Lower Upper
101 100
103 59.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#8
Diethyl Ether
Concen: 18.148 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 74 Resp: 11717
Ion Ratio Lower Upper
74 100
45 111.4 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.294 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:101 Resp: 2147
Ion Ratio Lower Upper
101 100
85 47.6 38.6 58.0
151 70.5 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#10
Methyl Iodide
Concen: 14.948 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion:142 Resp: 20760
Ion Ratio Lower Upper
142 100
127 52.6 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 119.170 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 59 Resp: 2897

Ion Ratio Lower Upper

59 100





57 10.5 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#12

1,1-Dichloroethene

Concen: 17.187 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 96 Resp: 18932


Ion Ratio Lower Upper

96 100

61 186.8 146.2 219.2

98 65.7 51.0 76.6





#13

Acrolein

Concen: 105.725 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 56 Resp: 2927

Ion Ratio Lower Upper





56 100

55 71.9 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#14

Allyl chloride

Concen: 19.172 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 41 Resp: 40362


Ion Ratio Lower Upper

41 100

39 72.5 60.6 90.8

76 28.6 24.9 37.3





#15

Acrylonitrile

Concen: 106.582 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBS01

Tgt Ion: 53	Resp: 74090
Ion Ratio	Lower Upper
53 100	
52 83.1	65.3 97.9
51 38.5	29.8 44.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#16

Acetone

Concen: 106.210 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion: 43	Resp: 73753
Ion Ratio	Lower Upper
43 100	
58 26.2	21.2 31.8





#17

Carbon Disulfide

Concen: 12.646 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 76 Resp: 33020

Ion Ratio Lower Upper

76 100





78 10.6 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#18

Methyl Acetate

Concen: 27.620 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 43 Resp: 44506

Ion Ratio Lower Upper

43 100

74 21.2 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 20.568 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion: 73 Resp: 7943
Ion Ratio Lower Upper
73 100
57 21.6 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#20
Methylene Chloride
Concen: 17.840 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27


Tgt Ion: 84 Resp: 23740
Ion Ratio Lower Upper
84 100
49 129.4 113.9 170.9
51 41.3 33.5 50.3
86 59.8 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 17.479 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01




Tgt Ion: 96 Resp: 1936
Ion Ratio Lower Upper
96 100
61 158.4 119.5 179.3
98 63.3 50.0 75.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#22
Diisopropyl ether
Concen: 20.604 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Tgt Ion: 45 Resp: 83789
Ion Ratio Lower Upper
45 100
43 95.4 66.6 100.0
87 23.3 19.8 29.6
59 9.8 8.6 12.8





#23
Vinyl Acetate
Concen: 96.761 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01




Tgt Ion: 43 Resp: 34608
Ion Ratio Lower Upper
43 100
86 9.2 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#24
1,1-Dichloroethane
Concen: 19.585 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27



Tgt Ion: 63 Resp: 44360
Ion Ratio Lower Upper
63 100
98 6.7 3.6 10.8
100 4.2 2.1 6.3





#25

2-Butanone

Concen: 110.466 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 43 Resp: 11136

Ion Ratio Lower Upper

43 100





72 21.4 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#26

2,2-Dichloropropane

Concen: 19.475 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 77 Resp: 34526

Ion Ratio Lower Upper

77 100

97 21.4 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 19.605 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27





Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:	96	Resp:	2614
Ion	Ratio	Lower	Upper
96	100		
61	158.2	0.0	322.8
98	62.0	0.0	129.0

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#28

Bromochloromethane

Concen: 21.860 ug/l




RT: 4.898 min Scan# 625


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	49	Resp:	23833
Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	4.0
130	68.0	56.2	84.2





#29

Tetrahydrofuran

Concen: 110.541 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 42 Resp: 6983

Ion Ratio Lower Upper

42 100

72 41.4 33.7 50.5





71 38.4 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#30

Chloroform

Concen: 20.005 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 83 Resp: 47227

Ion Ratio Lower Upper

83 100

85 64.1 49.3 73.9





#31

Cyclohexane

Concen: 16.791 ug/l

RT: 5.465 min Scan# 718

Delta R.T. 0.000 min

Lab File: VX046491.D


Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01



Tgt Ion: 56 Resp: 3465

Ion Ratio Lower Upper




56 100

69 31.5 24.4 36.6

84 83.8 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#32

1,1,1-Trichloroethane

Concen: 19.440 ug/l

RT: 5.373 min Scan# 703

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 97 Resp: 39783


Ion Ratio Lower Upper

97 100

99 65.2 51.8 77.6

61 50.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 50.130 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 65 Resp: 86820

Ion Ratio Lower Upper

65 100





67 48.8 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion:114 Resp: 164481


Ion Ratio Lower Upper

114 100

63 22.7 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 51.267 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBS01

Tgt Ion:	113	Resp:	6072
Ion	Ratio	Lower	Upper
113	100		
111	102.7	83.1	124.7
192	16.8	13.3	19.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#36

1,1-Dichloropropene

Concen: 17.344 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	75	Resp:	27602
Ion	Ratio	Lower	Upper
75	100		
110	33.2	16.3	48.9
77	32.1	24.3	36.5





#37

Ethyl Acetate

Concen: 20.211 ug/l

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 43 Resp: 39738

Ion Ratio Lower Upper





43 100

61 12.5 10.3 15.5

70 9.4 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#38

Carbon Tetrachloride

Concen: 18.402 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion:117 Resp: 32904


Ion Ratio Lower Upper

117 100

119 98.4 75.2 112.8

121 32.9 24.2 36.4





#39

Methylcyclohexane

Concen: 16.137 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 83 Resp: 3306

Ion Ratio Lower Upper

83 100

55 78.9 64.7 97.1





98 45.4 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#40

Benzene

Concen: 18.629 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 78 Resp: 86835

Ion Ratio Lower Upper

78 100


77 24.0 19.0 28.4





#41
Methacrylonitrile
Concen: 22.671 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.012 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01



Tgt Ion: 41 Resp: 23318
Ion Ratio Lower Upper
41 100
39 55.2 47.2 70.8
67 63.8 50.7 76.1
52 29.9 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#42
1,2-Dichloroethane
Concen: 19.809 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 62 Resp: 39852
Ion Ratio Lower Upper
62 100
98 7.3 0.0 15.2





#43

Isopropyl Acetate

Concen: 21.191 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBS01

Tgt Ion: 43	Resp: 6356
Ion Ratio	Lower Upper
43 100	
61 17.5	14.3 21.5
87 11.7	9.5 14.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#44

Trichloroethene

Concen: 18.361 ug/l




RT: 7.123 min Scan# 990


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:130	Resp: 20599
Ion Ratio	Lower Upper
130 100	
95 98.7	0.0 204.2





#45

1,2-Dichloropropane

Concen: 19.810 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 63 Resp: 2296

Ion Ratio Lower Upper

63 100





65 32.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#46

Dibromomethane

Concen: 18.913 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 93 Resp: 17290


Ion Ratio Lower Upper

93 100

95 84.3 65.6 98.4

174 85.7 68.2 102.2





#47

Bromodichloromethane

Concen: 19.645 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBS01

Tgt Ion:	83	Resp:	3537
Ion	Ratio	Lower	Upper
83	100		
85	62.2	50.5	75.7
127	7.5	6.5	9.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#48

Methyl methacrylate

Concen: 21.158 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	41	Resp:	32412
Ion	Ratio	Lower	Upper
41	100		
69	71.1	58.5	87.7
39	66.4	51.7	77.5





#49

1,4-Dioxane

Concen: 462.828 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 88 Resp: 13462

Ion Ratio Lower Upper

88 100

43 39.1 33.4 50.2





58 73.1 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#50

Toluene-d8

Concen: 47.253 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 98 Resp: 193713

Ion Ratio Lower Upper

98 100

100 66.1 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 108.282 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion: 43 Resp: 215594
Ion Ratio Lower Upper
43 100
58 35.9 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#52
Toluene
Concen: 19.143 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 92 Resp: 54715
Ion Ratio Lower Upper
92 100
91 168.9 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.148 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 75 Resp: 3064





Ion Ratio Lower Upper

75 100

77 33.9 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#54

cis-1,3-Dichloropropene

Concen: 19.420 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 75 Resp: 34352


Ion Ratio Lower Upper

75 100

77 32.4 25.4 38.0

39 62.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 20.509 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 97 Resp: 23114

Ion Ratio Lower Upper

97 100

83 85.8 69.0 103.6

85 56.1 44.3 66.5





99 63.9 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#56

Ethyl methacrylate

Concen: 20.636 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 69 Resp: 37067


Ion Ratio Lower Upper

69 100

41 79.5 60.8 91.2

39 49.5 39.0 58.6





#57

1,3-Dichloropropane

Concen: 19.578 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 76 Resp: 3962





Ion Ratio Lower Upper

76 100

78 33.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#58

2-Chloroethyl Vinyl ether

Concen: 106.806 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion: 63 Resp: 97808

Ion Ratio Lower Upper

63 100

106 27.2 21.5 32.3





#59

2-Hexanone

Concen: 110.961 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 43 Resp: 16345





Ion Ratio Lower Upper

43 100

58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#60

Dibromochloromethane

Concen: 20.474 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion:129 Resp: 25342

Ion Ratio Lower Upper

129 100

127 76.9 39.3 117.8





#61

1,2-Dibromoethane

Concen: 20.086 ug/l

RT: 9.610 min Scan# 1397

Delta R.T. 0.006 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:107 Resp: 23528





Ion Ratio Lower Upper

107 100

109 91.9 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#62

4-Bromofluorobenzene

Concen: 49.102 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion: 95 Resp: 77214


Ion Ratio Lower Upper

95 100

174 67.0 0.0 135.8

176 64.9 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27





Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:	117	Resp:	13945
Ion Ratio	Lower	Upper	
117	100		
82	63.8	50.6	76.0
119	33.8	25.8	38.6

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#64

Tetrachloroethene

Concen: 18.968 ug/l




RT: 9.269 min Scan# 1342


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	164	Resp:	18715
Ion Ratio	Lower	Upper	
164	100		
166	127.6	105.0	157.6
129	111.9	83.5	125.3
131	100.9	76.5	114.7









#65
Chlorobenzene
Concen: 19.461 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:112 Resp: 5939
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.209 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion:131 Resp: 21063
Ion Ratio Lower Upper
131 100
133 96.8 47.3 141.9
119 64.7 31.6 95.0





#67

Ethyl Benzene

Concen: 19.642 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 91 Resp: 105678





Ion Ratio Lower Upper

91 100

106 28.7 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#68

m/p-Xylenes

Concen: 39.802 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion:106 Resp: 78322

Ion Ratio Lower Upper

106 100

91 215.1 171.2 256.8









#69
o-Xylene
Concen: 20.613 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:106 Resp: 3954
Ion Ratio Lower Upper
106 100
91 227.1 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#70
Styrene
Concen: 20.492 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27


Tgt Ion:104 Resp: 64399
Ion Ratio Lower Upper
104 100
78 58.2 45.7 68.5
103 55.0 43.7 65.5





#71
Bromoform
Concen: 20.047 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01




Tgt Ion:173 Resp: 1568
Ion Ratio Lower Upper
173 100
175 48.7 23.4 70.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27



Tgt Ion:152 Resp: 63937
Ion Ratio Lower Upper
152 100
115 77.5 46.9 140.7
150 167.6 0.0 351.0





#73

Isopropylbenzene

Concen: 21.209 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:105 Resp: 10557

Ion Ratio Lower Upper

105 100





120 25.2 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#74

N-ethyl acetate

Concen: 21.810 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion: 43 Resp: 53645




Ion Ratio Lower Upper


43 100

70 37.4 30.9 46.3

55 22.5 18.7 28.1

61 21.5 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 21.336 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBS01

Tgt Ion:	83	Resp:	37218
Ion Ratio	Lower	Upper	
83	100		
131	10.1	5.0	14.9
85	64.6	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#76

1,2,3-Trichloropropane

Concen: 21.162 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	75	Resp:	32568
Ion Ratio	Lower	Upper	
75	100		
77	39.7	20.5	61.5





#77

Bromobenzene

Concen: 20.328 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBS01

Tgt Ion:	156	Resp:	2349
Ion Ratio	Lower	Upper	
156	100		
77	187.5	92.8	278.3
158	95.8	48.8	146.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#78

n-propylbenzene

Concen: 20.277 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	91	Resp:	117360
Ion Ratio	Lower	Upper	
91	100		
120	22.3	10.8	32.4





#79

2-Chlorotoluene

Concen: 20.296 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 91 Resp: 7576





Ion Ratio Lower Upper

91 100

126 30.9 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#80

1,3,5-Trimethylbenzene

Concen: 20.642 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion:105 Resp: 85839

Ion Ratio Lower Upper

105 100


120 46.8 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.143 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01



Tgt Ion: 75 Resp: 952
Ion Ratio Lower Upper
75 100
53 126.6 100.5 150.7
89 45.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#82
4-Chlorotoluene
Concen: 20.186 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion: 91 Resp: 83569
Ion Ratio Lower Upper
91 100
126 27.7 13.3 39.8





#83

tert-Butylbenzene

Concen: 20.930 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27





Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:	119	Resp:	8767
Ion Ratio	Lower	Upper	
119	100		
91	65.8	32.9	98.7
134	23.1	11.4	34.1

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#84

1,2,4-Trimethylbenzene

Concen: 20.630 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Tgt Ion:	105	Resp:	86879
Ion Ratio	Lower	Upper	
105	100		
120	42.5	21.2	63.6









#85
sec-Butylbenzene
Concen: 21.008 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01

Tgt Ion:105 Resp: 108045
Ion Ratio Lower Upper
105 100
134 19.6 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#86
p-Isopropyltoluene
Concen: 20.553 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27

Tgt Ion:119 Resp: 87255
Ion Ratio Lower Upper
119 100
134 24.8 12.5 37.5
91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 19.969 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion:146 Resp: 42110

Ion Ratio Lower Upper

146 100

111 46.1 22.1 66.3





148 63.9 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#88

1,4-Dichlorobenzene

Concen: 20.518 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27




Tgt Ion:146 Resp: 44193


Ion Ratio Lower Upper

146 100

111 43.5 21.3 63.9

148 62.1 31.9 95.5





#89

n-Butylbenzene

Concen: 20.149 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046491.D

Acq: 04 Jun 2025 11:27

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBS01

Tgt Ion: 91 Resp: 75030

Ion Ratio Lower Upper





91 100

92 53.2 26.9 80.7

134 23.8 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#90

Hexachloroethane

Concen: 19.602 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046491.D




Acq: 04 Jun 2025 11:27


Tgt Ion:117 Resp: 14661

Ion Ratio Lower Upper

117 100


201 66.5 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 20.986 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01



Tgt Ion:146 Resp: 4441
Ion Ratio Lower Upper
146 100
111 46.4 22.4 67.3
148 64.4 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 23.797 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27


Tgt Ion: 75 Resp: 9196
Ion Ratio Lower Upper
75 100
155 69.7 34.9 104.8
157 83.7 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 20.975 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01




Tgt Ion:180 Resp: 2549
Ion Ratio Lower Upper
180 100
182 92.9 47.9 143.8
145 33.3 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#94
Hexachlorobutadiene
Concen: 19.618 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Tgt Ion:225 Resp: 10415
Ion Ratio Lower Upper
225 100
223 64.3 30.8 92.4
227 64.9 30.9 92.7





#95
Naphthalene
Concen: 21.259 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBS01




Tgt Ion:128 Resp: 94780
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.0 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#96
1,2,3-Trichlorobenzene
Concen: 20.522 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046491.D
Acq: 04 Jun 2025 11:27



Tgt Ion:180 Resp: 25741
Ion Ratio Lower Upper
180 100
182 95.8 47.8 143.3
145 37.0 18.1 54.3



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VX0604WBSD01	SDG No.:	Q2201
Lab Sample ID:	VX0604WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046497.D	1		06/04/25 13:52	VX060425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	18.8		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.3		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	22.2		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	22.0		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	22.5		0.20	1.00	ug/L
71-43-2	Benzene	21.4		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.6		0.22	1.00	ug/L
79-01-6	Trichloroethene	21.3		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	22.2		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	20.6		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.7		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	52.1		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	49.0		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	84500	5.55			
540-36-3	1,4-Difluorobenzene	153000	6.757			
3114-55-4	Chlorobenzene-d5	133000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	62800	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046497.D
 Acq On : 04 Jun 2025 13:52
 Operator : JC/MD
 Sample : VX0604WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:51:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	84483	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152834	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81368	51.661	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.320%			
35) Dibromofluoromethane	5.385	113	57316	52.079	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.160%			
50) Toluene-d8	8.647	98	186769	49.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.060%			
62) 4-Bromofluorobenzene	11.079	95	75026	51.347	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.700%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	24632	19.049	ug/l	99
3) Chloromethane	1.307	50	23757	18.946	ug/l	97
4) Vinyl Chloride	1.374	62	21918	18.781	ug/l	98
5) Bromomethane	1.593	94	9220	17.033	ug/l	98
6) Chloroethane	1.672	64	15446	24.792	ug/l	94
7) Trichlorofluoromethane	1.880	101	35809	20.761	ug/l	97
8) Diethyl Ether	2.136	74	12725	21.673	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	21610	20.246	ug/l	97
10) Methyl Iodide	2.447	142	24073	19.060	ug/l	100
11) Tert butyl alcohol	2.971	59	27971	126.516	ug/l	99
12) 1,1-Dichloroethene	2.312	96	20333	20.297	ug/l	100
13) Acrolein	2.233	56	25349	100.678	ug/l	97
14) Allyl chloride	2.660	41	42621	22.262	ug/l	95
15) Acrylonitrile	3.062	53	73994	117.045	ug/l	99
16) Acetone	2.386	43	72770	115.231	ug/l	98
17) Carbon Disulfide	2.501	76	39956	16.824	ug/l	100
18) Methyl Acetate	2.703	43	45782	31.241	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	82120	23.382	ug/l	100
20) Methylene Chloride	2.782	84	24899	20.575	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	20553	20.402	ug/l	98
22) Diisopropyl ether	3.763	45	85964	23.245	ug/l	98
23) Vinyl Acetate	3.721	43	352257	108.297	ug/l	100
24) 1,1-Dichloroethane	3.605	63	45805	22.237	ug/l	98
25) 2-Butanone	4.562	43	110519	120.545	ug/l	99
26) 2,2-Dichloropropane	4.471	77	32995	20.465	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	26678	21.998	ug/l	98
28) Bromochloromethane	4.897	49	22848	23.044	ug/l	98
29) Tetrahydrofuran	5.013	42	69743	121.397	ug/l	99
30) Chloroform	5.092	83	48458	22.570	ug/l	97
31) Cyclohexane	5.458	56	39649	21.123	ug/l	96
32) 1,1,1-Trichloroethane	5.379	97	41875	22.500	ug/l	97
36) 1,1-Dichloropropene	5.690	75	30800	20.829	ug/l	99
37) Ethyl Acetate	4.721	43	40135	21.968	ug/l	98
38) Carbon Tetrachloride	5.678	117	33893	20.399	ug/l	97
39) Methylcyclohexane	7.379	83	37949	19.934	ug/l	92
40) Benzene	6.031	78	92846	21.436	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046497.D
 Acq On : 04 Jun 2025 13:52
 Operator : JC/MD
 Sample : VX0604WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:51:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	24996	26.155	ug/l	95
42) 1,2-Dichloroethane	6.086	62	40404	21.614	ug/l	99
43) Isopropyl Acetate	6.342	43	64639	23.192	ug/l	99
44) Trichloroethene	7.123	130	22185	21.281	ug/l	96
45) 1,2-Dichloropropane	7.427	63	23770	22.070	ug/l	98
46) Dibromomethane	7.580	93	18288	21.529	ug/l	99
47) Bromodichloromethane	7.818	83	36336	21.718	ug/l	99
48) Methyl methacrylate	7.696	41	33430	23.486	ug/l	99
49) 1,4-Dioxane	7.659	88	12861	475.862	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	220370	119.115	ug/l	99
52) Toluene	8.714	92	57666	21.713	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	30745	20.675	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	35234	21.437	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	23253	22.204	ug/l	96
56) Ethyl methacrylate	9.116	69	39616	23.736	ug/l	99
57) 1,3-Dichloropropane	9.305	76	41148	21.879	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	98765	116.070	ug/l	99
59) 2-Hexanone	9.427	43	168301	122.961	ug/l	99
60) Dibromochloromethane	9.518	129	24794	21.558	ug/l	99
61) 1,2-Dibromoethane	9.610	107	23687	21.762	ug/l	96
64) Tetrachloroethene	9.269	164	19397	20.578	ug/l	94
65) Chlorobenzene	10.079	112	62199	21.331	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21656	21.749	ug/l	99
67) Ethyl Benzene	10.195	91	113326	22.048	ug/l	99
68) m/p-Xylenes	10.299	106	80924	43.046	ug/l	94
69) o-Xylene	10.640	106	41048	22.397	ug/l	97
70) Styrene	10.652	104	67392	22.447	ug/l	97
71) Bromoform	10.799	173	15427	20.636	ug/l #	98
73) Isopropylbenzene	10.963	105	110318	22.550	ug/l	100
74) N-amyl acetate	10.841	43	55422	22.926	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	38648	22.544	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	32723m	21.634	ug/l	
77) Bromobenzene	11.195	156	24841	21.872	ug/l	99
78) n-propylbenzene	11.305	91	124223	21.838	ug/l	99
79) 2-Chlorotoluene	11.360	91	78907	21.507	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	92894	22.729	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	9297	20.011	ug/l	99
82) 4-Chlorotoluene	11.451	91	88804	21.826	ug/l	99
83) tert-Butylbenzene	11.713	119	92557	22.483	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	92688	22.395	ug/l	99
85) sec-Butylbenzene	11.890	105	111854	22.129	ug/l	99
86) p-Isopropyltoluene	12.006	119	93040	22.299	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	44995	21.707	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	44540	21.040	ug/l	98
89) n-Butylbenzene	12.329	91	79526	21.729	ug/l	98
90) Hexachloroethane	12.536	117	15772	21.457	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	46207	22.214	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	9047	23.821	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	25914	21.691	ug/l	99
94) Hexachlorobutadiene	13.725	225	11422	21.891	ug/l	100
95) Naphthalene	13.774	128	102748	23.449	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	26637	21.608	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046497.D
Acq On : 04 Jun 2025 13:52
Operator : JC/MD
Sample : VX0604WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

Quant Time: Jun 05 01:51:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion:168 Resp: 8448
Ion Ratio Lower Upper
168 100
99 72.2 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#2
Dichlorodifluoromethane
Concen: 19.049 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion: 85 Resp: 24632
Ion Ratio Lower Upper
85 100
87 30.8 15.7 47.1









#3
Chloromethane
Concen: 18.946 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion: 50 Resp: 2375
Ion Ratio Lower Upper
50 100
52 33.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#4
Vinyl Chloride
Concen: 18.781 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion: 62 Resp: 21918
Ion Ratio Lower Upper
62 100
64 32.7 25.2 37.8









#5
Bromomethane
Concen: 17.033 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion: 94 Resp: 9220
Ion Ratio Lower Upper
94 100
96 89.4 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#6
Chloroethane
Concen: 24.792 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52


Tgt Ion: 64 Resp: 15446
Ion Ratio Lower Upper
64 100
66 33.6 24.3 36.5





#7
Trichlorofluoromethane
Concen: 20.761 ug/l
RT: 1.880 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion:101 Resp: 35809
Ion Ratio Lower Upper
101 100
103 61.6 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#8
Diethyl Ether
Concen: 21.673 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52



Tgt Ion: 74 Resp: 12725
Ion Ratio Lower Upper
74 100
45 104.6 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.246 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion:101 Resp: 21610
Ion Ratio Lower Upper
101 100
85 49.2 38.6 58.0
151 72.2 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#10
Methyl Iodide
Concen: 19.060 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion:142 Resp: 24073
Ion Ratio Lower Upper
142 100
127 52.3 41.7 62.5
141 14.2 11.5 17.3





#11

Tert butyl alcohol

Concen: 126.516 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 59 Resp: 2797

Ion Ratio Lower Upper

59 100





57 10.3 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#12

1,1-Dichloroethene

Concen: 20.297 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 96 Resp: 20333


Ion Ratio Lower Upper

96 100

61 183.0 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 100.678 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 56 Resp: 25349





Ion Ratio Lower Upper

56 100

55 72.4 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#14

Allyl chloride

Concen: 22.262 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 41 Resp: 42621


Ion Ratio Lower Upper

41 100

39 70.8 60.6 90.8

76 29.2 24.9 37.3





#15

Acrylonitrile

Concen: 117.045 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBSD01

Tgt Ion:	53	Resp:	73994
Ion	Ratio	Lower	Upper
53	100		
52	82.9	65.3	97.9
51	38.0	29.8	44.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#16

Acetone

Concen: 115.231 ug/l




RT: 2.386 min Scan# 213


Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:	43	Resp:	72770
Ion	Ratio	Lower	Upper
43	100		
58	27.5	21.2	31.8





#17

Carbon Disulfide

Concen: 16.824 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 76 Resp: 3995





Ion Ratio Lower Upper

76 100

78 8.6 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#18

Methyl Acetate

Concen: 31.241 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion: 43 Resp: 45782

Ion Ratio Lower Upper

43 100

74 20.5 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 23.382 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 73 Resp: 82120

Ion Ratio Lower Upper

73 100





57 21.9 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#20

Methylene Chloride

Concen: 20.575 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion: 84 Resp: 24899




Ion Ratio Lower Upper


84 100

49 141.9 113.9 170.9

51 44.1 33.5 50.3

86 62.5 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 20.402 ug/l

RT: 3.087 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 96 Resp: 2055

Ion Ratio Lower Upper

96 100

61 152.2 119.5 179.3





98 64.9 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#22

Diisopropyl ether

Concen: 23.245 ug/l

RT: 3.763 min Scan# 439

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion: 45 Resp: 85964




Ion Ratio Lower Upper


45 100

43 80.9 66.6 100.0

87 24.7 19.8 29.6

59 10.3 8.6 12.8





#23

Vinyl Acetate

Concen: 108.297 ug/l

RT: 3.721 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 43 Resp: 35225

Ion Ratio Lower Upper

43 100





86 9.3 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#24

1,1-Dichloroethane

Concen: 22.237 ug/l

RT: 3.605 min Scan# 413

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 63 Resp: 45805


Ion Ratio Lower Upper

63 100

98 6.3 3.6 10.8

100 3.7 2.1 6.3





#25

2-Butanone

Concen: 120.545 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 43 Resp: 110519

Ion Ratio Lower Upper

43 100





72 23.5 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#26

2,2-Dichloropropane

Concen: 20.465 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion: 77 Resp: 32995

Ion Ratio Lower Upper

77 100

97 20.4 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 21.998 ug/l

RT: 4.489 min Scan# 557

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 96 Resp: 26678

Ion Ratio Lower Upper

96 100

61 158.5 0.0 322.8





98 65.9 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#28

Bromochloromethane

Concen: 23.044 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 49 Resp: 22848


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.0

130 68.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 121.397 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.012 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52





Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion:	42	Resp:	6974
Ion	Ratio	Lower	Upper
42	100		
72	41.9	33.7	50.5
71	38.4	31.4	47.2

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#30

Chloroform

Concen: 22.570 ug/l




RT: 5.092 min Scan# 657


Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:	83	Resp:	48458
Ion	Ratio	Lower	Upper
83	100		
85	63.6	49.3	73.9





#31

Cyclohexane

Concen: 21.123 ug/l

RT: 5.458 min Scan# 717

Delta R.T. -0.006 min

Lab File: VX046497.D


Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01



Tgt Ion: 56 Resp: 39649

Ion Ratio Lower Upper

56 100

69 29.4 24.4 36.6




84 79.0 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#32

1,1,1-Trichloroethane


Concen: 22.500 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52





Tgt Ion: 97 Resp: 41875


Ion Ratio Lower Upper

97 100

99 63.1 51.8 77.6

61 47.9 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 51.661 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 65 Resp: 8136

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion:114 Resp: 152834


Ion Ratio Lower Upper

114 100

63 24.9 0.0 49.2

88 17.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 52.079 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion:113 Resp: 57310

Ion Ratio Lower Upper

113 100

111 102.9 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#36

1,1-Dichloropropene

Concen: 20.829 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 75 Resp: 30800


Ion Ratio Lower Upper

75 100

110 31.5 16.3 48.9

77 30.0 24.3 36.5





#37

Ethyl Acetate

Concen: 21.968 ug/l

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBSD01

Tgt Ion: 43	Resp: 4013
Ion Ratio	Lower Upper
43	100
61	12.1 10.3 15.5
70	10.4 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
 Supervised By :Semsettin Yesilyurt 06/05/2025



#38

Carbon Tetrachloride

Concen: 20.399 ug/l




RT: 5.678 min Scan# 753


Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:117	Resp: 33893
Ion Ratio	Lower Upper
117	100
119	91.9 75.2 112.8
121	32.2 24.2 36.4





#39

Methylcyclohexane

Concen: 19.934 ug/l

RT: 7.379 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 83 Resp: 37949

Ion Ratio Lower Upper

83 100

55 87.5 64.7 97.1





98 53.5 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#40

Benzene

Concen: 21.436 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion: 78 Resp: 92846

Ion Ratio Lower Upper

78 100


77 23.1 19.0 28.4





#41
Methacrylonitrile
Concen: 26.155 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01



Tgt Ion: 41 Resp: 24990
Ion Ratio Lower Upper
41 100
39 56.0 47.2 70.8
67 59.5 50.7 76.1
52 28.1 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#42
1,2-Dichloroethane
Concen: 21.614 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion: 62 Resp: 40404
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 23.192 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBSD01

Tgt Ion: 43	Resp: 64639
Ion Ratio	Lower Upper
43 100	
61 17.6	14.3 21.5
87 11.7	9.5 14.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#44

Trichloroethene

Concen: 21.281 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:130	Resp: 22185
Ion Ratio	Lower Upper
130 100	
95 106.1	0.0 204.2





#45

1,2-Dichloropropane

Concen: 22.070 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 63 Resp: 23770

Ion Ratio Lower Upper

63 100





65 28.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#46

Dibromomethane

Concen: 21.529 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 93 Resp: 18288


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 84.8 68.2 102.2





#47

Bromodichloromethane

Concen: 21.718 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBSD01

Tgt Ion:	83	Resp:	36330
Ion	Ratio	Lower	Upper
83	100		
85	62.2	50.5	75.7
127	7.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#48

Methyl methacrylate

Concen: 23.486 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:	41	Resp:	33430
Ion	Ratio	Lower	Upper
41	100		
69	72.0	58.5	87.7
39	64.7	51.7	77.5





#49

1,4-Dioxane

Concen: 475.862 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 88 Resp: 1286

Ion Ratio Lower Upper

88 100

43 41.5 33.4 50.2





58 72.9 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#50

Toluene-d8

Concen: 49.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion: 98 Resp: 186769

Ion Ratio Lower Upper

98 100

100 64.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 119.115 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion: 43 Resp: 220370
Ion Ratio Lower Upper
43 100
58 35.4 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#52
Toluene
Concen: 21.713 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion: 92 Resp: 57666
Ion Ratio Lower Upper
92 100
91 173.6 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 20.675 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 75 Resp: 3074





Ion Ratio Lower Upper

75 100

77 31.8 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#54

cis-1,3-Dichloropropene

Concen: 21.437 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 75 Resp: 35234


Ion Ratio Lower Upper

75 100

77 30.9 25.4 38.0

39 60.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 22.204 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 97 Resp: 2325

Ion Ratio Lower Upper

97 100

83 89.2 69.0 103.6

85 58.8 44.3 66.5





99 62.9 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#56

Ethyl methacrylate

Concen: 23.736 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52




Tgt Ion: 69 Resp: 39616


Ion Ratio Lower Upper

69 100

41 75.9 60.8 91.2

39 47.0 39.0 58.6





#57

1,3-Dichloropropane

Concen: 21.879 ug/l

RT: 9.305 min Scan# 1114

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 76 Resp: 4114





Ion Ratio Lower Upper

76 100

78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#58

2-Chloroethyl Vinyl ether

Concen: 116.070 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion: 63 Resp: 98765

Ion Ratio Lower Upper

63 100

106 27.3 21.5 32.3





#59

2-Hexanone

Concen: 122.961 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 43 Resp: 16830

Ion Ratio Lower Upper

43 100





58 48.9 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#60

Dibromochloromethane

Concen: 21.558 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion:129 Resp: 24794

Ion Ratio Lower Upper

129 100

127 77.3 39.3 117.8









#61
1,2-Dibromoethane
Concen: 21.762 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion:107 Resp: 2368
Ion Ratio Lower Upper
107 100
109 95.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#62
4-Bromofluorobenzene
Concen: 51.347 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52


Tgt Ion: 95 Resp: 75026
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.1 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. 0.006 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01



Tgt Ion:117 Resp: 133225
Ion Ratio Lower Upper
117 100
82 60.8 50.6 76.0
119 31.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#64
Tetrachloroethene
Concen: 20.578 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion:164 Resp: 19397
Ion Ratio Lower Upper
164 100
166 116.4 105.0 157.6
129 101.6 83.5 125.3
131 96.3 76.5 114.7









#65
Chlorobenzene
Concen: 21.331 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion:112 Resp: 62199
Ion Ratio Lower Upper
112 100
114 32.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.749 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion:131 Resp: 21656
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 64.9 31.6 95.0





#67

Ethyl Benzene

Concen: 22.048 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 91 Resp: 113320





Ion Ratio Lower Upper

91 100

106 30.0 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#68

m/p-Xylenes

Concen: 43.046 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046497.D




Acq: 04 Jun 2025 13:52


Tgt Ion:106 Resp: 80924

Ion Ratio Lower Upper

106 100

91 223.3 171.2 256.8









#69
o-Xylene
Concen: 22.397 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion:106 Resp: 4104
Ion Ratio Lower Upper
106 100
91 230.6 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#70
Styrene
Concen: 22.447 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52


Tgt Ion:104 Resp: 67392
Ion Ratio Lower Upper
104 100
78 59.9 45.7 68.5
103 55.7 43.7 65.5





#71
Bromoform
Concen: 20.636 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01



Tgt Ion:173 Resp: 1542
Ion Ratio Lower Upper
173 100
175 48.1 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Tgt Ion:152 Resp: 62838
Ion Ratio Lower Upper
152 100
115 76.8 46.9 140.7
150 160.6 0.0 351.0





#73

Isopropylbenzene

Concen: 22.550 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion:105 Resp: 110318

Ion Ratio Lower Upper

105 100





120 25.4 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025

Supervised By :Semsettin Yesilyurt 06/05/2025



#74

N-ethyl acetate

Concen: 22.926 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion: 43 Resp: 55422




Ion Ratio Lower Upper


43 100

70 38.5 30.9 46.3

55 22.7 18.7 28.1

61 21.9 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 22.544 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBSD01

Tgt Ion:	83	Resp:	3864
Ion	Ratio	Lower	Upper
83	100		
131	9.7	5.0	14.9
85	62.8	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#76

1,2,3-Trichloropropane

Concen: 21.634 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:	75	Resp:	32723
Ion	Ratio	Lower	Upper
75	100		
77	41.8	20.5	61.5





#77

Bromobenzene

Concen: 21.872 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X





ClientSampleId :

VX0604WBSD01

Tgt Ion:	156	Resp:	2484
Ion Ratio	Lower	Upper	
156	100		
77	186.9	92.8	278.3
158	96.0	48.8	146.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#78

n-propylbenzene

Concen: 21.838 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.006 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Tgt Ion:	91	Resp:	124223
Ion Ratio	Lower	Upper	
91	100		
120	21.3	10.8	32.4









#79
2-Chlorotoluene
Concen: 21.507 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01

Tgt Ion: 91 Resp: 7890
Ion Ratio Lower Upper
91 100
126 31.6 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#80
1,3,5-Trimethylbenzene
Concen: 22.729 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52


Tgt Ion:105 Resp: 92894
Ion Ratio Lower Upper
105 100
120 46.2 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.011 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion: 75 Resp: 929
Ion Ratio Lower Upper
75 100
53 127.5 100.5 150.7
89 45.7 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#82
4-Chlorotoluene
Concen: 21.826 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52



Tgt Ion: 91 Resp: 88804
Ion Ratio Lower Upper
91 100
126 27.0 13.3 39.8





#83

tert-Butylbenzene

Concen: 22.483 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :





VX0604WBSD01

Tgt Ion:	119	Resp:	9255
Ion Ratio	Lower	Upper	
119	100		
91	66.6	32.9	98.7
134	22.8	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#84

1,2,4-Trimethylbenzene

Concen: 22.395 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52


Tgt Ion:	105	Resp:	92688
Ion Ratio	Lower	Upper	
105	100		
120	41.7	21.2	63.6





#85
sec-Butylbenzene
Concen: 22.129 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion:105 Resp: 111854
Ion Ratio Lower Upper
105 100
134 19.2 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#86
p-Isopropyltoluene
Concen: 22.299 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Tgt Ion:119 Resp: 93040
Ion Ratio Lower Upper
119 100
134 24.9 12.5 37.5
91 27.6 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 21.707 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion:146 Resp: 44998
Ion Ratio Lower Upper
146 100
111 43.5 22.1 66.3
148 63.1 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#88
1,4-Dichlorobenzene
Concen: 21.040 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52



Tgt Ion:146 Resp: 44540
Ion Ratio Lower Upper
146 100
111 44.0 21.3 63.9
148 64.9 31.9 95.5





#89

n-Butylbenzene

Concen: 21.729 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52





Instrument :

MSVOA_X

ClientSampleId :

VX0604WBSD01

Tgt Ion: 91	Resp: 79520
Ion Ratio	Lower Upper
91	100
92	52.2 26.9 80.7
134	23.5 11.8 35.3

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025

#90

Hexachloroethane

Concen: 21.457 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046497.D

Acq: 04 Jun 2025 13:52


Tgt Ion: 117	Resp: 15772
Ion Ratio	Lower Upper
117	100
201	61.7 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 22.214 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01



Tgt Ion:146 Resp: 4620
Ion Ratio Lower Upper
146 100
111 46.5 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 23.821 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52


Tgt Ion: 75 Resp: 9047
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 91.2 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 21.691 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion:180 Resp: 25914
Ion Ratio Lower Upper
180 100
182 95.2 47.9 143.8
145 35.5 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025




#94
Hexachlorobutadiene
Concen: 21.891 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Tgt Ion:225 Resp: 11422
Ion Ratio Lower Upper
225 100
223 61.4 30.8 92.4
227 61.7 30.9 92.7





#95
Naphthalene
Concen: 23.449 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
VX0604WBSD01




Tgt Ion:128 Resp: 102748
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 10.7 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/05/2025
Supervised By :Semsettin Yesilyurt 06/05/2025



#96
1,2,3-Trichlorobenzene
Concen: 21.608 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046497.D
Acq: 04 Jun 2025 13:52



Tgt Ion:180 Resp: 26637
Ion Ratio Lower Upper
180 100
182 94.6 47.8 143.3
145 37.5 18.1 54.3



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC020	VX046041.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:13 AM	MMDadoda	5/6/2025 12:42:46 PM	Peak Integrated by Software
VSTDICCC050	VX046042.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:18 AM	MMDadoda	5/6/2025 12:42:48 PM	Peak Integrated by Software
VSTDICC100	VX046043.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:22 AM	MMDadoda	5/6/2025 12:42:50 PM	Peak Integrated by Software
VSTDICC150	VX046044.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:27 AM	MMDadoda	5/6/2025 12:42:53 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	Ethyl Acetate	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,4-Dichlorobenzene	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Bromochloromethane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Ethyl Acetate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Methyl methacrylate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICV050	VX046048.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:45 AM	MMDadoda	5/6/2025 12:41:37 PM	Peak Integrated by Software



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Manual Integration Report

Sequence:	VX060425	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046488.D	1,2,3-Trichloropropane	MMDadoda	6/5/2025 4:44:18 PM	SAM	6/5/2025 4:47:19 PM	Peak Integrated by Software
VSTDCCC050	VX046488.D	Methacrylonitrile	MMDadoda	6/5/2025 4:44:18 PM	SAM	6/5/2025 4:47:19 PM	Peak Integrated by Software
VX0604WBS01	VX046491.D	1,2,3-Trichloropropane	MMDadoda	6/5/2025 4:44:20 PM	SAM	6/5/2025 4:47:20 PM	Peak Integrated by Software
VX0604WBSD01	VX046497.D	1,2,3-Trichloropropane	MMDadoda	6/5/2025 4:44:23 PM	SAM	6/5/2025 4:47:24 PM	Peak Integrated by Software
VSTDCCC050	VX046515.D	1,2,3-Trichloropropane	SAM	6/5/2025 4:47:29 PM	MMdadoda	6/6/2025 1:00:26 AM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046038.D	05 May 2025 09:37	JC/MD	Ok
2	VSTDIC001	VX046039.D	05 May 2025 10:49	JC/MD	Not Ok
3	VSTDIC005	VX046040.D	05 May 2025 11:12	JC/MD	Not Ok
4	VSTDIC020	VX046041.D	05 May 2025 11:35	JC/MD	Ok,M
5	VSTDIC050	VX046042.D	05 May 2025 11:58	JC/MD	Ok,M
6	VSTDIC100	VX046043.D	05 May 2025 12:21	JC/MD	Ok,M
7	VSTDIC150	VX046044.D	05 May 2025 12:45	JC/MD	Ok,M
8	IBLK	VX046045.D	05 May 2025 13:08	JC/MD	Ok
9	VSTDIC005	VX046046.D	05 May 2025 16:04	JC/MD	Ok,M
10	VSTDIC001	VX046047.D	05 May 2025 16:27	JC/MD	Ok,M
11	VSTDICV050	VX046048.D	05 May 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046487.D	04 Jun 2025 09:43	JC/MD	Ok
2	VSTDCCC050	VX046488.D	04 Jun 2025 10:12	JC/MD	Ok,M
3	VX0604MBL01	VX046489.D	04 Jun 2025 10:40	JC/MD	Ok
4	VX0604WBL01	VX046490.D	04 Jun 2025 11:04	JC/MD	Ok
5	VX0604WBS01	VX046491.D	04 Jun 2025 11:27	JC/MD	Ok,M
6	Q2169-03DL	VX046492.D	04 Jun 2025 11:55	JC/MD	Ok
7	Q2168-08DL	VX046493.D	04 Jun 2025 12:18	JC/MD	Ok
8	Q2168-12DL	VX046494.D	04 Jun 2025 12:41	JC/MD	Ok
9	Q2169-01	VX046495.D	04 Jun 2025 13:05	JC/MD	Not Ok
10	Q2175-05	VX046496.D	04 Jun 2025 13:28	JC/MD	Ok
11	VX0604WBSD01	VX046497.D	04 Jun 2025 13:52	JC/MD	Ok,M
12	Q2200-01DL	VX046498.D	04 Jun 2025 14:15	JC/MD	Ok
13	Q2200-02	VX046499.D	04 Jun 2025 14:39	JC/MD	Ok
14	Q2200-05	VX046500.D	04 Jun 2025 15:02	JC/MD	Dilution
15	Q2175-06	VX046501.D	04 Jun 2025 15:26	JC/MD	Dilution
16	IBLK	VX046502.D	04 Jun 2025 15:50	JC/MD	Ok
17	IBLK	VX046503.D	04 Jun 2025 16:13	JC/MD	Ok
18	Q2200-03	VX046504.D	04 Jun 2025 16:37	JC/MD	Ok
19	Q2201-01	VX046505.D	04 Jun 2025 17:01	JC/MD	Ok
20	Q2198-05	VX046506.D	04 Jun 2025 17:25	JC/MD	ReRun
21	IBLK	VX046507.D	04 Jun 2025 17:49	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

22	Q2200-01	VX046508.D	04 Jun 2025 18:12	JC/MD	Dilution
23	Q2200-05DL	VX046509.D	04 Jun 2025 18:36	JC/MD	Ok
24	Q2200-06	VX046510.D	04 Jun 2025 19:00	JC/MD	Ok
25	Q2175-06DL	VX046511.D	04 Jun 2025 19:24	JC/MD	Ok,M
26	VX0604MBS01	VX046512.D	04 Jun 2025 19:47	JC/MD	Ok,M
27	Q2168-11ME	VX046513.D	04 Jun 2025 20:11	JC/MD	Dilution
28	Q2168-07ME	VX046514.D	04 Jun 2025 20:35	JC/MD	Ok
29	VSTDCCC050	VX046515.D	04 Jun 2025 20:59	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046487.D	04 Jun 2025 09:43		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046488.D	04 Jun 2025 10:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0604MBL01	VX0604MBL01	VX046489.D	04 Jun 2025 10:40		JC/MD	Ok
4	VX0604WBL01	VX0604WBL01	VX046490.D	04 Jun 2025 11:04		JC/MD	Ok
5	VX0604WBS01	VX0604WBS01	VX046491.D	04 Jun 2025 11:27	BS failed low for comp. #17	JC/MD	Ok,M
6	Q2169-03DL	303-PPR-2DL	VX046492.D	04 Jun 2025 11:55	vial B pH<2	JC/MD	Ok
7	Q2168-08DL	B3DL	VX046493.D	04 Jun 2025 12:18	vial B pH#5.0	JC/MD	Ok
8	Q2168-12DL	C2DL	VX046494.D	04 Jun 2025 12:41	vial B pH#5.0	JC/MD	Ok
9	Q2169-01	303-PPR-1	VX046495.D	04 Jun 2025 13:05	vial B pH<2;not req	JC/MD	Not Ok
10	Q2175-05	52525-B	VX046496.D	04 Jun 2025 13:28	vial A pH<2 foamy sample	JC/MD	Ok
11	VX0604WBSD01	VX0604WBSD01	VX046497.D	04 Jun 2025 13:52		JC/MD	Ok,M
12	Q2200-01DL	RMW-02B-66-060325D	VX046498.D	04 Jun 2025 14:15	vial A pH<2	JC/MD	Ok
13	Q2200-02	RMW-03B-90-060325	VX046499.D	04 Jun 2025 14:39	vial A pH<2	JC/MD	Ok
14	Q2200-05	MW-11B-37.5-060325	VX046500.D	04 Jun 2025 15:02	vial A pH<2 Need 200X	JC/MD	Dilution
15	Q2175-06	EGR-LIQUID	VX046501.D	04 Jun 2025 15:26	vial A pH<2 Need 2000X	JC/MD	Dilution
16	IBLK	IBLK	VX046502.D	04 Jun 2025 15:50		JC/MD	Ok
17	IBLK	IBLK	VX046503.D	04 Jun 2025 16:13		JC/MD	Ok
18	Q2200-03	EB01-060325	VX046504.D	04 Jun 2025 16:37	vial A pH<2 EB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060425

Review By	Maresh Dadoda	Review On	6/5/2025 4:44:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:48:33 PM
SubDirectory	VX060425	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134124		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134125,VP134126		

19	Q2201-01	MW-01-6.5-060325	VX046505.D	04 Jun 2025 17:01	vial A pH<2	JC/MD	Ok
20	Q2198-05	B-202-GW01	VX046506.D	04 Jun 2025 17:25	vial A pH<2 BS failed low for comp. #17	JC/MD	ReRun
21	IBLK	IBLK	VX046507.D	04 Jun 2025 17:49		JC/MD	Ok
22	Q2200-01	RMW-02B-66-060325	VX046508.D	04 Jun 2025 18:12	vial B pH<2 Need 100X	JC/MD	Dilution
23	Q2200-05DL	MW-11B-37.5-060325D	VX046509.D	04 Jun 2025 18:36	vial B pH<2	JC/MD	Ok
24	Q2200-06	TB-01-060325	VX046510.D	04 Jun 2025 19:00	vial A pH<2 TB	JC/MD	Ok
25	Q2175-06DL	EGR-LIQUIDDL	VX046511.D	04 Jun 2025 19:24	vial B pH<2	JC/MD	Ok,M
26	VX0604MBS01	VX0604MBS01	VX046512.D	04 Jun 2025 19:47		JC/MD	Ok,M
27	Q2168-11ME	C2ME	VX046513.D	04 Jun 2025 20:11	Need 5X	JC/MD	Dilution
28	Q2168-07ME	B3ME	VX046514.D	04 Jun 2025 20:35		JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX046515.D	04 Jun 2025 20:59		JC/MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2201

Test : VOCMS Group3

Prepbatch ID :

Sequence ID/Qc Batch ID: VX060425,

Standard ID :

VP132035,VP132096,VP133174,VP133887,VP133935,VP133953,VP133991,VP134124,VP134125,VP134126,

Chemical ID :

V13391,V13457,V13460,V13465,V13466,V13706,V14290,V14432,V14435,V14503,V14504,V14525,V14526,V14613,V14614,V14620,V14626,V14630,V14631,V14632,V14633,V14711,V14717,V14718,V14721,V14749,V14750,V14811,V14812,V14843,V14921,V14944,V14945,V14946,V14947,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final
Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final
Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133887	05/12/2025	06/23/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/14/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14432 + 1.00000ml of V14435 + 1.00000ml of V14503 + 1.00000ml of V14504 + 1.00000ml of V14525 + 1.00000ml of V14526 + 1.00000ml of V14711 + 1.00000ml of V14717 + 1.00000ml of V14718 + 1.00000ml of V14721 + 1.00000ml of V14749 + 1.00000ml of V14750 + 1.00000ml of V14811 + 1.00000ml of V14812 + 10.60000ml of V14921 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP133935	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
05/21/2025								

FROM 0.25000ml of V14290 + 24.75000ml of V14921 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP133953	05/19/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
05/21/2025								

FROM 0.25000ml of V13391 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP133991	05/22/2025	06/19/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/24/2025

FROM 1.00000ml of V14944 + 1.00000ml of V14945 + 1.00000ml of V14946 + 1.00000ml of V14947 + 21.00000ml of V14620 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP134124	06/04/2025	06/05/2025	John Carlone	None	None	Maresh Dadoda
								06/06/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP133953 = Final Quantity: 40.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134125	06/04/2025	06/05/2025	John Carlone	None	None	Maresh Dadoda
FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP133935 + 0.01250ml of VP132035 + 0.01250ml of VP133887 + 0.01250ml of VP133991 = Final Quantity: 40.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134126	06/04/2025	06/05/2025	John Carlone	None	None	Maresh Dadoda
FROM 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP133935 + 0.01250ml of VP132035 + 0.01250ml of VP133887 + 0.01250ml of VP133991 = Final Quantity: 40.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	05/12/2025 / SAM	08/15/2024 / SAM	V14432

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14503

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14504

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14525

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14526

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14620

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14711

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14717

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14718

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14721

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/13/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14749

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14750

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	11/12/2025	05/12/2025 / SAM	01/08/2025 / SAM	V14811

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	06/30/2026	05/12/2025 / SAM	01/08/2025 / SAM	V14812

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14945

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14946

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14947

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KME)	Lot	DR	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
														(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102366	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 787mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG02JXK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Pentachloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.6	22.9	156-59-2	N/A	or-rat 848mg/kg
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	156-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.6	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.6	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.6	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.6	20.5	74-95-3	N/A	or-rat 106mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.6	20.4	594-20-7	N/A	or-rat 725mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.6	20.5	75-34-3	100 ppm	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.6	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-83-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloropropane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloroethane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.6	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10081-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	109-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-rat 1062mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-rat




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Hexamethylcyclotriphosphorotriazine/Hexyl acetate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethane	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.53
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,2-Trichloropropane	37.77
38	n-Propylbenzene	37.92
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03/17/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		


Compound	(K049)	Lot	Dir.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(Solvent Safety Info. On Attached pg.)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg	
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 75mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg	
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 787mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	H4002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	or-rat 848mg/kg	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg	
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 106mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	or-rat 725mg/kg	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	or-rat 725mg/kg	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-83-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg	
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A	or-rat 2699mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-63-6	N/A	or-rat 5g/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	109-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg	
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1060mg/kg	
64. 1,3-Dichlorobenzene</																	



Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methoxyvinyltoluene/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethane	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.24
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	32.84
28	1,2-Dibromomethane	33.26
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.76
34	Isopropylbenzene/Bromobenzene	36.40
35	cis-1,4-Dichloro-3-butene	37.23
36	1,1,2,2-Tetrachloroethane	37.77
37	n-Propylbenzene	37.92
38	trans-1,4-Dichloro-3-butene	38.05
39	Bromobenzene	38.14
40	1,3,5-Trimethylbenzene	38.50
41	2-Chlorobenzene	38.63
42	4-Chlorobenzene	38.77
43	tert-Butylbenzene	39.76
44	1,2,4-Trimethylbenzene	39.91
45	Perchloroethane	40.17
46	sec-Butylbenzene	40.52
47	p-Isopropylbenzene	41.02
48	1,3-Trichlorobenzene	41.42
49	1,4-Dichlorobenzene	41.83
50	n-Butylbenzene	42.52
51	1,2-Dichlorobenzene	42.18
52	1,2-Dibromopropane	46.12
53	Nitrobenzene	46.40
54	1,2,4-Trifluorobenzene	49.28
55	Hexachlorocyclopentadiene	49.72
56	Naphthalene	50.56
57	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water

   Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>	
Formulated By: Justin Dippold	091424 DATE
<i>Pedro L. Rentas</i>	
Reviewed By: Pedro L. Rentas	091424 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>	
Formulated By:	Justin Dippold
091424 DATE	
<i>Pedro L. Rentas</i>	
Reviewed By:	Pedro L. Rentas
091424 DATE	

Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97
	CAS#: 67-56-1	

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97
	CAS#: 67-56-1	

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97
	CAS#: 67-56-1	

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022

Balance Serial # B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek America, Inc. 11/15/24

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105


Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948



		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water
Lot#: 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 051925
Description: Acrolein

Solvent(s): Water
Lot#: 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Jacobs
ADDRESS: 412 Mt. Kenble Ave., Suite 100
CITY: Morrisstown STATE: NJ ZIP: 07960
ATTENTION: John Yinfante John.Yinfante@Jacobs.com
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: STC Princeton
PROJECT NO.: D3868221 LOCATION: Princeton Junction
PROJECT MANAGER: Mary Murphy
e-mail: Mary.Murphy@Jacobs.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Mary Murphy PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) RUSH 7 day TAT (5 business day) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☒ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

Site specific VOC's (per NJ law)

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS		
			COMP	GRAB	DATE	TIME		F/E	1	2	3	4	5	6	7	8		9	
1.	MW-01-6.5-060325	GW		X	6/3/25	1400	2	X											F = ascorbic acid preservative
2.																			
3.																			
4.																			
5.																			
6.																			
7.																			
8.																			
9.																			
10.																			

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP <u>3.4</u> °C
1. <u>Don Holmes</u>	<u>6/3/25 1600</u>	1. <u>[Signature]</u>	Comments: <u>See work order for list of site specific VOC's</u>
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	PO # <u>148064311</u>
2. <u>[Signature]</u>		2. <u>[Signature]</u>	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>	<u>6.3.2023</u>	3. <u>[Signature]</u>	

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2201	JACO05	Order Date : 6/3/2025 4:10:00 PM	Project Mgr :
Client Name : JACOBS Engineering Grou		Project Name : Former Schlumberger STC	Report Type : Level 3
Client Contact : John Ynfante		Receive DateTime : 6/3/2025 ~~12:00:00~~ AM	EDD Type : CH2MHILL
Invoice Name : JACOBS Engineering Grou		Purchase Order : 17:20	Hard Copy Date :
Invoice Contact : John Ynfante			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2201-01	MW-01-6.5-060225-060325	Water	06/03/2025	14:00	VOCMS Group3		8260-Low	5 Bus. Days	

Relinquished By : 

Date / Time : 6/4/25 1040

SAMPLES RECEIVED ON 6/3/25
SAMPLES PLACED IN SM-REF-2

Received By : 

Date / Time : 6/4/25 10:40 AM RGH 4

Storage Area : VOA Refridgerator Room