

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2201

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX046497.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0604WBSD01	Vinyl chloride	20	18.8	ug/L	94	19		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	20.3	ug/L	102	17		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	22.2	ug/L	111	12		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	22.0	ug/L	110	12		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	22.5	ug/L	113	15		70 (80)	130 (108)	20 (20)
	Benzene	20	21.4	ug/L	107	14		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.6	ug/L	108	9		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	21.3	ug/L	106	14		70 (77)	130 (113)	20 (20)
	1,1,2-Trichloroethane	20	22.2	ug/L	111	7		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	20.6	ug/L	103	8		70 (67)	130 (123)	20 (20)