

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789

8900, Fax: 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID :	Q2202
Test :	Sulfate,TDS
Prepbatch ID :	
Sequence ID/Qc Bat	ch ID: LB135991,LB136008,
	36,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192,WP113193,WP113194,WP VP113333,WP113357,WP113358,
Chemical ID: M6041 W2647 W311	2,W3163,W3180,W3197,
W00+1,**20+1,**011	2,**0 100,**0 100,**0 101,



Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME_	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Jignesh Parikh	
4035	IC ELUENT CONCENTRATE FOR IC-1	<u>WP112796</u>	04/22/2025	10/22/2025	lwona Zarych	WETCHEM_S CALE_5 (WC	None	04/22/2025	
	SC-5)								

			,
п	FROM	2.10000gram of W2647 + 84.75000gram of W3163 + 913.15000ml of W3112 = Final Quantity: 1000.00	າ∩ ml
п	FRON	2.10000gram of \$12047 + 04.75000gram of \$15.15000ml of \$15.15000ml of \$15.15000ml of \$15.15000ml of \$15.15000ml	<i>J</i> O 1111

Recipe ID	<u>NAME</u>	<u>NO.</u>	Prep Date	Expiration Date	<u>Prepared</u> <u>By</u>	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh
2487	Anions 300/9056 calibration standard 1	<u>WP113186</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	05/22/2025

FROM 10.00000ml of W3112 = Final Quantity: 10.000 ml



Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME_	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Jignesh Parikh	
24	Anions 300/9056 calibration standard 2	<u>WP113187</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	05/22/2025	
	(WC)								

<u>FROM</u>	0.20000mi of W3180 + 9.80000 mi of W3112 = Final Quantity: 10.000 mi

Recipe ID	<u>NAME</u>	NO.	Prep Date	Expiration Date	<u>Prepared</u> <u>By</u>	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh
25	Anions 300/9056 calibration standard 3	<u>WP113188</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_P IPETTE_3 (WC)	05/22/2025

FROM 0.40000ml of W3180 + 9.60000ml of W3112 = Final Quantity: 10.000 ml





Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	<u>NAME</u>	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Jignesh Parikh	
26	Anions 300/9056 calibration standard 4	<u>WP113189</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	05/22/2025	
	(WC)								

<u>FROM</u>	0.50000mi of w3180 + 9.50000mi of w3112 = Final Quantity: 10.000 mi

Recipe ID	<u>NAME</u>	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh
3680	Anions 300/9056 calibration standard 5-CCV	<u>WP113190</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_P IPETTE_3 (WC)	05/22/2025

FROM 5.00000ml of W3180 + 95.00000ml of W3112 = Final Quantity: 50.000 ml



Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh	
3679	Anions 300/9056 calibration standard 6	<u>WP113191</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	J	
FROM	(WC)								

Recipe ID	NAME	<u>NO.</u>	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Jignesh Parikh
3681	Anions 300/9056 calibration standard 7	<u>WP113192</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3 (WC)	•

FROM 2.50000ml of W3180 + 7.50000ml of W3112 = Final Quantity: 10.000 ml



Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME.	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh			
3233	Anions 300/9056 ICV-LCS std	<u>WP113193</u>	05/22/2025	05/23/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	05/22/2025			
FROM	(WC)										

FROM	45.00000ml of W3112 + 5.00000 ml of W3197 = Final	Quantity: 50.000 ml

Recipe				Expiration	Prepared			Supervised By
<u>ID</u> 4036	NAME IC ELUENT FOR IC-1	<u>NO.</u> WP113194	Prep Date 05/22/2025	<u>Date</u> 06/22/2025	By Iwona Zarych	<u>ScaleID</u> None	PipetteID Glass	Jignesh Parikh
					ĺ		Pipette-A	05/22/2025

1980.00000ml of W3112 + 20.00000ml of WP112796 $\,=\,$ Final Quantity: 2000.000 $\,$ ml **FROM**



Alliance

Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Jignesh Parikh		
4037	IC H2SO4 FOR IC-1	WP113195	05/22/2025	06/22/2025	lwona Zarych	None	Glass Pipette-A	05/22/2025		
FROM	FROM 5.60000ml of M6041 + 994.40000ml of W3112 = Final Quantity: 1000.000 ml									

		-	

Recipe				Expiration	<u>Prepared</u>			Supervised By
<u>ID</u>	NAME	<u>NO.</u>	Prep Date	<u>Date</u>	<u>By</u>	<u>ScaleID</u>	<u>PipetteID</u>	Iwona Zarych
3233	Anions 300/9056 ICV-LCS std	WP113332	06/03/2025	06/04/2025	Rubina Mughal	None	WETCHEM_F	•
							IPETTE_3	06/04/2025

FROM 45.00000ml of W3112 + 5.00000ml of W3197 = Final Quantity: 50.000 ml



Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	<u>NAME</u>	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Iwona Zarych			
3680	Anions 300/9056 calibration standard 5-CCV	<u>WP113333</u>	06/03/2025	06/04/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3	06/04/2025			
FDOM	(WC)										

<u>FROM</u>	45.00000ml of $\sqrt{3112} + 5.00000ml$ of $\sqrt{3180} = \text{Final Quantity: } 50.000 \text{ ml}$

Recipe ID	<u>NAME</u>	<u>NO.</u>	Prep Date	Expiration Date	<u>Prepared</u> <u>By</u>	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Jignesh Parikh
3680	Anions 300/9056 calibration standard 5-CCV	<u>WP113357</u>	06/04/2025	06/05/2025	lwona Zarych	None	WETCHEM_F IPETTE_3 (WC)	06/05/2025

FROM 45.00000ml of W3112 + 5.00000ml of W3180 = Final Quantity: 50.000 ml





Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID 3233	NAME Anions 300/9056 ICV-LCS std	<u>NO.</u> WP113358	Prep Date 06/04/2025	Expiration Date 06/05/2025	Prepared By Iwona Zarych	<u>ScaleID</u> None	PipettelD WETCHEM_P IPETTE_3	Supervised By Jignesh Parikh 06/05/2025
FROM	45.00000ml of W3112 + 5.00000ml o	of W3197 =	Final Quantity	7: 50.000 ml			(WC)	



CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	03/20/2028	08/16/2024 / mohan	08/16/2024 / mohan	M6041
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3506-5 / SODIUM BICARBONATE, PWD, ACS, 2.5KG	0000240594	06/03/2026	02/24/2020 / AMANDEEP	01/20/2020 / apatel	W2647
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / Iwona	W3112
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date /	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-SX0395-3 / SODIUM CARBONATE ANHYDR 2.5KG	24E3156178	09/30/2027	12/10/2024 / Iwona	12/10/2024 / Iwona	W3163
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	300-CAL-A-500ML / 300.0 Calibration Standard, 500 ml	V2-MEB742616	02/19/2026	02/19/2025 / Iwona	01/27/2025 / Iwona	W3180
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	300-CAL-A-500ML / 300.0 Calibration Standard, 500 ml	040525	04/05/2027	04/08/2025 / Iwona	04/08/2025 / Iwona	W3197

Sodium Bicarbonate, Powder BAKER ANALYZED® A.C.S. Reagent

(sodium hydrogen carbonate)



Material No.: 3506-05 Batch No.: 0000240594

Manufactured Date: 2019/06/05 Retest Date: 2026/06/03

Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (NaHCO3) (dried basis)	99.7 - 100.3 %	100.1
Insoluble Matter	<= 0.015 %	< 0.002
Chloride (Cl)	<= 0.003 %	0.003
Phosphate (PO4)	<= 0.001 %	0.001
Sulfur Compounds (as SO4)	<= 0.003 %	0.003
Calcium (Ca)	<= 0.02 %	0.02
Frace Impurities – Iron (Fe)	<= 0.001 %	0.001
Magnesium (Mg)	<= 0.005 %	0.005
Potassium (K)	<= 0.005 %	0.005
Ammonium (NH4)	<= 5 ppm	5
Trace Impurities – ACS – Heavy Metals (as Pb)	<= 5 ppm	5

For Laboratory, Research or Manufacturing Use Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC



Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium





Material No.: 9673-33

Batch No.: 23D2462010 Manufactured Date: 2023-03-22

Retest Date: 2028-03-20

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS - Assay (H2SO4)	95.0 - 98.0 %	96,1 %
Appearance	Passes Test	Passes Test
ACS - Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS - Substances Reducing Permanganate (as SO2)	≤ 2 ppm	< 2 ppm
Ammonium (NH ₄)	≤ 1 ppm	1 ppm
Chloride (CI)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO ₃)	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO ₄)	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (AI)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities - Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities - Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities - Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities - Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities - Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities - Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities - Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities - Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid BAKER INSTRA-ANALYZED® Reagent For Trace Metal Analysis Low Selenium





Material No.: 9673-33 Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities - Sodium (Na)	≤ 500.0 ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA

Packaging Site: Phillipsburg Mfg Ctr & DC





W3163 Rec. on 12/10/24 by IZ

Certificate of Analysis

Material BDH9284-2.5KG

Material Description BDH SODIUM CARB ANHYD ACS 2.5KG

Grade USPREAGENT (ACS GRADE)

Batch 24E3156178
Reassay Date 09/30/2027
CAS Number 497-19-8
Molecular Formula Na2CO3
Molecular Mass 105.99

Date of Manufacture 09/01/2023

Storage Room Temperature

Material is hygroscopic. Protect from Moisture.

Additional Product Description:

Characteristics	Specifications	Measured Values
Appearance	Fine white granular powder	Fine white granular powder
Calcium	<= 0.03 %	0.003 %
Chloride	<= 0.001 %	0.0003 %
Heavy Metals (as Pb)	<= 0.0005 %	0.0001 %
Insolubles	<= 0.01 %	0.001 %
Iron	<= 0.0005 %	0.0001 %
Loss on Heating	<= 1.0 %	0.03 %
Magnesium	<= 0.005 %	0.001 %
Phosphate	<= 0.001 %	0.001 %
Potassium	<= 0.005 %	0.003 %
Purity	>= 99.5 %	100.0 %
Silica	<= 0.005 %	0.001 %
Sulfur Compounds	<= 0.003 %	0.002 %
Extra Description:	Meets Reagent Specifications for testing USP/NF monographs	

Internal ID #: 710

Signature Additional Information

We certify that this batch conforms to the specifications listed above.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC.

28600 Fountain Parkway, Solon OH 44139 USA

Analysis may have been rounded to significant digits in specification limits

Product meets analytical specifications of the grades listed.

VWR International LLC, Radnor Corporate Center, Suite 200, 100 Matsonford Road, Radnor, PA 19087, USA

Date Printed: 05/31/2024



Certificate of Analysis

300 Technology Drive Christiansburg, VA 24073 USA inorganicventures.com

P: 800-669-6799/540-585-3030 F: 540-585-3012 info@inorganicventures.com

1.0 **ACCREDITATION / REGISTRATION**

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



2.0 PRODUCT DESCRIPTION

Product Code: Multi Analyte Ion Chromatography Solution

Catalog Number: 300-CAL-A

Lot Number: V2-MEB742616

Matrix: H2O

Value / Analyte(s): 150 µg/mL ea:

Sulfate,

100 µg/mL ea: Bromide. 50 µg/mL ea:

o-Phosphate as P, 30 µg/mL ea:

Chloride, Nitrite as N,

25 µg/mL ea: Nitrate as N, 20 µg/mL ea: Fluoride

3.0 **CERTIFIED VALUES AND UNCERTAINTIES**

ANALYTE CERTIFIED VALUE CERTIFIED VALUE ANALYTE 100.0 ± 0.5 μg/mL 30.01 ± 0.13 µg/mL Bromide, Br

Chloride, Cl

20.00 ± 0.07 µg/mL 25.00 ± 0.10 µg/mL Fluoride, F-Nitrate as N, NNO3-

30.00 ± 0.10 μg/mL 50.00 ± 0.18 μg/mL Nitrite as N, NNO2o-Phosphate as P, PPO4

Sulfate, SO4 150.0 ± 0.8 µg/mL

0.999 g/mL (measured at 20 ± 4 °C) Density:

Assay Information:

METHOD	NIST SRM#	SRM LOT#
IC Assay	3184	151130
Fajans	999c	999c
IC Assay	3182	190830
Fajans	999c	999c
IC Assay	3183	140203
IC Assay	3185	170309
IC Assay	Traceable to 40H	08228TH-H2
Calculated	40h	40h
IC Assay	3186	170606
IC Assay	3181	080603
	IC Assay Fajans IC Assay Fajans IC Assay IC Assay IC Assay IC Assay Calculated IC Assay	IC Assay 3184 Fajans 999c IC Assay 3182 Fajans 999c IC Assay 3183 IC Assay 3185 IC Assay Traceable to 40H Calculated 40h IC Assay 3186

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods	Characterization of CRM/RM by One Method
Certified Value, $X_{\text{CRM/RM}}$, where two or more methods of characterization are used is the weighted mean of the results:	Certified Value, X _{CRM/RM} , where one method of characterization is used is the mean of individual results:
$X_{CRM/RM} = \Sigma(w_i) (X_i)$	$X_{CRM/RM} = (X_a) (u_{char \ a})$
X _i = mean of Assay Method i with standard uncertainty u _{char i}	X _a = mean of Assay Method A with
$\mathbf{w_i}$ = the weighting factors for each method calculated using the inverse square of the variance:	\mathbf{u}_{char}^{-} a = the standard uncertainty of characterization Method A
$\mathbf{w_i} = (1/u_{\text{char }i})^2 / (\Sigma(1/(u_{\text{char }i})^2))$	
CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k (u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$	CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k (u^2_{char} a + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$
k = coverage factor = 2	k = coverage factor = 2
$u_{char} = [\Sigma((w_i)^2 (u_{char})^2)]^{1/2}$ where u_{char} are the errors from each characterization method	u _{char a} = the errors from characterization
u _{bb} = bottle to bottle homogeneity standard uncertainty	u _{bb} = bottle to bottle homogeneity standard uncertainty
u _{lts} = long term stability standard uncertainty (storage)	u _{lts} = long term stability standard uncertainty (storage)
u _{ts} = transport stability standard uncertainty	u _{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 CHROMATOGRAM

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures <u>Terms and Conditions of Sale</u>, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° 30° C while in sealed TCT bag.
- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.
- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° 24° C to minimize the effects of transpiration. Use at $20^{\circ} \pm 4^{\circ}$ C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.
- For more information, visit www.inorganicventures.com/TCT

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

April 02, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- April 02, 2029
- The date after which this CRM/RM should not be used.
- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: __
- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS Certificate Prepared By:

Uyen Truong Custom Processing Supervisor

Certificate Approved By:

Thomas Kozikowski Stock VS Manager

Paul R Laine

Certifying Officer:

Paul Gaines Chairman / Senior Technical Director

300.0 / 9056A
Method:
Analyst: NF
Instrument IC-1

On BR- Con NO3 Con HPO4 Con SO4 Method name date time 0 0 0 0 IC1-032125 3/21/2025 10:45 3 3.994 1.001 1.993 5.998 IC1-032125 3/21/2025 11:20 1 2.075 0.523 1.052 3.247 IC1-032125 3/21/2025 11:20 1 1.0.03 2.493 4.968 14.842 IC1-032125 3/21/2025 11:20 1 1.0.35 5.011 10.256 30.502 IC1-032125 3/21/2025 11:20 1 1.0.364 2.568 14.842 IC1-032125 3/21/2025 12:30 1 1.0.364 2.568 14.842 IC1-032125 3/21/2025 12:30 1 1.0.364 2.568 15.189 IC1-032125 3/21/2025 12:30 1 1.0.364 2.569 15.269 IC1-032125 3/21/2025 13:30 1 1.0.387 2.561 15.185 IC1-032125 3/21/2025 13:30 1 0 0 0 0 0 0 1 0 0 0 0 0
Con BR- Con NO3 Con HPO4 Con SO4 2.075 0.523 1.052 3.247 3.994 1.001 1.993 5.998 4.904 1.226 2.407 7.216 10.03 2.493 4.968 14.842 19.975 5.011 10.256 30.502 25.022 6.247 12.323 36.695 10.364 2.568 5.199 14.842 25.022 6.247 12.323 36.695 10.364 2.568 5.199 15.269 0 0 0 0 10.387 2.561 5.226 15.185 10.387 2.561 5.266 10 0.215 0 0 0 0.221 0 2.367 24.892 10.101 2.479 1.573 24.485 0 0 0 0 0 0 0 0 0 0 0 0
Con BR- Con NO3 2.075 0.523 3.994 1.001 4.904 1.226 10.03 2.493 19.975 5.011 25.022 6.247 10.383 2.559 0 0 10.387 2.561 0 0 0.221 0 0 0.0221 0 0 0 0.221 0 0 0 0.63 0.296 0 0 0.643 2.589 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0
Con BR- Con NO3 2.075 0.523 3.994 1.001 4.904 1.226 10.03 2.493 19.975 5.011 25.022 6.247 10.383 2.559 0 0 10.387 2.561 0 0 0.221 0 0 0.0221 0 0 0 0.221 0 0 0 0.63 0.296 0 0 0.643 2.589 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0
Con 10 10 10 10 10 10 10 10 10 10 10 10 10
0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0
0.63 1.20 1.20 1.46 2.998 3.08 3.098 3.098 3.098 3.098 0.0 0.0 0.0 0.0
Con CL- Con 0 0.619 1.199 1.475 3.009 5.988 7.51 3.038 0.122 3.119 0 26.59 28.881 20.1 0.047 427.687 2.364 3.687 3.107 3
0 0.421 0.795 0.977 1.993 4.034 4.979 2.031 0 2.028 0 3.32 2.301 2.172 0 0.275 0 0.433 0.054 0.071 2.072
Ident STD1 STD2 STD3 STD4 STD5 STD6 STD7 ICV ICW ICB CCCB LB135296BSW LB135296BLW Q1711-01 Q1711-02MS Q1711-03MSD Q1711-04 Q1711-04 Q1711-04 Q1711-04 CCV CCB CCC CCC CCCC CCCC CCCC CCCC CC