

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: BNA_F Calibration Date/Time: 06/06/2025 11:36

Lab File ID: BF142640.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.121		-5.6	
Benzaldehyde	0.859	0.766		-10.8	
Phenol-d6	1.429	1.356		-5.1	
Phenol	1.608	1.531		-4.8	20.0
bis(2-Chloroethyl)ether	1.157	1.138		-1.6	
2-Chlorophenol	1.293	1.241		-4.0	
2-Methylphenol	1.010	0.964		-4.6	
2,2-oxybis(1-Chloropropane)	1.944	1.880		-3.3	
Acetophenone	0.443	0.410		-7.4	
3+4-Methylphenols	1.293	1.177		-9.0	
n-Nitroso-di-n-propylamine	0.886	0.812	0.050	-8.4	
Nitrobenzene-d5	0.367	0.343		-6.5	
Hexachloroethane	0.507	0.489		-3.5	
Nitrobenzene	0.331	0.311		-6.0	
Isophorone	0.613	0.566		-7.7	
2-Nitrophenol	0.177	0.174		-1.7	20.0
2,4-Dimethylphenol	0.312	0.292		-6.4	
bis(2-Chloroethoxy)methane	0.383	0.361		-5.7	
2,4-Dichlorophenol	0.283	0.273		-3.5	20.0
Naphthalene	0.985	0.926		-6.0	
4-Chloroaniline	0.399	0.381		-4.5	
Hexachlorobutadiene	0.192	0.181		-5.7	20.0
Caprolactam	0.080	0.071		-11.3	
4-Chloro-3-methylphenol	0.291	0.262		-10.0	20.0
2-Methylnaphthalene	0.620	0.577		-6.9	
Hexachlorocyclopentadiene	0.387	0.355	0.050	-8.3	
2,4,6-Trichlorophenol	0.387	0.375		-3.1	20.0
2-Fluorobiphenyl	1.490	1.353		-9.2	
2,4,5-Trichlorophenol	0.408	0.381		-6.6	
1,1-Biphenyl	1.552	1.469		-5.3	
2-Chloronaphthalene	1.146	1.112		-3.0	
2-Nitroaniline	0.331	0.307		-7.3	
Dimethylphthalate	1.320	1.176		-10.9	
Acenaphthylene	1.936	1.812		-6.4	
2,6-Dinitrotoluene	0.285	0.258		-9.5	
3-Nitroaniline	0.311	0.273		-12.2	
Acenaphthene	1.182	1.087		-8.0	20.0
2,4-Dinitrophenol	0.147	0.127	0.050	-13.6	
4-Nitrophenol	0.230	0.182	0.050	-20.9	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: BNA_F Calibration Date/Time: 06/06/2025 11:36

Lab File ID: BF142640.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.701	1.563		-8.1	
2,4-Dinitrotoluene	0.372	0.330		-11.3	
Diethylphthalate	1.289	1.095		-15.1	
4-Chlorophenyl-phenylether	0.646	0.584		-9.6	
Fluorene	1.314	1.186		-9.7	
4-Nitroaniline	0.282	0.233		-17.4	
4,6-Dinitro-2-methylphenol	0.112	0.113		0.9	
n-Nitrosodiphenylamine	0.684	0.680		-0.6	20.0
2,4,6-Tribromophenol	0.222	0.192		-13.5	
4-Bromophenyl-phenylether	0.236	0.239		1.3	
Hexachlorobenzene	0.262	0.263		0.4	
Atrazine	0.182	0.168		-7.7	
Pentachlorophenol	0.148	0.135		-8.8	20.0
Phenanthrene	1.069	1.013		-5.2	
Anthracene	1.091	1.048		-3.9	
Carbazole	0.933	0.847		-9.2	
Di-n-butylphthalate	1.021	0.903		-11.6	
Fluoranthene	0.999	0.880		-11.9	20.0
Pyrene	1.866	1.563		-16.2	
Terphenyl-d14	1.464	1.196		-18.3	
Butylbenzylphthalate	0.516	0.527		2.1	
3,3-Dichlorobenzidine	0.405	0.465		14.8	
Benzo (a) anthracene	1.336	1.219		-8.8	
Chrysene	1.200	1.183		-1.4	
Bis(2-ethylhexyl)phthalate	0.660	0.831		25.9	
Di-n-octyl phthalate	1.290	1.594		23.6	20.0
Benzo (b) fluoranthene	1.184	1.106		-6.6	
Benzo (k) fluoranthene	1.109	0.999		-9.9	
Benzo (a) pyrene	1.116	1.052		-5.7	20.0
Indeno (1,2,3-cd) pyrene	1.501	1.380		-8.1	
Dibenzo (a,h) anthracene	1.218	1.119		-8.1	
Benzo (g,h,i) perylene	1.218	1.111		-8.8	
1,2,4,5-Tetrachlorobenzene	0.574	0.558		-2.8	
1,4-Dioxane	0.475	0.477		0.4	20.0
2,3,4,6-Tetrachlorophenol	0.341	0.306		-10.3	

All other compounds must meet a minimum RRF of 0.010.