

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: BNA\_M Calibration Date/Time: 06/10/2025 04:48

Lab File ID: BM050262.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND                    | RRF   | RRF040 | MIN<br>RRF | %D    | MAX%D |
|-----------------------------|-------|--------|------------|-------|-------|
| 2-Fluorophenol              | 1.199 | 1.232  |            | 2.8   |       |
| Benzaldehyde                | 0.995 | 0.967  |            | -2.8  |       |
| Phenol-d6                   | 1.578 | 1.600  |            | 1.4   |       |
| Phenol                      | 1.670 | 1.704  |            | 2.0   | 20.0  |
| bis(2-Chloroethyl)ether     | 1.343 | 1.373  |            | 2.2   |       |
| 2-Chlorophenol              | 1.319 | 1.338  |            | 1.4   |       |
| 2-Methylphenol              | 1.102 | 1.105  |            | 0.3   |       |
| 2,2-oxybis(1-Chloropropane) | 0.933 | 1.118  |            | 19.8  |       |
| Acetophenone                | 0.500 | 0.498  |            | -0.4  |       |
| 3+4-Methylphenols           | 1.474 | 1.472  |            | -0.1  |       |
| n-Nitroso-di-n-propylamine  | 0.983 | 1.010  | 0.050      | 2.7   |       |
| Nitrobenzene-d5             | 0.385 | 0.405  |            | 5.2   |       |
| Hexachloroethane            | 0.555 | 0.579  |            | 4.3   |       |
| Nitrobenzene                | 0.350 | 0.366  |            | 4.6   |       |
| Isophorone                  | 0.664 | 0.664  |            | 0.0   |       |
| 2-Nitrophenol               | 0.151 | 0.156  |            | 3.3   | 20.0  |
| 2,4-Dimethylphenol          | 0.310 | 0.307  |            | -1.0  |       |
| bis(2-Chloroethoxy)methane  | 0.434 | 0.449  |            | 3.5   |       |
| 2,4-Dichlorophenol          | 0.287 | 0.284  |            | -1.0  | 20.0  |
| Naphthalene                 | 1.034 | 1.006  |            | -2.7  |       |
| 4-Chloroaniline             | 0.441 | 0.433  |            | -1.8  |       |
| Hexachlorobutadiene         | 0.190 | 0.181  |            | -4.7  | 20.0  |
| Caprolactam                 | 0.091 | 0.088  |            | -3.3  |       |
| 4-Chloro-3-methylphenol     | 0.306 | 0.302  |            | -1.3  | 20.0  |
| 2-Methylnaphthalene         | 0.624 | 0.605  |            | -3.0  |       |
| Hexachlorocyclopentadiene   | 0.351 | 0.358  | 0.050      | 2.0   |       |
| 2,4,6-Trichlorophenol       | 0.380 | 0.384  |            | 1.1   | 20.0  |
| 2-Fluorobiphenyl            | 1.477 | 1.442  |            | -2.4  |       |
| 2,4,5-Trichlorophenol       | 0.417 | 0.418  |            | 0.2   |       |
| 1,1-Biphenyl                | 1.498 | 1.476  |            | -1.5  |       |
| 2-Chloronaphthalene         | 1.164 | 1.145  |            | -1.6  |       |
| 2-Nitroaniline              | 0.256 | 0.296  |            | 15.6  |       |
| Dimethylphthalate           | 1.353 | 1.338  |            | -1.1  |       |
| Acenaphthylene              | 1.883 | 1.863  |            | -1.1  |       |
| 2,6-Dinitrotoluene          | 0.273 | 0.287  |            | 5.1   |       |
| 3-Nitroaniline              | 0.304 | 0.322  |            | 5.9   |       |
| Acenaphthene                | 1.178 | 1.097  |            | -6.9  | 20.0  |
| 2,4-Dinitrophenol           | 0.146 | 0.058  | 0.050      | -60.3 |       |
| 4-Nitrophenol               | 0.259 | 0.266  | 0.050      | 2.7   |       |

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: BNA\_M Calibration Date/Time: 06/10/2025 04:48

Lab File ID: BM050262.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND                   | RRF   | RRF040 | MIN<br>RRF | %D    | MAX%D |
|----------------------------|-------|--------|------------|-------|-------|
| Dibenzofuran               | 1.723 | 1.686  |            | -2.1  |       |
| 2,4-Dinitrotoluene         | 0.360 | 0.377  |            | 4.7   |       |
| Diethylphthalate           | 1.342 | 1.344  |            | 0.1   |       |
| 4-Chlorophenyl-phenylether | 0.638 | 0.622  |            | -2.5  |       |
| Fluorene                   | 1.320 | 1.281  |            | -3.0  |       |
| 4-Nitroaniline             | 0.285 | 0.303  |            | 6.3   |       |
| 4,6-Dinitro-2-methylphenol | 0.109 | 0.068  |            | -37.6 |       |
| n-Nitrosodiphenylamine     | 0.628 | 0.624  |            | -0.6  | 20.0  |
| 2,4,6-Tribromophenol       | 0.227 | 0.228  |            | 0.4   |       |
| 4-Bromophenyl-phenylether  | 0.213 | 0.212  |            | -0.5  |       |
| Hexachlorobenzene          | 0.249 | 0.248  |            | -0.4  |       |
| Atrazine                   | 0.200 | 0.207  |            | 3.5   |       |
| Pentachlorophenol          | 0.162 | 0.224  |            | 38.3  | 20.0  |
| Phenanthrene               | 1.143 | 1.117  |            | -2.3  |       |
| Anthracene                 | 1.143 | 1.125  |            | -1.6  |       |
| Carbazole                  | 1.041 | 1.038  |            | -0.3  |       |
| Di-n-butylphthalate        | 1.133 | 1.226  |            | 8.2   |       |
| Fluoranthene               | 1.205 | 1.223  |            | 1.5   | 20.0  |
| Pyrene                     | 1.348 | 1.288  |            | -4.5  |       |
| Terphenyl-d14              | 1.055 | 1.019  |            | -3.4  |       |
| Butylbenzylphthalate       | 0.450 | 0.493  |            | 9.6   |       |
| 3,3-Dichlorobenzidine      | 0.386 | 0.426  |            | 10.4  |       |
| Benzo (a) anthracene       | 1.266 | 1.246  |            | -1.6  |       |
| Chrysene                   | 1.204 | 1.189  |            | -1.2  |       |
| Bis(2-ethylhexyl)phthalate | 0.693 | 0.784  |            | 13.1  |       |
| Di-n-octyl phthalate       | 1.028 | 1.197  |            | 16.4  | 20.0  |
| Benzo (b) fluoranthene     | 1.163 | 1.171  |            | 0.7   |       |
| Benzo (k) fluoranthene     | 1.187 | 1.165  |            | -1.9  |       |
| Benzo (a) pyrene           | 1.093 | 1.115  |            | 2.0   | 20.0  |
| Indeno (1,2,3-cd) pyrene   | 1.288 | 1.368  |            | 6.2   |       |
| Dibenzo (a,h) anthracene   | 1.045 | 1.119  |            | 7.1   |       |
| Benzo (g,h,i) perylene     | 1.046 | 1.121  |            | 7.2   |       |
| 1,2,4,5-Tetrachlorobenzene | 0.578 | 0.570  |            | -1.4  |       |
| 1,4-Dioxane                | 0.526 | 0.542  |            | 3.0   | 20.0  |
| 2,3,4,6-Tetrachlorophenol  | 0.361 | 0.354  |            | -1.9  |       |

All other compounds must meet a minimum RRF of 0.010.