

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/06/2025

Lab File ID: BF142640.D Time Analyzed: 11:36

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	126582	6.893	487381	8.18	254808	9.94
UPPER LIMIT	253164	7.393	974762	8.681	509616	10.439
LOWER LIMIT	63291	6.393	243691	7.681	127404	9.439
EPA SAMPLE NO.						
01 MW-12-20250603	126541	6.89	428600	8.19	250553	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.