

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: Q2202

SDG NO.: Q2202

Lab File ID: BM050193.D

DFTPP Injection Date: 06/05/2025

Instrument ID: BNA_M

DFTPP Injection Time: 08:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.3
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	37.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.6
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	74.4
443	15.0 - 24.0% of mass 442	14.8 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050194.D	06/05/2025	09:20
SSTDICC005	SSTDICC005	BM050195.D	06/05/2025	09:59
SSTDICC010	SSTDICC010	BM050196.D	06/05/2025	10:38
SSTDICC020	SSTDICC020	BM050197.D	06/05/2025	11:17
SSTDICCC040	SSTDICCC040	BM050198.D	06/05/2025	11:57
SSTDICC050	SSTDICC050	BM050199.D	06/05/2025	12:36
SSTDICC060	SSTDICC060	BM050200.D	06/05/2025	13:16
SSTDICC080	SSTDICC080	BM050201.D	06/05/2025	13:56